A photograph of a woman with blonde hair tied back, climbing a grey, mossy rock face. She is wearing a blue and white patterned tank top and a black climbing shoe with a yellow circular logo on the sole. Her right foot is hooked over a rock ledge, and her left hand is reaching up to grip the rock. The background is a blurred green forest. The text is overlaid on the right side of the image.

Ponseti's clubfoot treatment

**a method in need
of correction?**

Bob Giesberts

Ponseti's clubfoot treatment:
a method in need of correction?

Bob Giesberts

November 1, 2018

Graduation committee

Chairman & secretary

Prof. dr. G.P.M.R. Dewulf (University of Twente)

Supervisor

Prof. dr. ir. G.J. Verkerke (University of Twente)

Co-supervisor

Dr. P.G.M. Maathuis (University Medical Center Groningen)

Members

Prof. dr. ir. A. de Boer (University of Twente)

Prof. dr. ir. H.F.J.M. Koopman (University of Twente)

Dr. M.C. van der Steen (Catharina Hospital Eindhoven)

Dr. R.J.K. Vegter (University Medical Center Groningen)

Prof. dr. S.K. Bulstra (University Medical Center Groningen)

Dr. ir. G.D.S. Ludden (University of Twente)

Paranymphs

Niek Beckers (University of Twente)

Tycho ter Brug (University of Twente)

Printed by Gildeprint, www.gildeprint.nl

Cover image: Paul Lahaye

ISBN: 978-90-365-4587-7

DOI: 10.3990/1.9789036545877

Copyright © 2018 by R.B. Giesberts, Enschede, The Netherlands This dissertation is published under the terms of the Creative Commons Attribution NonCommercial 4.0 International License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is given the appropriate credit, and any changes to the used material are indicated.

PONSETI'S CLUBFOOT TREATMENT:
A METHOD IN NEED OF CORRECTION?

DISSERTATION

to obtain
the degree of doctor at the University of Twente,
on the authority of the rector magnificus
Prof.dr. T.T.M. Palstra,
on account of the decision of the graduation committee,
to be publicly defended
on Wednesday the 28th of November, 2018, at 16:45

by

Robert Bram Giesberts
born on the 7th of March, 1987
in Groningen

iii

This thesis has been approved by:

Prof. dr. ir. G.J. Verkerke (supervisor)

Dr. P.G.M. Maathuis (co-supervisor)

ISBN: 978-90-365-4587-7

Copyright ©2018, R.B. Giesberts

This work is part of the research programme Symbionics with project number 13527, which is (partly) financed by the Netherlands Organisation for Scientific Research (NWO). The funding party had no role in study design, analysis, manuscript preparation, or decision to publish.

This work has also benefited from the advice and financial support of the following companies and organisations. Their support is thankfully acknowledged.

BAAT
MEDICAL

Contents

1	General introduction	1
1.1	The clubfoot	1
1.2	The Ponseti method	5
1.3	Correction	8
1.4	Project goals	11
1.5	Outline of this thesis	12
2	Review	13
2.1	Introduction	15
2.2	Materials and methods	16
2.3	Results	19
2.4	Discussion	26
2.5	Conclusions	30
3	Quantifying the Ponseti Method	31
3.1	Introduction	33
3.2	Materials and methods	35
3.3	Results	37
3.4	Discussion	40
3.5	Conclusion	46

4	Force sensor	47
4.1	Introduction	49
4.2	Sensor design	51
4.3	Experimental section	54
4.4	Results	57
4.5	Discussion	63
4.6	Conclusions	67
5	Dupuytren	69
5.1	Introduction	71
5.2	Methods	73
5.3	Results	78
5.4	Discussion	81
5.5	Conclusions	86
6	Ponseti forces	87
6.1	Introduction	89
6.2	Methods	90
6.3	Results	96
6.4	Discussion	99
6.5	Conclusion	101
7	Ten Cold Clubfeet	103
7.1	Introduction	105
7.2	Patients and methods	106
7.3	Results	110
7.4	Discussion	112
8	Brace design	119

8.1	Introduction	119
8.2	Requirements	123
8.3	Concepts	123
8.4	Reflection	129
8.5	Conclusions and future work	132
9	General discussion	133
9.1	Introduction	133
9.2	Interpretation of findings	134
9.3	Clubfoot mechanics	135
9.4	Limitations	137
9.5	Conclusion	140
	Acknowledgments	169
	About the author	175
	Publications	177

Summary

A clubfoot is a congenital deformity of the foot. It is often treated with the Ponseti method which consists of weekly manipulation and the application of a plaster cast to gradually correct the position of the foot. This method is highly effective, but since it is applied in the first weeks after birth, the plaster casts are inconvenient for the parents making bathing impossible and cuddling unpleasant. Additionally, literature suggests that the treatment can be accelerated without harmful effects.

An instrumented rubber clubfoot model was made to quantify the forces in the Ponseti method. This test showed that the applied forces vary a lot among physicians. A force sensor was developed to measure how these forces on the clubfoot change over time as a result of tissue adaptation in the clubfoot. In all measurements the applied forces gradually decreased to reach an equilibrium within several hours, questioning the validity of the currently accepted weekly cast change interval. Another observation was that the evaporation of the excess water from the plaster casts cause an uncomfortable temperature drop that lasts for 12+ hours. These findings suggest that despite being highly effective, the Ponseti method is not perfect and should be improved.

Concepts for a dynamic clubfoot brace are developed as a possible alternative treatment. Such a brace would closely follow the Ponseti principles but, rather than applying a constant position with a plaster cast, it would apply a constant

low force to co-adapt with the clubfoot. We hypothesize that with such a brace initial correction can be achieved within a week. It has a softer look and feel and can be temporarily taken off to allow easy diaper changes and bathing. More work and research is needed to validate the design and test its efficacy.

Samenvatting

Een klompvoet is een aangeboren afwijking aan de voet. In Nederland wordt de aandoening behandeld met gipsredressies volgens de Ponseti-methode om de voet geleidelijk te corrigeren. In deze methode wordt de voet richting de gewenste stand gemanipuleerd door op specifieke punten op de voet kracht te zetten. In deze stand wordt de voet vastgezet met gipsverband. Na een week wordt het gips vervangen en wordt de voet weer iets verder gecorrigeerd. De behandeling is vrijwel altijd succesvol, maar aangezien de behandeling al snel na de geboorte wordt gestart, is het gips onhandig voor de ouders. Zo kan knuffelen door het ruwe gips onplezierig zijn en mag het gips niet nat worden waardoor badderen onmogelijk is. Met een literatuurstudie is aangetoond dat de behandeling versneld kan worden zonder enig negatief effect.

Met een klompvoetmodel van rubber met ingebouwde druksensoren zijn de gebruikte krachten in de Ponseti-methode in kaart gebracht. Deze meting laat zien dat er grote verschillen zijn in hoe hard artsen tijdens de behandeling op een klompvoet drukken. Een krachtsensor is ontwikkeld om te onderzoeken hoe de krachten op de klompvoet in verloop van tijd veranderen doordat het weefsel in de voet zich aanpast aan de stand die door het gips wordt opgelegd. Deze meting laat zien dat de kracht van de voet op het gips snel afneemt en binnen enkele uren stabiliseert, wat suggereert dat de weefsels zich op dat moment al hebben aangepast aan de nieuwe stand. Dit resultaat trekt de noodzaak van

de huidige wekelijkse gipswissel in twijfel. In deze studie is ook opgemerkt dat het droogproces van het natte gips ervoor zorgt dat de voeten oncomfortabel koud worden gedurende meer dan twaalf uur. Hoewel de Ponseti-methode in de praktijk vrijwel altijd succesvol is, tonen de resultaten van deze studies aan dat de behandeling op bepaalde punten verbeterd kan worden.

Als een mogelijke alternatief voor de Ponseti-methode zijn concepten ontwikkeld voor een dynamische klompvoetbrace. Een dergelijke brace zou de principes van de Ponseti-methode volgen, maar in plaats van de voet vast te zetten met gips zou er een constante kleine kracht worden aangebracht op de voet. Op deze manier past de brace zich aan, aan de correctie van de voet. Met een dergelijke brace zou het mogelijk moeten zijn om de klompvoet binnen een week in de gecorrigeerde stand te krijgen. De brace heeft een zachte uitstraling en kan tijdelijk worden afgedaan zodat het verschonen van een luier makkelijker is en zodat badden ook mogelijk is. De concepten moeten verder ontwikkeld worden en er is meer onderzoek nodig om het bestaande ontwerp te valideren en de werking van de brace te testen.

Chapter 1

General introduction

1.1 The clubfoot

An idiopathic clubfoot — or *congenital talipes equinovarus* (CTEV) — is a congenital deformity of the foot. Approximately 1 per 1000 children are born with unilateral or bilateral clubfoot, which makes it one of the most common congenital deformities (Werler et al., 2013). In the Netherlands every year approximately 190 children are born with a clubfoot (Besselaar, Kamp, Reijman, & van der Steen, n.d.). Its incidence is higher in boys than in girls and in about half of the cases it is bilateral (both feet are affected). However, sex and laterality do not influence the severity of the deformity (Zionts, Jew, Ebrahimzadeh, & Sangiorgio, 2017).

In approximately 20% of cases, clubfoot is associated with arthrogryposis, congenital myotonic dystrophy, myelomeningocele, amniotic band constriction or other genetic syndromes (Gurnett, Boehm, Connolly, Reimschisel, & Dobbs, 2008). In all other cases the deformity is isolated and the cause unknown, hence its classification as idiopathic clubfoot. The aetiology of a clubfoot re-

mains largely unknown, but mechanical or constraining influences such as breach delivery, twins and a deformation of the uterus have been largely ruled out as pathogenesis (Werler et al., 2013).

In normal embryogenesis, the soles of the feet initially face each other. The feet gradually rotate to have the soles face downward. In a clubfoot the primary defect appears to be developmental arrest of the hindlimb from around day 44 of gestation (Zosia, 2003). The early appearance suggests that the deformity is not caused by a defect in the development of the bones but rather a defect in the development of the soft tissue. The aetiology of clubfoot is probably multifactorial, but there is evidence for a genetic contribution to the aetiology (Zosia, 2003). The transcription factors *PITX1* and *TBX4* have both been related to clubfoot (Dobbs & Gurnett, 2012; Gurnett, Alaei, et al., 2008).

1.1.1 Pathoanatomy

A clubfoot can be early recognized with prenatal sonography (Treadwell, Stanitski, & King, 1999) or at birth (FIGURE 1.1). It is a complex combination of four deformations of the foot, resembling appearance of a golf club, hence the name clubfoot (Pittner, Klingele, & Beebe, 2008):

- *cavus*, relative pronation of the forefoot with respect to the hindfoot
- *adductus*, inward rotation of the forefoot
- *varus*, inward rotation of the hindfoot
- *equinus*, plantar flexion of the hindfoot

Studies with electron microscopy (Sano, Uhthoff, Jarvis, Mansingh, & Wenckebach, 1998), scanning acoustic microscopy (Hattori et al., 2007) and biochemical analysis (Wess, 2009) show that the soft tissues at the medioplantar aspect of the clubfoot are contracted. Affected structures that are commonly identified as

(a) Early diagnosis with ultrasound

(b) Diagnosis at birth

(c) A clubfoot left untreated

(d) Treatment with plaster casts

Figure 1.1 | Clubfoot. Clubfoot is diagnosed before or at birth. If left untreated it can be a serious disability. Photos by CURE International.

tight are the Achilles tendon, the *posterior tibial tendon*, the *peroneal tendon*, *plantar talocalcaneonavicular ligament* and the *calcaneofibular ligament* (Eastwood, 2009).

Since the deformity is apparent at birth, its diagnosis and treatment are hardly ever missed. In some low-income countries proper healthcare can be difficult to find and if left untreated, clubfoot results in the patient walking on the lateral side of the foot (FIGURE 1.1C). This can be very painful and often results in skin breakdown, inflammation and even amputation.

1.1.2 Severity

There are two scoring systems commonly used to quantify the severity of clubfoot and to monitor improvements during treatment.

Pirani

The Pirani score (Pirani, 2009) is based on visual evaluation of six clinical signs scored with a 0 for normal, 0.5 for slightly abnormal and 1 for severely abnormal. There are three signs for the hindfoot (e.g. severity of the posterior crease) and three for the midfoot (e.g. severity of the medial crease). A Pirani score of 6 is the most severe whereas a 0 represents a normal foot. If after treatment the Pirani score is one or higher, it is often considered as a failure (Sharma, Yadav, Verma, Gohiya, & Gaur, 2016).

Diméglio

The Diméglio score (Diméglio, Bensahel, Souchet, Mazeau, & Bonnet, 1995) is based on goniometric measurements of the deformation in four planes. The foot is manipulated into its end range while measuring the angle of deflection in four planes. Every angle is given a score ($-20^\circ - 0^\circ = 1$ point, $0^\circ - 20^\circ = 2$ points,

Figure 1.2 | Prognosis. Zhang et al. (2012) found the initial Diméglio score to be a good predictor for the need for surgery. Adapted from (Zhang et al., 2012).

20° – 45° = 3 points, 45° – 90° = 4 points) that represents the severity of the clubfoot. Four additional points can be given for posterior crease, medial crease, cavus and poor muscle condition. The total amount of points is used to grade the severity of the clubfoot. Total points < 5 equals grade I (Benign), 5–9 for grade II (Moderate), 10–14 for grade III (Severe) and 15–20 for grade IV (Very severe).

1.2 The Ponseti method

The Ponseti method (Ponseti, 2008; Ponseti & Smoley, 1963) is worldwide accepted as the golden standard for the treatment of a clubfoot (Hennessey, 2012). It consists of serial manipulation and casting to correct the position of the foot, targeting all four aspects of the deformity in the order as presented in the list

Figure 1.3 | Ponseti method. Many cases of clubfoot are corrected with five casts.

above. To correct the adductus and the varus, in the Ponseti method the physician manipulates the foot by applying pressing on the medial aspect of the first metatarsal (FM) while giving counter pressure on the lateral aspect of the talar neck (TN). In this newly imposed position the foot is fixated with the application of plaster cast, which is typically changed on a weekly basis. After approximately 4 to 6 casts — and in many cases a percutaneous Achilles tenotomy to correct the equinus — most clubfeet are corrected. Finally an abduction brace is worn for several years to maintain the correction (FIGURE 1.3) (Dobbs et al., 2004).

1.2.1 Prognosis

Clubfeet with a higher initial severity score can expect a longer and more intensive treatment. They require more Ponseti casts (Lampasi, Abati, Bettuzzi, Stilli, & Trisolino, 2018), are more likely to require an Achilles tenotomy (Chu, Labar, Sala, van Bosse, & Lehman, 2010; Dyer & Davis, 2006; Scher, Feldman, van Bosse, Sala, & Lehman, 2004), are more likely to experience a relapse (Sangiorgio, Ebrahimzadeh, Morgan, & Zions, 2017), and are more likely to require additional, more invasive surgery (Goldstein, Seehausen, Chu, Sala, & Lehman, 2014; Zhang et al., 2012) (see FIGURE 1.2).

1.2.2 Ponseti - benefits

The Ponseti method has earned its de facto golden standard status for its efficacy, both short-term and long-term (Laaveg & Ponseti, 1980). It gives superior results

over surgery (Herzenberg, Radler, & Bor, 2002), Kite's method (Derzsi, Nagy, Gozar, Gurzu, & Pop, 2015; Matos & de Oliveira, 2010) and similar to the French method (Richards et al., 2008). Studies showing high initial success rates often exceeding 90% are published on a monthly basis (e.g. this month (Jul-2018) Kannan & Umashankar, 2018; S. Mohan, Kumar, Prashanth, & Krishnagopal, 2018; Salim, Ahmed, & Rehman, 2018). Long-term results are mostly affected by brace compliance (Dobbs et al., 2004).

Besides its efficacy, the Ponseti method is relatively easy to implement almost anywhere in the world (Shabtai, Specht, & Herzenberg, 2014), it is non-invasive and it only requires low-cost materials and equipment. As such, in the Netherlands, the guidelines for the treatment of clubfoot have a clear recommendation for the Ponseti method (Besselaar et al., 2017).

1.2.3 Ponseti - downsides

Since a clubfoot is treated soon after birth, the plaster casts of the Ponseti method can be a burden for the caretakers of the patients. It can be difficult to change diapers without soiling the cast and bathing the child is not possible for an entire week when the cast is on (Nogueira et al., 2013). Other downsides that are mentioned by caretakers are the wet and cold casts, the different feeling of a hard cast when cuddling with their child and judgmental reactions from others in response to the casts (Pletch, Morcuende, Barriga, Segura, & Salas, 2015). The frequent hospital visits both during the casting phase and the bracing phase can be cumbersome for the caretakers.

Some aspects of the Ponseti method such as the manipulation time and the required pressure are not described in much detail in the literature and its exact mechanics are not fully understood (Cohen, 2015; Cummings, Davidson, Armstrong, & Lehman, 2002). That is, we know that it works well, but we don't know why it works so well. It is not always necessary to fully understand a

treatment for it to be implemented successfully (e.g. deep brain stimulation for Parkinson (de Hemptinne et al., 2015)). However, a better understanding can help in further improvement of the treatment method.

1.3 Correction

It has been known for centuries that biological tissues can be shaped, given that ancient cultures have found ways to extend their lower lips, their necks or earlobes. In all these examples an external force is used to permanently stretch the involved parts of the human body.

1.3.1 Mechanostat

How biological structures respond to external forces is known as mechanotransduction and studied in the field of mechanobiology (which is much different from biomechanics). The theory of a *mechanostat* has been proposed, which states that the living body has a regulatory system to control the stresses in the body (Frost, 1987, 2003). Simply put, where stresses are low, tissue is removed (e.g. bone loss around implants) and where stresses are high, tissue is reenforced (e.g. added bone around screws). The theory applies well on bones (i.e. Wolff's law (Wolff, 1892)) and may have analogs in soft tissues (Frost, 2003).

If a load is applied, the mechanosensors (e.g. osteocytes) in the body detect a mechanical stress and stimulate fibroblasts and fibroclast to renew the tissues (de Morree, 2014). The effective response of the tissues is to reduce the stress by lengthening the tissues. Research with rat tail tendons shows a clear response on an imposed position: the force reduces over time (Sanjeevi, 1982).

The immediate response to an applied load is viscoelastic stress-relaxation, attributed in part to the inhibition of reflex activity which is readily reversible (Magnusson, 1998). In muscles, longer periods of uninterrupted stretching can

stimulate the addition of sarcomeres in series, effectively lengthening the muscle-fibers (Tabary, Tabary, Tardieu, Tardieu, & Goldspink, 1972). This possibly already happens after 15 minutes (Williams, 1990). In tendons, in their relaxed state the fibers are curled up (known as crimp) and allow for a large amount of stretch due to ‘uncrimping’ when tension is applied. It is this uncrimping behavior that allows the correction in the Ponseti method (Ponseti, 2008).

1.3.2 Stress-relaxation vs creep

In material science two phenomena describe the behavior of a viscoelastic material (Özkaya, Nordin, Goldsheyder, & Leger, 2012). If a constant stretch is applied, the resulting stress decreases over time. This is known as stress-relaxation. In contrast, if a constant stress is applied, the resulting stretch increases over time. This is known as creep. Both processes are explained in **FIGURE 1.4**.

1.3.3 High load brief stress vs low load prolonged stress

In physical therapy a distinction is made between high load brief stress (HLBS) and low load prolonged stress (LLPS) (Glasgow, Tooth, & Fleming, 2010). With HLBS a “dose” of treatment can be applied in a short period of time, such as with joint mobilization or passive exercise. In contrast, with LLPS the same dose is applied over an extended period of time, such as with splinting and casting. A certain amount of stress over time is needed to realize a certain amount of permanent stretch of the biological tissue. This principle is used in the treatment of deformities and contractures.

1.3.4 Constant-position vs constant-force

Static splints and plaster casts apply a constant position, inducing a stress-relaxation response in the affected biological materials. A dynamic splint applies

Figure 1.4 | Creep vs stress-relaxation. **A** Stress-relaxation is the process in which a constant position (or *stretch*) is applied and the material responds by a decreasing *stress*. An example can be found in elastic bands in cloths that wear out over time. **B** Alternatively, creep is the process in which a constant *stress* (or force) is applied and the material responds by a increasing *stretch*.

a constant force, inducing a creep response in the affected biological materials. There is a high level of evidence for the use of mobilizing splinting in the management of joint contractures (Glasgow et al., 2010).

1.3.5 Trends

Following the general trend of minimally invasive surgery, the treatment of deformities and contractures has become more conservative, abandoning the option of surgery for its downsides on the long-term (e.g. Herzenberg et al., 2002; Hueston, n.d.). A more recent trend is the use of orthoses rather than serial casting (Farmer & James, 2009). However, this trend has not yet been adopted in the treatment of clubfoot.

1.4 Project goals

The goal of this project is to investigate if it would be possible to apply the strategy of using an orthosis to the treatment of clubfoot, and as such develop concepts for this strategy, while still following the proven Ponseti method. In comparison to plaster casts an orthosis can allow bathing, will be more hygienic and will not cool the baby, making it a more user-friendly treatment option.

Our philosophy is that a force-controlled approach could be more efficient than a position-controlled approach. Our hypothesis is that the clubfoot treatment can be improved by the development of an orthosis that is based on a force-controlled approach instead of a position-controlled approach, because it will accelerate the correction process, making it a more efficient treatment option.

1.5 Outline of this thesis

This thesis presents original work done on the treatment of clubfeet and covers several studies that have been performed in the course of this PhD project.

There is no theoretical evidence which proves that a clubfoot would need a week to adapt to the position of the plaster cast. In CHAPTER 2 a review is performed on the length of the cast change interval, to see if practical evidence can be found which supports favoring the traditional weekly interval over more accelerated versions.

Little is known about the forces used in the Ponseti method and a quantification of these forces is much needed. CHAPTER 3 describes how we determined how much force physicians apply to manipulate the clubfoot. To investigate our hypothesis that the treatment process can be accelerated with a force-controlled approach, we have measured how the forces under the plaster cast change over time. To do so, first a new force sensor was developed (CHAPTER 4) which was tested in the treatment of Dupuytren fingers with a hand splint (CHAPTER 5). CHAPTER 6 describes how the forces under the cast decrease over time and whether or not a weekly cast change interval is justified. As a side product of this study, the temperature under the cast was also measured, which gave surprising results, presented in CHAPTER 7.

From the insights gathered from studying these aspects of the treatment of clubfoot, CHAPTER 8 presents how a dynamic brace could be a more user-friendly — and possibly more efficient — method of treating clubfoot, without abandoning the sequence and force application method as defined by Ponseti.

Finally, CHAPTER 9 presents a general discussion.

Chapter 2

Influence of cast change interval in the Ponseti method

BOB GIESBERTS, MARIEKE VAN DER STEEN, PATRICK MAATHUIS, ARNOLD BESSELAAR, EDSKO HEKMAN AND BART VERKERKE

This chapter is published as R.B. Giesberts, M.C. van der Steen, P.G.M. Maathuis, A.T. Besselaar, E.E.G. Hekman and G.J. Verkerke (2018) Influence of cast change interval in the Ponseti method: a systematic review, *PLOS One*, 13(6):e0199540. doi: 10.1371/journal.pone.0199540.

Abstract

Background: Clubfeet are commonly treated using the Ponseti method. This method involves weekly manipulation and casting which gradually corrects the position of the foot. However, the reasons for following a weekly interval are not clear.

Question / Purpose: The aim is to investigate the influence of the cast change interval on treatment outcomes in the Ponseti method.

Methods: We performed a systematic review of comparative studies in which the cast change interval was varied. Scientific databases were searched for relevant publications, screened for eligibility and assessed for a risk of bias. A 'best evidence' synthesis tool was used to synthesize the results of the included studies and draw conclusions from relevant clinical outcomes.

Results: Nine papers matched the inclusion criteria, which provided data of 587 subjects who had a total of 870 clubfeet. There is strong evidence for a positive relation between cast change interval and treatment duration. However, there is no evidence for any relation between the cast change interval and the required number of casts, tenotomy rate, required surgery or failure rate.

Conclusions: Accelerated versions are as effective and safe as the traditional Ponseti method. However, more research is needed to assess the long-term results and to identify an optimal cast change interval.

2.1 Introduction

Clubfeet are commonly treated with serial casting, termed the Ponseti method (Ponseti & Smoley, 1963). In this method, the clubfoot is manipulated and fixated in a corrected position using a plaster cast. This cast is typically changed every week. Occasionally, the cast is changed more frequently in an attempt to minimize the treatment duration (e.g. Harnett, Freeman, Harrison, Brown, & Beckles, 2011). In low-income countries this can be especially important if practitioners are scarce and patients live far away from the hospital.

The original papers on the Ponseti method state that a new cast should be applied at four to seven (Ponseti & Smoley, 1963) or five to seven-day intervals (Cooper & Dietz, 1995). However, the time period is not motivated here nor have we found solid motivation elsewhere in literature. The reason for having a weekly treatment interval seems to be largely pragmatic. Clustering patients on one fixed day of the week is convenient for hospital planning and makes it easier to organize the necessary care around a newborn. Additionally, accommodating all clubfoot patients on a single day gives parents the opportunity to share their concerns with other parents (Terrazas-Lafargue & Morcuende, 2007).

Shortening the interval time could be advantageous for several reasons. Parents who do not have ready access to treatment facilities may have to leave home for the treatment duration, putting strain on the family. But even for those who can make outpatient visits to a clinic, shortening the treatment time will reduce the time during which their family life is interrupted. However, shortening the treatment time can only be done if a shorter interval is not detrimental to treatment outcome, if the number of hospital visits does not increase, and if no additional discomfort or pain is caused to the children by increasing the rate of correction. With this review we want to assess the influence of a shorter cast change interval.

When looking at the application of serial casting to stretch soft tissue in

other disorders than clubfoot, we have found inconclusive or even contradictory evidence for the optimal cast change interval. In hand therapy, this is suggested to be two days (Bell-Krotoski & Figarola, 1995). However, removing the cast of a contracted finger after three days resulted in an increase in range of motion of 3.0° although removing the cast after six days resulted in an increase of 5.2° (Flowers & LaStayo, 1994). When the cast change interval was changed in the treatment of elbow, knee, wrist and ankle contractures, it was found that an interval of 1–3 days resulted in a shorter treatment with fewer complications when compared to an interval of 5–7 days (Pohl et al., 2002). As the underlying mechanisms causing these contractures might differ from the more complex clubfoot, the question remains how these findings on cast change intervals apply to the treatment of clubfoot.

The current systematic review compares the results from applying accelerated versions of the Ponseti method to the results of weekly cast changes, to investigate the influence of the cast change interval on treatment outcomes.

2.2 Materials and methods

A systematic review was performed on the existing literature dealing with the Ponseti method regarding the influence of cast change interval on treatment outcomes.

2.2.1 Search protocol

A PRISMA-driven (Hutton et al., 2014) systematic search of the PubMed, COCHRANE, WebOfKnowledge, Scopus, PeDRO, CINAHL and Google Scholar databases was conducted in October 2017 to identify relevant papers published between January 2005 and October 2017. The used search string was “*Ponseti AND clubfoot AND duration AND cast* AND (Pirani OR Dimeglio)*”. Reference lists of the full-texts

retrieved for eligibility were screened to identify further relevant studies.

2.2.2 Eligibility criteria and study selection

Comparative studies in which the cast change interval was varied were included. Full-text papers needed to be available written in English, German, French or Dutch. Exclusion criteria were conference abstracts, meta-analyses or review papers. Additional exclusion criteria were studies on non-idiopathic clubfoot, modifications to the original Ponseti method other than to the cast change interval and studies without a control group. Two reviewers (RBG, MCvdS) independently assessed the relevance of the identified papers based on the title and abstract. In a second stage, full-text papers were checked against inclusion and exclusion criteria (by RBG and MCvdS). Any doubts about eligibility were resolved by discussions between the two reviewers.

2.2.3 Data extraction

Each selected paper was reviewed (by RBG) to extract relevant patient data, cast change interval, number of casts, treatment duration, required surgery, relapse rate and failure rate. Attempts were made to contact the authors of each selected paper for clarification and to access the raw data for a deeper analysis of the presented results. For each paper, this data was extracted for both the Normal group (weekly cast changes) and the Accelerated group (shortened interval). Complications were extracted as defined by the selected paper. Treatment duration was defined as the time from the application of the first cast until the removal of the final cast prior to Achilles tenotomy. Required surgery was defined as any form of surgery after Ponseti treatment, including re-tenotomy. As defined by the Iowa Group, relapse was considered to be the appearance of any of the components of the deformity, including cavus, adductus, varus, and equinus (Morcuende, Abbasi, Dolan, & Ponseti, 2005). Failure was defined as

2 a post-cast-treatment Pirani score higher than 1.0, as defined by others (Gilani, Ahmed, Obaid-Ur-Rehman, & Bashir, 2014; Harnett et al., 2011; Sharma et al., 2016).

2.2.4 Quality assessment

The selected papers were independently assessed by the two reviewers (RBG, MCvdS) using Cochrane Collaborations tools for assessing risk of bias (Higgins, Altman, & Sterne, 2011), scoring the papers with a '+', '?' or '-'. Criteria used were selection bias (1 - randomization of groups, 2 - comparability of both groups), attrition bias (3 - sufficiency of follow-up, 4 - definition of treatment outcomes), reporting bias (5 - documentation of treatment outcomes), detection bias (6 - blinded measurement of treatment outcomes) and performance bias (7 - blinded participants and personnel). Disagreements were solved during a consensus meeting. To be classified as low risk of bias, items 1, 2, 4 and 5 of the quality assessment needed to be scored as positive.

2.2.5 Qualitative synthesis

Pooling of the data was considered impossible due to the clinical differences of the included studies and differences in reporting methods. Therefore, a 'best evidence' synthesis was performed as qualitative synthesis of the results. Based on the system used by (Van Tulder, Furlan, Bombardier, Bouter, & Group, 2003), the ranking of levels of evidence was used from the method formulated by (van Meer et al., 2015) (TABLE 2.1).

2.2.6 Statistical analysis

If statistical testing had not been performed in the original paper but the available extracted data allowed for it, differences in surgery, relapse and failure rate were

Table 2.1 | The ranking of the levels of evidence is based on van Meer et al. (2015); Van Tulder et al. (2003)

Level	Description
Strong evidence	Two or more studies with low risk of bias and by generally consistent findings in all studies ($\geq 75\%$ of the studies reported consistent findings)
Moderate evidence	One low risk of bias study and two or more high risk of bias studies and by generally consistent findings in all studies ($\geq 75\%$)
Limited evidence	One or more high risk of bias studies or one low risk of bias study and by generally consistent findings ($\geq 75\%$)
Conflicting evidence	Conflicting findings ($< 75\%$ of the studies reported consistent findings)
No evidence	No studies could be found

tested for statistical significance using the two-tailed χ^2 test or Fisher's exact test. Statistical significance was defined as $p < 0.05$.

2.3 Results

The abstracts of 389 papers were screened for relevance. A total of 96 papers were found to contain relevant information of which nine met the inclusion criteria (Elgohary & Abulsaad, 2015; Gilani et al., 2014; Harnett et al., 2011; Ibraheem et al., 2017; Mageshwaran, Murali, Devendran, Yoosuf, & Anandan, 2016; Morcuende et al., 2005; Sahu, Rajavelu, & Tudu, 2015; Sharma et al., 2016; R. J. Xu, 2011). FIGURE 2.1 shows the flow diagram of the papers.

Figure 2.1 | PRISMA flow diagram. Process of study identification and selection for outcome analysis (Hutton et al., 2014).

2.3.1 Study characteristics

Characteristics of the selected papers are presented in TABLE 2.2, and these papers cover the treatment of a total of 587 subjects with a total of 870 clubfeet.

2.3.2 Quality assessment

The results of the risk of bias analysis are presented in TABLE 2.3. Three studies were classified as having a low risk of bias (Elgohary & Abulsaad, 2015; Harnett et al., 2011; Sharma et al., 2016). The other six were classified as having a high risk of bias (Gilani et al., 2014; Ibraheem et al., 2017; Mageshwaran et al., 2016;

Table 2.2 | The ranking of the levels of evidence is based on van Meer et al. (2015); Van Tulder et al. (2003)

Study	Study design	Group	Nr of subjects	Nr of clubfeet	Age at onset (months)	Prior treatment (%)	Pre-treatment Pirani score	Follow-up (months)
Elgohary (2015)	Prospective	N	20	34	2.5	11.8	5.17	25.25
	randomized	A	21	32	2.7	15.6	5.13	23.38
Gilani (2014)	Prospective	N	40	61	5.09	0 ^b	4.12	-
	randomized	A	40	62	4.57		4.35	
Harnett (2011)	Prospective	N	21	32	1.0	-	5.0 ^a	8.0
	randomized	A	19	29	0.7		5.5 ^a	8.5
Ibraheem (2017)	Prospective	N	14	23	1.1	0 ^b	5	-
	randomized	A	14	22	1.9		4.8	-
Mageshwaran (2016)	Prospective	N	20	26	0.93	-	4.97	3 or 6
	randomized	A	20	25	0.92		5.025	3 or 6
Morcuede (2005)	Retrospective	N	111	162	5	72 ^c	-	-
	non-randomized	A	108	157	3			
Sahu (2015)	Prospective	N	27	40	1.3	0 ^b	5.03	11 ^{c,e}
	randomized	A	26	40	1.21		5.3	
Sharma (2016)	Prospective	N	20	26	0.75	0 ^b	5.32	7.7
	randomized	A	20	27	0.77		5.21	8.2
Xu (2011)	Unknown	N	20	32	2.10	-	4.0 ^d	48 ^c
	non-randomized	A	26	40	3.09		4.1 ^d	

N = Normal group, A = Accelerated group.

Unless indicated otherwise, data is presented as it is in the selected paper, as means.

Not reported data is indicated with '-'.^a

^a presented as median

^b patients with any form of prior treatment were excluded

^c for both groups combined

^d calculated over presented data

^e personal communication

Morcuende et al., 2005; Sahu et al., 2015; R. J. Xu, 2011).

2.3.3 Qualitative synthesis

The extracted data from the selected papers is presented in TABLE 2.4. The only surgery, relapse or failure rate data that scored positively when tested for a statistically significant difference was the number of relapses in Morcuende et al. (2005), as had already been reported in the original paper.

Only negligible differences — if any — in terms of required number of casts, tenotomy rate, required surgery or failure rate were reported between the groups in the selected papers (TABLE 2.4). The best evidence synthesis revealed a strong evidence for the absence of a relation between cast change interval and these clinical outcomes (TABLE 2.5). Five of the selected studies report that no (short-term) complications were observed in either group (Elgohary & Abulsaad, 2015; Harnett et al., 2011; Mageshwaran et al., 2016; Sharma et al., 2016; R. J. Xu, 2011), and the remaining studies did not report on complications.

2.3.4 Relapse

The best evidence synthesis gave moderate evidence for the absence of a relation between cast change interval and relapse TABLE 2.5. Morcuende et al. (2005) report a significant difference in number of relapses (11/108 in the accelerated group vs 25/111 in the traditional group, $p = 0.01$) while others did not report such an observation. The division of groups in this study was not random but based on geographical location, and relapse rates are highly influenced by brace compliance (Dobbs et al., 2004). The only study (R. J. Xu, 2011) with sufficiently long follow-up did not report on long-term results such as relapse. Those that did report on relapse rate had either a short (Elgohary & Abulsaad, 2015; Harnett et al., 2011; Mageshwaran et al., 2016; Sahu et al., 2015; Sharma et al., 2016) and/or poorly-defined (Morcuende et al., 2005; Sahu et al., 2015) follow-up.

Table 2.3 | Risk of bias analysis

Study	1 - Randomization of groups	2 - Comparability of groups	3 - Follow-up sufficiency	4 - Definition of outcomes	5 - Documentation of outcomes	6 - Blinded assessment of outcomes	7 - Blinded participants and personnel	Risk of bias
Elgohary (2015)	+	+	?	+	+	?	-	Low
Gilani (2014)	+	?	-	-	-	?	-	High
Harnett (2011)	+	+	?	+	+	?	-	Low
Ibraheem (2017)	+	+	-	+	?	-	-	High
Mageshwaran (2016)	+	+	-	?	+	?	-	High
Morcuende (2016)	-	-	?	+	-	?	-	High
Sahu (2015)	+	+	-	+	?	?	-	High
Sharma (2016)	+	+	-	+	+	?	-	Low
Xu (2011)	-	?	+	+	+	?	-	High

'+' was given if the used methodology was clear and adequate, and all required data was present
'?' was given if the used methodology was unclear or statistical information was missing
'-' was given if the used methodology was faulty or data was missing or not presented per subgroup
Note that in order to be classified as low risk of bias, items 1, 2, 4 and 5 needed to be positively scored

Table 2.4 | Extracted data from the selected papers

Study	Group	Interval (days)	Average nr of casts	Duration (days)	Complications	Tenotomy rate (%)	Failure rate (Pirani > 1.0) (n(%))	Surgery (n(%))	Relapse (n(%))
Elgohary (2015)	N	7	4.88	33.36	0 ^b	91.2	0	3 (9%)	5 (15%)
	A	3.5	5.16	18.13		93.8	0	3 (9%)	5 (16%)
Gilani (2014)	N	7	5.2	36.4^d	-	71.2 ^b	2 (5%)	2 (5%)	-
	A	3.5	5.12	17.92^d			4 (10%)	4 (10%)	
Harnett (2011)	N	7	5 ^a	42	0 ^b	52	2 (10%)	2 (10%)	0 ^b
	A	2.3	5 ^a	16		79	3 (16%) ^e	1 (5%)	
Ibraheem (2017)	N	7	5.26	52	-	96	0	-	-
	A	3.5	6.23	39		100	0		
Mageshwaran (2016)	N	7	5.55	52.8	0 ^b	11.5	-	0	3 (15%)
	A	3.5	5.95	39.65		24		1 (5%)	4 (20%)
Morcuende (2005)	N	7	4 ^b	24	-	81	-	21 (10%) ^b	25 (23%)
	A	5		16		85			11 (10%)
Sahu (2016)	N	7	6.2 ^c	57.4^c	-	78	-	0	9 (23%)
	A	3	7.4 ^c	23.8^c		83	-	1 (4%)	13 (33%)
Sharma (2016)	N	7	5.08	35.24	0 ^b	77	1 (5%)	1 (5%)	0 ^b
	A	3.5	4.15	14.19		74	1 (5%)	1 (5%)	
Xu (2011)	N	7	5.25	35.35	0 ^b	87.5	5 (16%)	4 (13%)	-
	A	3.5	5.04	20.61		87.5	6 (15%)	6 (15%)	

N = Normal group, A = Accelerated group.

Unless indicated otherwise, data is presented as it is in the selected paper.

Not reported data is indicated with '-'.^a

Numbers in **bold** represent a statistical significant difference ($p < 0.05$).

^a presented as median

^b for both groups combined

^c including tenotomy cast

^d calculated from cast x interval, not included in the best evidence synthesis

^e three subjects crossed-over to the control group because they still had Pirani > 1.0 after 21 days

Table 2.5 | Best evidence synthesis of outcome measures

	Nr of studies	Statistically significant difference	No statistically significant difference	Best evidence synthesis	Comments
Average nr of casts	7	LR: 0 HR: 0	LR: 3 HR: 4	Strong evidence no relation	Statistical significance not reported in Sahu et al. (2015). Mean 6.2 (range 4–10) vs 7.4 (5–10) casts: inconclusive. No statistical test in Gilani et al. (2014). Mean±SD 5.2±1.62 vs 5.12±1.53 casts: assumed as no statistical significant difference.
Tenotomy rate	8	LR: 0 HR: 0	LR: 3 HR: 5	Strong evidence no relation	
Duration	6	LR: 3 HR: 3	LR: 0 HR: 0	Strong evidence positive relation	Statistical significance not reported in Sahu et al. (2015). Mean 57.4 vs 23.8 days: inconclusive
Surgery	7	LR: 0 HR: 0	LR: 3 HR: 4	Moderate evidence no relation	Morcuende et al. (2005) surgery data for both groups combined: inconclusive
Relapse	4	LR: 0 HR: 1	LR: 1 HR: 2	Strong evidence no relation	
Failure (Pirani > 1.0)	5	LR: 0 HR: 0	LR: 3 HR: 2	Strong evidence no relation	No statistical test in Elgohary and Abulsaad (2015). No difference in post-treatment Pirani score (reported p = 0.89), all scores were ≤ 1.0: assumed as no statistical significant difference.

LR = Low Risk of bias, HR = High Risk of bias

2.3.5 Treatment duration

There is strong evidence for a relation between cast change interval and treatment duration, but not for the number of casts required to correct clubfoot. Since a shorter interval decreases the time per cast, but the total number of casts does not change, the treatment duration is significantly shorter in the accelerated groups of all studies.

2.4 Discussion

An increasing number of studies continues to be published in which the Ponseti methods interval is varied from its traditional weekly level. Only few studies use a control group to investigate the influence of the cast change interval on treatment outcomes. Individually, the selected nine studies were able to show the feasibility of an accelerated Ponseti method, but could not identify clear differences in clinical treatment outcomes. The current systematic review attempts to present additional conclusive evidence based on a best evidence synthesis, and to identify missing information.

Overall, this review suggests that the accelerated versions of the Ponseti method are in fact as effective as the traditional method in the initial correction of idiopathic clubfoot. The shorter cast change intervals cause a decreased treatment duration without deteriorating clinical outcome.

None of the other selected studies can confirm the conclusion of Morcuende et al. (2005) who, based on anecdotal observations only, suggest that a 5-day interval is the shortest safe interval. In line with this, cohort studies involving weekly cast-changes report approximately 20% of patients experiencing short-term complications such as blisters and skin problems, compared to 16% of those undergoing a cast-change twice a week (Baindurashvili, Kenis, & Stepanova, 2012; S. M. Mohan, Pawan, Mohan, & Madhu, 2014; Sugandhavesa, Cheewawat-

tanachai, Luevitoonvechkij, & Khunsree, 2015). However, insufficient follow-up was available in the selected papers. As such, this review could not address any possible problems that might arise during the bracing period, for example brace compliance. Furthermore, long term results of the accelerated Ponseti method including proper relapse rate and functional outcome remain unknown.

The biggest advantage of a shorter treatment duration is for the caregivers of the patients. Especially caregivers in low-income countries with limited access to healthcare centers would benefit from a shortened casting period (Harnett et al., 2011; Ibraheem et al., 2017; Ramahenina, OConnor, & Chamberlain, 2016; Sharma et al., 2016). Other authors suggest that benefits of a shorter treatment are the reduced risk of skin problems, cast slipping and osteopenia (Harnett et al., 2011; Sharma et al., 2016).

2.4.1 Limitations

Since the outcome of the Ponseti method is affected by factors such as experience (Cosma & Vasilescu, 2014) or the strictness of the adherence to the Ponseti method (Miller et al., 2015), treatment outcomes such as the required number of casts vary among health care institutions. A form of bias would be introduced if results from different institutions would be compared. We have therefore deliberately chosen to only include controlled trials. Inevitably, this choice has limited the selected number of studies which in combination with the heterogeneity in research methodology and data reporting did not allow for any quantitative meta-analysis within this systematic review.

It would not have been possible to blind participants and personnel involved from the cast change interval used during the treatment. Therefore this item was excluded for the classification of low/high risk of bias. Non-blinding during the assessment of treatment outcomes in follow-up might have led to performance bias. Similarly, insufficient or poorly defined follow-up increases the risk of attri-

tion bias, but this domain was excluded from the judgement of risk since mostly short-term outcome measures were extracted (e.g. treatment duration, casts, tenotomy). The long-term effects of shorter cast change intervals remain poorly documented.

For all selected studies, it was assumed that the practitioners themselves had assessed the treatment outcomes. Although non-independent assessment is a risk of attrition bias, we assume this risk to be minimal as outcome measures (for example, whether or not tenotomy or surgery was performed) were mostly objective.

Researchers who altered the cast change intervals and achieved poor results might have refrained from publishing them. Therefore, this risk of publication bias might have caused the best evidence synthesis in the current systematic review to be too optimistic.

For practical reasons, we have limited our search to the languages English, German, French and Dutch, and to the translated abstracts of leading journals in other languages which we found during our search.

2.4.2 Optimal cast change interval

According to the best evidence synthesis, no evidence exists to support the use of a cast change interval of one week. The combined results of the best evidence synthesis in this report can be used to suggest some fundamental principles for the correction of a clubfoot. One scholar mentioned that *“the tissues might need some time in the corrected position in the cast to be able to adapt through this growth and change”* (Radler, 2013). If the adaptation time were to be longer than the used cast change interval (e.g. 5 days versus half a week), then the accelerated method would only achieve partial adaptation. Such incomplete adaptation would have resulted in adverse treatment outcomes (for example, more casts, more complications and more surgery). However, none of the studies made such

observations, which implies that the tissues might need less than a third of a week to adapt, as shown by Harnett et al. (2011). As long as treatment outcomes do not worsen as a result of increasingly shorter cast change intervals, the interval still exceeds the required adaptation time, and there is still margin for further acceleration of the treatment process.

Elgohary and Abulsaad (2015) asked the question “*what is the least time interval between castings to be applied safely?*”. From the numbers available, we are unable to definitively answer that question, but it does seem that the limit has not yet been reached. To the best of our knowledge, the shortest reported interval can be found in a case study by Sutcliffe, Vaea, Poulivaati, and Evans (2013). After four casts in one week, the Pirani scores of two young patients were 1.5 and 3.5. In this case study the correction was still insufficient and more casts might have been required to achieve optimum correction. Anecdotal information exists for even shorter intervals with even daily cast changes. Cummings et al. (2002) mention that “*More rapid correction has been achieved with more frequent (daily) cast changes and manipulation*”, but without going into detail. A large randomized controlled trial with multiple cast change intervals (weekly, twice per week, daily, etc.) with adequate follow up is necessary to determine the optimal cast change interval.

2.4.3 Mechanobiology

What the shortest acceptable time interval is might be ascertained by considering theories from the discipline of mechanobiology. Much of the knowledge about the viscoelastic behavior of biological materials is based on experiments using material from cadavers, such as rat tail tendons (Sanjeevi, 1982). When subjected to a constant strain, the resulting stress reduces over time, known as stress-relaxation. The stresses in dead collagen fibers obtained from rat tail tendons reached equilibrium within several minutes (Sanjeevi, 1982; Sanjeevi, Somanathan, & Ramaswamy, 1982).

However, little is yet known about the combined behavior of stress-relaxation, tissue remodeling and growth of living biological tissue. During research into limb lengthening, it was observed that most of the relaxation occurs within the first two hours after distraction (Rahman, Akins, Wee, & Mackenzie, 2011). In serial casting of contracted elbows, knees and ankles, it was observed that the tension within the cast dropped by 80% within the first 24 hours (Strik, 2000). A high initial pressure was measured underneath the Ponseti cast although pressure sores are rarely observed, which suggests a rapid decrease of this pressure resulting from a short adaptation time (Giesberts, Hekman, Maathuis, & Verkerke, 2017, CHAPTER 3). Preliminary cast / clubfoot interface pressure measurements by one of the authors suggest that the tissues may have reached full adaptation just several hours after casting.

2.5 Conclusions

There is strong evidence that accelerated versions of the Ponseti method can safely be used in the treatment of clubfoot without risking any increase in the number of casts, the failure rate or the surgery rate. Shorter intervals significantly decrease treatment duration, which means that for each patient the most convenient duration may be selected. More research is needed to determine any existing optimal cast change interval and to investigate the long-term effects of shorter cast change intervals.

Chapter 3

Quantifying the Ponseti Method

BOB GIESBERTS, EDSKO HEKMAN, PATRICK MAATHUIS AND BART
VERKERKE

This chapter is published as R.B. Giesberts, E.E.G. Hekman, P.G.M. Maathuis and G.J. Verkerke (2017) Quantifying the Ponseti Method, *Journal of the Mechanical Behavior of Biomedical Materials*, 66, 45-49. doi: 10.1016/j.jmbbm.2016.10.021.

Abstract

The Ponseti method is the accepted treatment of idiopathic clubfoot. Although the method of manipulating the baby feet is described in great detail, current study aimed to investigate the magnitude and course of the applied forces in order to optimise the treatment of clubfoot. An instrumented clubfoot model was constructed with force sensors on the location of the First Metatarsal (FM) and the Talar Neck (TN) and treated with the Ponseti method by 17 practitioners. Applied forces on FM and TN were measured during manipulation (4.2 N; 12 N), during casting (3.2 N; 3.5 N) and after casting (2.9 N; 2.2 N). The forces during manipulation were significantly higher than during casting on TN ($p < 0.001$) but not on FM ($p = 0.129$). No 'correct' amount of force could be determined and inter-practitioner variability was measured to be 70%. The resulting pressure of the cast on the clubfoot model as measured directly after casting was significantly higher than local tissue perfusion. The results of this study suggest potential for the optimisation of the application of the Ponseti method.

3.1 Introduction

An idiopathic clubfoot (talipes equinovarus) is a common congenital deformity of the foot and is usually treated with the Ponseti method (Ponseti & Smoley, 1963). In this method, the practitioner applies a manual force on the medial side of the First Metatarsal (FM) with a counter force on the lateral side of the Talar Neck (TN). This abducts the foot to realign the navicular bone with the talus. The manipulation stretches the tissues on the medial side of the foot and is maintained for a week with a plaster cast. Most cases of clubfoot are corrected after five or six cast changes and, in many cases, a percutaneous Achilles tenotomy (Ponseti, 2008). After the casting period, a foot abduction brace is used for several years to prevent relapse (Dobbs et al., 2004; Ponseti, 2008).

All steps of the Ponseti method are described in great detail and it seems that one of the only aspects that allows for a personal interpretation is the amount of stretch or force that is required to correct the clubfoot. In literature only subjective measures are given for this measure, like “*the ligaments should never be stretched beyond the natural amount of give*” (Ponseti, 2008), “*Use light pressure*” (Staheli, 2009) or “*gentle manipulation*” (Radler, 2013). Without any objective measures for the magnitude and course of these forces, it can be very difficult to obtain the craft (Frick, 2009). Experienced practitioners of the Ponseti method state that they feel how much force is safe to apply to a clubfoot to correct its position.

Mathematical models have been established for strain dependent stress relaxation of ligaments and tendons (Davis & De Vita, 2012). These indicate that the higher the initially applied force, the greater the adaptation rate of the tissue. One can imagine that a certain minimum amount of force is required to induce any correction at all, and that too much force will cause tissue damage. Looking at the casting period only, the most efficient treatment would have the shortest treatment duration (low burden on the family, low costs) with the least complica-

tions (skin problems, relapses, etc.). Therefore, it is reasonable to expect that for each clubfoot an optimum force exists. With time and experience practitioners learn to get close to this optimum force.

When applying a force, complete closure of the capillaries (ischemia) should be avoided. P. W. Brand and Hollister (1999) states that the point when only some of the capillaries are closed and the skin is pale with pink spots ('early blanching') is the safe limit. The magnitude of this pressure resembles the diastolic blood pressure. For adults this is 70 to 80 mmHg or about 1.0 N/cm^2 (Marieb & Hoehn, 2007). The thicker the layer of soft tissue over the underlying bone, the higher the allowed pressure is before the point of blanching is reached. The lower the pressure, the longer it can be tolerated. To be on the safe side, an optimal pressure of 0.50 N/cm^2 is extrapolated from animal experiments for finger orthoses in the treatment of finger contractures (P. W. Brand & Hollister, 1999). A similar situation occurs in the Ponseti method where the cast applies a pressure on the clubfoot. Children who undergo clubfoot correction may, due to their age, be expected to have a reduced diastolic blood pressure of 45 mmHg or 0.60 N/cm^2 (Kent, Kecskes, Shadbolt, & Falk, 2007). The Ponseti casts are well tolerated by the children (Baindurashvili et al., 2012), thus suggesting that sustained pressure of the cast is below the safe limit.

During casting, there are a few layers of padding and cast tape between the clubfoot and the fingers of the practitioner. We therefore suspected that keeping the force focussed at the exact location of the FM and TN would be challenging, which would lower the measured force during casting.

The aim of the present study is to quantify the applied forces in the Ponseti method in order to find room for optimization of the treatment of clubfoot. We stated the following two hypotheses about the Ponseti method:

1. *The forces during manipulation are higher than during casting*
2. *The cast applies a pressure on the foot below local tissue perfusion*

3.2 Materials and methods

In this descriptive study, the Ponseti method was executed on an instrumented clubfoot model, and forces on two locations (FM and TN) were measured as the primary outcome measures. It was conducted at multiple Dutch hospitals that are acknowledged as clubfoot centres and at one British hospital. This study should give information on the applied forces on these locations during manipulation, during casting and after casting.

Seventeen experienced practitioners of the Ponseti method were asked to participate in this research. Only practitioners who had been using the Ponseti method for more than one year and who treated at least 10 clubfeet per year were included.

3.2.1 Instrumented clubfoot model

To investigate the applied forces, all participants used the same standardised clubfoot model. This model (FIGURE 3.1) was derived from the first phase clubfoot training model (*MD Orthopaedics*, Inc., Wayland, IA, USA). The instrumented model was made from silicone rubber with stiffness 33 ± 2 ShA (HT33, *Zhermack*, Badia Polesine, RO, Italy). Two coronary stent balloons ($\varnothing$ 3.5mm, $\leftrightarrow$ 10mm, *Blue Medical*, Helmond, the Netherlands) filled with demineralised water were embedded at the location of the FM and TN and connected to two intravenous blood pressure sensors (501669001, *Merit OEM*, Maastricht, the Netherlands). Communication with the computer was performed with an Arduino system.

3.2.2 Calibration

The sensors give a measure for the pressure in the coronary stent balloons. These values were used to measure the *pressure* from the cast on the ‘skin’ of the rubber clubfoot model (second hypothesis). To relate this to applied *force*, the instru-

mented clubfoot model was calibrated with a weighing scale. A perpendicular force (0 to 30 N in steps of 5 N) was applied manually on the location of FM and TN. By doing so the sensors showed a resolution $\pm$ error of 0.048 ± 0.002 N / LSB and 0.136 ± 0.005 N / LSB respectively (LSB = Least Significant Bit). Bending the foot compresses the balloon in the ankle joint (TN) which results in an internal pressure. This influence was analysed and the force at TN could be corrected by subtracting $1.01\pm0.02 \cdot F_{FM}$ from F_{TN} .

3.2.3 Measurement protocol

The practitioners were given information on the nature of the study and were asked to provide consent. They were asked to treat the clubfoot model with the Ponseti method. No instructions were given whether to address the cavus, adductus or varus. A few moments were given to allow practitioners to get a good feeling of the foot before the measurements began. The clubfoot model was manipulated and held in its corrected position for five seconds. The mean value over this interval was determined. This process was repeated five times and the mean force was calculated over these five values.

Next, the clubfoot was casted following the Ponseti method. During casting the mean force was calculated over the period when all plaster was applied and the position of the hands was fairly stable.

The mean force after casting was defined as the first 2.5 seconds (50 data points) after the practitioner released the foot (see highlighted areas in FIGURE 3.2).

3.2.4 Data collection

Force data were collected during the five manipulations and during casting of the clubfoot, recording up until the point that the cast was removed from the

clubfoot model. All data were processed with Microsoft Excel 2013 and SPSS 20.

3.2.5 Statistical tests

Shapiro-Wilk normality tests were performed and showed lack of normality only for FM after casting. Therefore, mean and range were reported. Significance was set at $\alpha < 0.05$ for all statistical tests. Forces at FM and TN during manipulation and during casting were compared using Student *t*-tests. 95% Confidence Intervals were used to compare pressure values with hypothesised values. The coefficient of variation (CV) was used as a measure for inter-practitioner variability.

3.3 Results

During manipulation one practitioner did not press on the TN but rather on the lateral malleolus and one practitioner pulled on the forefoot. During casting two practitioners pulled on the forefoot and one held the foot with his/her whole hand instead of using just two fingers. In these cases, the manipulation and casting was not in accordance with the guidelines of the Ponseti method and it resulted in the sensors not registering the correct force. Additionally, sensor malfunctioning was observed in the period after casting in one measurement. The data of these measurements (two during manipulation, three during casting and four after casting) were not included in the analysis. In two cases a force below zero was measured. Although this is an unrealistic value, there was no reason to exclude this data and it was treated as noise. This resulted in 15 successful measurements during manipulation, 14 during casting and 13 after casting.

The results of a typical force measurement are presented in FIGURE 3.2. Measured mean forces on FM and TN during manipulation (4.2 N (0.1 to 10); 12.5 N (1.6 to 27)), during casting (3.2 N (0.0 to 4.9); 3.5 N (-2.3 to 7.1)) and

Figure 3.1 | Photo of the instrumented clubfoot model. Note the two coronary stent balloons inside the rubber clubfoot model.

after casting (2.9 N (-1.4 to 4.5); 2.2 N (-3.8 to 5.7)) are presented in Figure 3. The measured force on TN was significantly higher during manipulation than during casting ($p < 0.001$), but not for FM ($p = 0.129$).

After casting, the remaining pressures on the FM and TN were 0.89 N/cm^2 (95% CI: 0.59–1.2) and 1.7 N/cm^2 (95% CI: 0.64–2.7) respectively.

Coefficients of variability (CV) showing inter-practitioner variability for applied force during manipulation, during casting and after casting are presented in TABLE 3.1.

Figure 3.2 | Typical force measurement during the application of the plaster cast. The grey areas represent the periods over which the mean was calculated. Note that a filter was used for graphical purposes only. (FM = First Metatarsal, TN = Talar Neck)

Phase	FM	TN
Manipulation	67%	63%
During casting	45%	59%
After casting	61%	112%
Overall average	70%	

Table 3.1 | Coefficients of variability for FM (First Metatarsal) and TN (Talar Neck)

Figure 3.3 | Applied force on the instrumented clubfoot model. (FM = First Metatarsal, TN = Talar Neck)

3.4 Discussion

A clubfoot is usually treated with the Ponseti method: manipulations and serial casting. Little is known about the magnitude and course of the applied forces in the treatment of clubfoot. This study aimed to quantify the applied forces in the Ponseti method in order to allow for possible optimisation of the treatment of clubfoot.

3.4.1 Optimum force

Considering that in this study all practitioners were experienced with the Ponseti method and each practitioner was presented with exactly the same clubfoot model, it is remarkable to observe such a great inter-practitioner variability (FIGURE 3.4). In practice this variability is even greater due to the fact that each

Figure 3.4 | Distribution of the applied forces on the instrumented clubfoot model. Dots represent data from one practitioner, crosses the mean value for that part of the treatment (manipulation, during casting and after casting). (FM = First Metatarsal, TN = Talar Neck)

Figure 3.5 | Comparison of both clubfoot training models from *MD Orthopaedics*. The first (left) model was used for current study while Cohen, Altioik, Tarima, Smith, and Harris (2012) used the second model (right).

treated foot is unique.

The Ponseti method is taught without objective measures for the required amount of force, while its success rate is reported to be around 90% (Jowett, Morcuende, & Ramachandran, 2011). There is no evidence of a different success rate for the practitioners who participated in current study and therefore it can be assumed that almost all measured values in this study would result in effective correction of the clubfoot. In other words, all force values apparently fall within the acceptable range which is required for correction. However, the treatment outcome (complications, number of casts, etc.) might vary between the practitioners. If the applied force could be related to treatment outcome, an optimum force could possibly be determined which would help in educating new Ponseti practitioners.

3.4.2 Applied force

Concerning the first hypothesis, *the forces during manipulation are higher than during casting*, this was confirmed for TN ($p < 0.001$) but not for FM ($p = 0.129$). The lower force on TN during casting might be explained by the difficulty of palpating TN and FM. Clinical experience shows that especially for TN it might be difficult to find the correct location underneath a few layers of padding and plaster cast. If the practitioner does not press on the correct location, a lower force will be measured, but this would also result in a greater CV. However, for TN CV is very similar during casting (63% vs. 69%). Therefore, it is expected that the cause for the lower force during casting can be found in the padding rather than difficulty of palpation. The padding disperses the applied pressure over a greater area, resulting in a decreased measured force at the location of the sensor.

In their study of creep activity of plaster cast (Cohen et al., 2012) also measured the forces on a clubfoot model during manipulation. Two orthopaedic

surgeons were asked to manipulate the model to correct the adductus while the force on FM was measured with a force scale. They found the force on FM to be 15 ± 3 N, much higher than the measured value of current study (4.2 N). While in current study the first phase clubfoot model was used, (Cohen et al., 2012) used the slightly bigger second model in which the cavus had already been corrected (see FIGURE 3.5). While no instructions were given as to how to treat the clubfoot model, two practitioners mentioned that, in a clubfoot as presented to them in the clubfoot model, only the cavus should be addressed. The adductus and varus should be addressed in successive casts. The Ponseti method (Ponseti, 2008; Staheli, 2009) stresses the importance of the correct order of cavus, adductus, varus and then equinus, but is not very clear about whether or not cavus and adductus can be addressed simultaneously in the first cast. In practice, depending on the flexibility of the feet, it appears to be possible to address cavus and adductus simultaneously.

3.4.3 Adaptation to the applied force

Our second hypothesis stated: *the cast applies a pressure on the foot below local tissue perfusion*. The hypothesis is refuted. After casting, the pressure of the cast on the clubfoot model was significantly higher than the hypothesised 0.50 N/cm^2 . With the measured values exceeding the threshold value, it would be expected that (even if casting was adequate and carefully applied) pressure marks on the skin or even damage the tissue due to ischemia would occur, but this is rarely observed: (Baindurashvili et al., 2012) observed soft tissue lesions in the areas where pressure is applied as specific feature of the treatment method in only 6% of cases. In current study practitioners wore gloves or had their hands covered with plaster and therefore it could not be verified whether their fingertips turned pale during casting.

One explanation could be that the pressure drops below a safe margin before the cast can do any damage due to the beneficial adaptive properties of the tissue

of the new born. MRI studies have shown that several minutes after casting the cartilage anlagen of the tarsal bones have adapted (R. Brand, Siegler, Pirani, Morrison, & Udupa, 2006). Studies with accelerated Ponseti methods show that the foot adapts well within one week (Elgohary & Abulsaad, 2015; Gilani et al., 2014; Harnett et al., 2011; R. J. Xu, 2011). Previous work by one of the co-authors (P.G.M.M.) suggests an adaptation rate of just a few hours (Maathuis, de Jong, Hekman, Bulstra, & Verkerke, 2009). We have planned a next study that aims to measure how fast the clubfoot tissue adapts to the static loads of serial casting.

Another explanation is that the safe pressure limit of 0.50 N/cm^2 is based on old experiments with paralyzed dogs (Reswick & Rogers, 1976). How this translates to pressures on a baby foot is unknown. Although P. W. Brand and Hollister (1999) has found that pressures on the finger below this limit were safe, it was not tested but merely assumed that pressures above this limit would result in skin problems. Above all, the safe pressure limit might be different for every person Coleman et al. (2014).

Both time and force influence the efficacy of serial casting (Cyr & Ross, 1998) and thus the Ponseti method. One theory combining both factors looks at the treatment dose, the amount of force times the duration of application (McClure, Blackburn, & Dusold, 1994). This can be expressed as dose per week (e.g. for knee contractures a minimum treatment dose is suggested to be $11000 \text{ Nm}\cdot\text{min}/\text{wk}$ (Uhl & Jacobs, 2012)). To the best of our knowledge such data does not exist for the treatment of clubfoot, but with our planned next study it might be possible to measure the treatment dose of the Ponseti method. An objective measure for the optimum time and force would help in optimising the treatment of clubfoot.

3.4.4 Limitations of the study

Applied forces on the FM and TN of an instrumented clubfoot model were measured during manipulation, during casting and after casting. How well the measured values represent forces on a real clubfoot depends on some limitations to the used measurement device.

Firstly, the use of coronary stent balloons created an elegant and robust measurement instrument but forces are not directly measured. They are derived from measured pressure which is converted to a force based on a calibration with known forces. Some error may be introduced if the area on which the force is applied by the practitioner is different from the area used in the calibration procedure.

A second source of error may be that the practitioner applies force in a location or direction which slightly deviates from the calibration procedure. Since the Ponseti method is very clear about location and direction of force we assumed the measured force to be correct. However, with three practitioners, an incorrect technique was clearly visible.

Thirdly, the used clubfoot model is not a real foot. It does not have baby fat, bones or skin, nor does it inhibit short-term viscoelastic behaviour. However, it did give the opportunity to present every practitioner with the same clinical case.

Practitioners mentioned that the clubfoot model was rather stiff, giving the feeling of a syndromic or plain unrealistic foot. However, the influence of stiffness can be assumed to be minimal if it is force (rather than deflection) that is leading, and the observed results are representative of normal treatment. Indeed most practitioners said it was force and some even closed their eyes temporarily during the manipulation to get a better feeling for the applied force. We therefore assume that for this purpose, the instrumented clubfoot model is a realistic simulation and the measured forces give a good first indication for the actual applied forces during treatment.

3.5 Conclusion

The current study aimed to quantify the used forces in the Ponseti method with an instrumented standardised clubfoot model. Large differences between practitioners were observed with an average inter-practitioner variability of 70%. The measured forces on TN were significantly higher during manipulation than during casting but this was not found for FM. The resulting pressures of the cast on the skin were above local tissue perfusion while skin problems are rarely observed in the Ponseti method. Even when presented with one specific case, practitioners display a large variation in applied corrective forces. This result indicates potential for optimising the manner in which the Ponseti method is practiced.

Chapter 4

Design and Test of a new Inductive Force Sensor

BOB GIESBERTS, VICTOR SLUITER AND BART VERKERKE

This chapter is published as R.B. Giesberts, V.I. Sluiter, and G.J. Verkerke (2018) Design and Test of a new Inductive Force Sensor, *Sensors*, 18(7), 2079. doi: 10.3390/s18072079.

Abstract

The currently accepted interval of weekly cast changes in the treatment of clubfeet seems unsubstantiated. A force sensor is needed to determine the adaptation rate of a clubfoot to establish what cast change interval would be most effective and efficient. We developed a force sensor based on the principle that the resonance frequency of an LC-tank changes when a metal target is brought in close proximity. A thin rubber ring between the LC-tank and the metal target transformed this proximity sensor into a force sensor. With a Static Load Test and an Incremental Load Test the performance of the constructed force sensors was characterized. The custom made sensor showed excellent sensitivity ($(1.7 \pm 0.8 \cdot 10^5)$ counts/N), resolution ((0.15 ± 0.06) mN) and accuracy ((3.5 ± 3.0) %) for the application. The observed drift was (2.1 ± 0.7) %/ $\log_{10}(\text{hr})$, which is lower than other thin force sensors. Preliminary results of measurements in the treatment of Dupuytren fingers and clubfeet show good functioning for long-term force measurements.

4.1 Introduction

A clubfoot is a congenital deformity of the foot which is commonly treated with the Ponseti method (Ponseti, 2008). In this treatment, in the first weeks after birth the foot is manipulated towards a more corrected position and fixated with a plaster cast (FIGURE 4.1A). This weekly procedure is repeated several times until the deformity is fully corrected. At first the foot resists the newly imposed position by pressing against the rigid plaster cast, but over time the biological tissues inside the clubfoot adapt to the new position: the force against the plaster cast decreases (FIGURE 4.1B). The adaptation rate of a clubfoot can be determined by measuring the decrease of force between the cast and the clubfoot. This information is essential to improve the treatment method because it can determine when the next correction is due. This could be much faster than after one week.

To measure the adaptation rate of the clubfoot a force sensor is needed with very specific requirements. The specific application that is described above re-

Figure 4.1 | Sensor location. a The clubfoot is fixated in a corrected position with a plaster cast. b Initially the foot presses against the cast but over time this force decreases due to adaptation of the biological tissues.

Table 4.1 | Sensor requirements. A small Lipo battery (165 mAh) should be able to power the system for at least one week (168 h), hence the set requirement for the maximum power consumption.

Requirements		Value	Unit
Dimensions	max	3×10 (ϕ)	mm
Dynamic range	min	0–10	N
Resolution	max	0.01	N
Time drift	max	1	% / day
Bandwidth	min	0.1	Hz
Current consumption	max	1.0	mA @ 3.7 V

quires a sensor that is thin enough to fit underneath the cast without causing any damage to the skin, accurate enough for a measurement over the period of one week and its power should be supplied by a small battery that would be safe in a portable device for a baby. The requirements used for the development of the force sensor described in this paper are summarized in TABLE 4.1.

Numerous force sensors with varying functional principles are commercially available, but, to the best of our knowledge, none of them are suitable for long term precision measurements within a portable setting. Other sensors do not fit our requirements because they are either too bulky or experience too much drift. OEM load cells (e.g. *Futek* Inc., Irvine, CA, USA) are too large for our requirements and it would be challenging to make a low-power amplifier with high resolution for load cells based on strain gauges. Force Sensitive Resistors (FSR, e.g. *FlexiForce*, *Tekscan*, Boston, MA, USA) are susceptible to drift (Hollinger & Wanderley, 2006; Khodasevych, Parmar, & Troynikov, 2017; Parmar, Khodasevych, & Troynikov, 2017). OptoForce sensors (*OptoForce*, Budapest, Hungary) are elegant yet not thin enough to be placed below the plaster cast. TakkTile sensors (*RightHand Robotics*, Somerville, MA, USA) are a relatively good candidate but we found that their measurement range is difficult to adjust and their

accuracy over long term measurements is unknown.

The aim of this paper is to describe the performance and accuracy of a new inductive force sensor for the application of long-term force measurements in a portable setting.

4.2 Sensor design

An inductive force sensor was developed based on the evaluation kit of the inductance to digital converter LDC1000 (*Texas Instruments*, Dallas, TX, USA) and its successor LDC1614 (*Texas Instruments*, 2014). This integrated circuit is able to accurately determine the resonance frequency (f_{sensor}) of an LC-tank. Its main output parameter is LHR_DATA, a digital value that is proportional to frequency, with 28 bits of measurement resolution. The resonance frequency is influenced by the distance of the coil to any metal target. In most applications a non-ferromagnetic and conductive metal (e.g. Aluminum) is used to modulate the inductance of the coil. Closer proximity of the target to the coil will increase Eddy currents in the target, thereby decreasing the inductance of the coil. In close proximity the relation between distance and resonating frequency can be linearized with small errors. The exact functioning and examples for sensor design can be found in documentation provided by Texas Instruments (e.g. Kasemsadeh & LaPointe, 2016; Oberhauser, 2015).

To convert this proximity sensor into a force sensor, an elastic medium with a known stiffness was added between the coil and the target. Our initial approach consisted of a simple sandwich of these three components. Although this solution has been researched by others (Wang et al., 2017), and patented by others (a.o. Bonifas, Kim, & Suh, 2015; Zinober & Has, 2014), it did not appear as commercial product at the time of writing of this manuscript. Preliminary tests with this approach showed a high accuracy and sensitivity but mediocre long-term results with values for time drift as high as 50 %/day. Several tests were done to identify

the cause for this drift. If the sensor was not loaded with any weight, no drift was observed. Creep of the rubber medium was rejected with a creep test under 50 % compression for 14 days which showed no deformation. Replacing the compressible medium by an equally thin sheet of (incompressible) glass did not resolve the drift. Neither did replacing the aluminum target for a polyethylene target. One of the few hypotheses remaining was plastic deformation of the PCB.

In the final design a base-plate was added and the rubber medium was shaped into a 'ring', enclosing the PCB (see FIGURE 4.2 and FIGURE 4.3). With this approach loading the sensor compresses the rubber ring around the PCB but prevents loading – and possibly deforming – the PCB. With the rubber ring this solution is a compromise on the size ($\phi 10$ mm vs $\phi 7$ mm). A DS1825 thermometer (Maxim Integrated, 2005) was added to correct for the influence of temperature.

The terminology used for the sensor, the Acquisition Unit (AU) and their combination as Sensing and Logging Unit (SLU) throughout this paper are presented in FIGURE 4.2.

4.2.1 Acquisition unit

The sensor was connected to a custom made acquisition unit (AU) to log LHR.DATA over time. The AU was equipped with the LDC1614 (Texas Instruments, 2014) and a KL25Z microprocessor (NXP, 2014) that was programmed to write the collected data to a microSD memory card. The AU was powered with a small single cell Lipo battery (165 mAh).

Programming of the AU was optimized for high sensitivity, low power consumption and little time drift. This means that for the internal settings of the LDC1614, CHx_RCOUNT was set to maximum (0xFFFF) for increased precision at the expense of bandwidth, SETTLECOUNT_CHx to 50 (to allow a sensor settling time of $t_{settle} = 20$ ms), and CHx_DRIVE was automatically determined using the method described by Holubeva (2015) and set to 23. The resulting sample rate

Figure 4.2 | Two Sensors and an Acquisition Unit. The acquisition unit (top) allows the connection of two sensors (below) and a thermometer (right). Schematic representation of a sensor and logger to explain the distinction between sensor, Acquisition Unit (AU) and the combination of the two: the Sensing and Logging Unit (SLU).

Figure 4.3 | Sensor specifications. Front and top view of the sensor in four stages of the assembly process. The sensor is build out of four parts of which the technical specifications are given here.

was 18 Hz. The used code is available online (Giesberts, 2016).

4.3 Experimental section

4.3.1 Data collection

To reduce energy consumption, the AU was programmed to write packages of 60 data samples at a time to the microSD card, rather than sample-per-sample. The thermometer was programmed to do one temperature measurement per package.

4.3.2 Test setup

A calibration rig was fabricated which allowed us to hang known weights on the sensor. The applied load (F , in N) compresses the sensor, thereby changing the oscillation frequency. `LHR_DATA`, proportional to this frequency, is measured and stored by the AU. None of the parts used in close proximity of the sensor were made of metal to prevent any interference with the LC-tank. A hanger of 127g was used as a pre-load to which weights of 100g each were attached (see FIGURE 4.4).

4.3.3 Static Load Test

A static load test was performed to carefully assess time drift of the Sensor and Logging Unit (SLU). The SLU was placed in the calibration rig with the pre-load in place. During the first 2 hours, the force sensor was not loaded, then a weight of 500g was hung on the force sensor and left there for at least 12 hours after which the load was removed. Measurements continued for another 2 hours.

The following characteristics were determined:

Figure 4.4 | Calibration rig. Each sensor was placed in the calibration rig. Steel disks of 100g each (bottom left) were placed on the hanger to apply a load.

Noise was defined as the standard deviation of the output signal (LHR.DATA), filtered with a high-pass filter (0.01 Hz). Furthermore, the noise performance was characterized by measuring the Allan deviation (Riley, 2008, p.15) as a function of averaging time using a sensor that was not loaded for eight hours.

Temperature drift of the SLU was defined as the change in LHR.DATA per change of the room temperature (heating of the empty office used for our tests was controlled centrally and caused fluctuations in temperature). It was calculated as $\Delta \text{LHR.DATA} / \Delta T$ over the period while the sensor was loaded and the input parameter (F) was constant, in counts/ $^{\circ}\text{C}$.

Time drift of the SLU was defined as the relative change of LHR.DATA over time, after a 30sec settling time. Two calculation methods were used:

- Drift was calculated as the gain of the best linear fit of the signal over

logarithmic time, in $\%/ \log(\text{hr})$ (Ferguson-Pell, Hagisawa, & Bain, 2000).

- Drift was calculated as the relative difference between the initial value and the final value after several specific times (Parmar et al., 2017).

4.3.4 Incremental Load Test

The SLU was loaded with incremental 30sec-long steps of 100g each up to a maximum of 1kg. With this test the following characteristics were determined:

Sensitivity of the SLU was defined as the change of the output signal per change of input parameter ($\Delta \text{LHR_DATA} / \Delta F$) (Carr, 1993). It was determined as the gain of a best linear fit in counts/N. *Non-linearity* was defined as the greatest deviation to this linear fit (Carr, 1993).

Accuracy of the SLU was defined as the greatest deviation to a polynomial fit that was used to allow conversion of LHR_DATA to force, as percentage of the full scale output.

Hysteresis of the SLU was defined as the maximum difference between the loading and unloading curves, as percentage of the full scale output.

4.3.5 Data processing

The collected data were processed using Matlab R2017b. In post-processing the LHR_DATA for the Static Load Tests were corrected for an apparent start-up effect and for the influence of temperature changes using the identified temperature drift.

Start-up effect

In its intended application in force measurements in the treatment of clubfoot the SLU uses a logging protocol of four 30sec periods per hour to preserve battery

Figure 4.5 | Start-up effect. The output signal gradually decreases every time the SLU is started. For this graph the SLU was programmed to take ten 150sec measurements per hour, presented in grey. For all timestamps the mean was calculated (darkgrey) and least squares regression was used to find the best fit to a double exponential function (orange). The typical sensitivity (TABLE 4.2) can be used to convert LHR_DATA to a measure for force: 1500 counts approximate 0.01 N.

power. Every time the SLU is activated an initial start-up effect is observed, as shown in FIGURE 4.5. Seemingly independent of external influences the output signal consistently decreases over time to reach a plateau after several minutes. This behavior was observed in all SLUs with almost identical parameters. For all sensors this behavior was identified, quantified and corrected.

4.4 Results

The Static Load Test was performed with nine sensors and the Incremental Load Test with thirteen. Typical examples of the resulting data of both tests are presented in FIGURE 4.6 and FIGURE 4.7 respectively. The performance of the

Table 4.2 | Sensor performance. All characteristics are presented as mean $\pm$ standard deviation.

	Static Load Test	Incremental Load Test
	(n = 9)	(n = 13)
Sensitivity (10^5 counts / N)	1.7 $\pm$ 0.8	1.3 $\pm$ 0.6
Non-linearity (%)		12 $\pm$ 10
Accuracy (%)		3.5 $\pm$ 3.0
Hysteresis (%)	5.8 $\pm$ 1.4	4.9 $\pm$ 2.1
Noise (10^{-3} N)	1.2 $\pm$ 0.4	
Effective Resolution (10^{-3} N)	0.15 $\pm$ 0.06	
Time drift (%/log ₁₀ (hr))	2.1 $\pm$ 0.7	
Time drift after 8hrs (%)	5.0 $\pm$ 1.4	
Temperature drift (N/°C)	-0.088 $\pm$ 0.030	

inductive force sensors are summarized in TABLE 4.2 and described in more detail below.

The influence of temperature was (-0.088 ± 0.030) N/°C ($R^2 = 0.84$). The results of a typical temperature analysis are shown in FIGURE 4.8.

The time drift of the SLU was determined to be (2.1 ± 0.7) %/log₁₀(hr). FIGURE 4.9 shows the observed drift during a typical incremental load measurement.

The noise characteristics are presented in the Allan deviation graph in FIGURE 4.10. It shows that the contribution of white noise is relatively small and that the sensor response is dominated by drift.

4.4.1 Total error

The greatest sources of error are hysteresis (both ways, ± 2.5 %) and drift (± 2.1 %/log₁₀(hr)). When the output signal is corrected for the start-up effect and for the influence of

Figure 4.6 | Static Load Test. The results of a typical Static Load Test. The typical sensitivity (TABLE 4.2) can be used to convert LHR_DATA to a measure for force: 1500 counts approximate 0.01 N.

Figure 4.7 | Incremental Load Test. The results of a typical Incremental Load Test.

Figure 4.8 | Temperature influence. About 85 % of the changes in output signal can be explained by fluctuations in temperature. It is not clear if this temperature influence is independent of the observed time drift and the load applied. The typical sensitivity (TABLE 4.2) can be used to convert LHR_DATA to a measure for force: 1500 counts approximate 0.01 N.

Figure 4.9 | Time drift. The results of a typical static load test, during the loaded period, with time on the x-axis on a logarithmic scale. **Orange** dots represent the observed drift. A best linear fit and its 95% prediction interval is presented as a **blue** line. The **blue** dots represent the drift at specific times, as per Parmar et al. (2017)'s method.

Figure 4.10 | Allan deviation The brief decline at the onset of the curve indicates that the contribution of white noise is relatively small. The upward slope shows that the sensor response is dominated by drift. It can be concluded that the optimal averaging time is around 2 seconds.

temperature, the combined error is approximately $\pm 5\%$ of the full scale output. Depending on the application the actual error might be higher due to influences that were not present in the tests used in this study.

4.5 Discussion

4.5.1 External influences

A negative linear relationship was observed between a change in temperature and the output signal. This means that with an temperature increase, the output signal suggests a larger distance between the LC-tank and the aluminum target. Thermal expansion of the rubber medium might be a cause for the observed temperature drift.

Most of the fluctuations of the sensor signal under a constant load could be explained by changes in temperature ($R^2 > 0.84$) and the signal was corrected for this influence. However, some residual fluctuations could be observed — other than noise or drift — with an unknown source. These fluctuations might be caused by higher order effects of temperature or by any other external influence, but this was not studied extensively.

4.5.2 Effective resolution

The `LHR_DATA` consists of 28-bit numbers, but the resolution of the full output range is closer to 23–24 bit. This is due to 20–22 count differences between possible sample values. This quantization fits the documentation of Oberhauser (2016), which predicts a LDC resolution of 3 Hz (corresponding to ~ 20 counts in our reading) when an external oscillator of 40 MHz is used. As the sensitivity is specified using `LHR_DATA`, the effective resolution is 20–22 times the reciprocal of the sensitivity.

The sensors were calibrated over a range of 10 N, which generated an output signal covering approximately 1% of the 28-bit LDC range. This means that over the relevant force used, the sensor signal spans a 16-bit range. In practice, the sensor is able to detect changes in load in the range of tens of mN.

4.5.3 Drift

Although drift is a well-known phenomenon, different methods exist for its quantification. Because of its apparent randomness, it can be difficult to give a clear figure for drift. It is presented as a percentage after a specific time (Khodasevych et al., 2017; Parmar et al., 2017) or as a percentage per logarithmic hour (Ferguson-Pell et al., 2000; Interlink Electronics, n.d.; Tekscan, 2012).

Reports on force sensors with similar designs do not present a characterization of drift (Wang et al., 2017) and did not allow for any meaningful comparison, so we have compared the observed drift in our sensor to that of commonly used force sensors. We chose Parmar et al. (2017)’s definitions and methods, who evaluated several force sensitive resistors (FSR) and applied 72.7 mmHg to their sensors. After 8 hours the signal had drifted with 10–25% compared to the value after an initial settling time of 26 seconds. Using the same protocol Khodasevych et al. (2017) found 0.4–24% drift after 8 hours for more FSRs. In our Static Load Test a weight of 500g was placed on the sensor with 10 mm diameter, equal to 74.6 mmHg. After 8 hours, our sensors showed 5.0 ± 1.4 % drift.

The chosen *settling time* has a large influence on the figure for drift. Every sensor has some transient behavior before reaching its steady state, but the exact end of the transient behavior is hard to find and might be after several hours. The choice for a settling time of 26sec (Parmar et al., 2017) or 30sec (our study) is arbitrary and creates a risk of bias. In most cases drift is not purely linear but rather logarithmic, so the choice for a longer settling time generates a lower value for drift. Additionally, a longer settling time creates a higher reference value (P_i

in Parmar et al. (2017)), which in turn also lowers the value for drift.

FIGURE 4.7 shows an apparent time response (whether it be called response time or drift) after every change in load in the Incremental Load Test. While other sensors might have a clear distinction between response time and drift, our sensor does not: these characteristics rather seem to overlap. Due to drift of the sensor signal, it is impossible to determine a stable ‘end’ value without choosing an arbitrary time span. It is therefore also not possible to determine how long it takes to reach this value. With the drift present, any calculation of response time would be influenced by the choice for an arbitrary time span and is therefore omitted in this paper. Our choice for a settling time of 30sec is arbitrary and of pragmatic nature.

Any quantification of drift should be practical, and presenting it as a percentage per logarithmic time gives a better idea about the error that has to be taken into account when studying the sensor data. According to the datasheets of various FSRs (Interlink Electronics, n.d.; Tekscan, 2012), their drift is $< 5 \text{ \%}/\log_{10}(\text{hr})$. Using the same calculation method, our sensor shows a logarithmic drift of $(2.1 \pm 0.7) \text{ \%}/\log_{10}(\text{hr})$.

4.5.4 Reflection

For the specific application of long-term force measurements in the treatment of clubfoot, the performance of the developed force sensor fits most requirements. The sensor is thin enough to be placed underneath the cast without causing skin problems and the AU is neither bulky nor heavy. The observed drift is better than for most FSRs (Interlink Electronics, n.d.; Parmar et al., 2017; Tekscan, 2012) but slightly higher than required. In our application the force is expected to slowly decrease to a certain minimum level. In those measurements the error caused by drift might be limited since the observed drift appears to be load-dependent.

The influence of temperature can not be corrected perfectly and some residual

error can be expected. However, the temperature underneath a plaster cast is not expected to fluctuate much and by using an external thermometer most of the temperature influence can be corrected in post-processing. Hysteresis can not be corrected for since the actual load on the sensor is unknown and therefore the loading history can not be compensated for.

All sensors were assembled by hand which surely has introduced some variations in the amount of glue and subsequent distance between the LC-tank and the aluminum target. Any excessive glue might have created unforeseen connections between the rubber medium and the LC-tank, which is likely to have created variations in drift and hysteresis. Still the differences between the sensors tested were fairly small. A more controlled production method will decrease these differences and can be expected to improve the reproducibility of the performance of the sensors.

This sensor is very suitable for other biomedical applications where gradual changes of force should be measured in a small volume. Applications that come to mind are prosthetic socket fitting and prevention of pressure ulcers. For applications requiring a higher bandwidth, a sacrifice in power consumption should be made in the settings of the LDC1614. For applications requiring other force ranges, the stiffness of the medium may be changed.

4.5.5 Limitations

We did not have the means to perform the tests in an environment with a controlled temperature, barometric pressure and humidity. Instead we used the fluctuations in temperature to determine its influence and correct the signal for it. This may have introduced a confirmation bias. Additionally, the temperature influence was determined with a load of 500g and we did not test if this influence was load dependent.

During the Incremental Load Test, the maximum load (1kg) was usually ap-

plied longer than the other loads. Due to the time drift, this might have lead to an overestimation of the calculated hysteresis.

4.5.6 Sensor in practice

Preliminary results of measurements in the treatment of habitual toe walkers, Dupuytren fingers (CHAPTER 5) and clubfeet (CHAPTER 6) show excellent functioning for long term measurement. Encountered problems in these applications were of practical nature and were resolved with improved design: delamination of the aluminum target at the removal of the sensor was resolved by adding a post-it on top, battery failure was resolved by enhanced feedback with two LEDs on the AU and problems with identifying the start of important events in the logging was resolved by adding an input button on the AU.

4.6 Conclusions

A practical small-sized force sensor has been developed for long-term precision measurements in a portable setting. The used approach of inductive sensing combined with a compressible medium shows good applicability for the proposed force measurements between the skin and a cast, and great potential for many other applications.

Chapter 5

Tissue adaptation rate in the treatment of Dupuytren

5

BOB GIESBERTS, ANNE TER HAAR, MARTIJN SANDERMAN, EDSKO HEKMAN
AND BART VERKERKE

This chapter is accepted for publication (22-Sept 2018) as R.B. Giesberts, A. ter Haar, M. Sanderman, E.E.G. Hekman, and G.J. Verkerke (n.d.) Tissue adaptation rate in the treatment of Dupuytren contracture, *Journal of Hand Therapy*

Abstract

Study design: Basic research. (cross-sectional)

Introduction: Dupuytren disease can cause disabling contractures of the finger joints. After partial fasciectomy, postoperative hand splinting helps to maintain extension range of motion.

Purpose of the study: To measure how the contraction forces of the finger on the splint change over time.

Methods: Subjects that were treated for Dupuytren contracture with partial fasciectomy were invited to participate in this study. Force sensors were placed in their dorsal extension splint and the applied force was measured continually for several weeks.

Results: 11 Subjects (59–75 yrs old) with the MCP (8) or PIP (3) as their most severely affected finger joint participated. Each night, the measured force consistently decreased to reach a plateau after about three hours (adaptation time 2.55 (95% CI: 0.2–31.8) hrs). The time to reach this plateau decreased with time after surgery ($\approx 5\%$ /day, $p = 0.0005$, $R^2 = 0.08$).

Conclusions: The observed rate of decrease in the measured force indicates a tissue adaptation time of approximately three hours.

5.1 Introduction

Dupuytren is a genetic disease that affects the palmar aponeurosis, often resulting in contractures of the finger joints (Shaw, Chong, Zhang, Hentz, & Chang, 2007). A contracture causes a deficit in range of motion (ROM) for a specific joint. This can make ordinary tasks like a handshake or grasping a bottle of water very difficult, making it a real handicap for the patient (Foissac, Camuzard, Dumas, Dumontier, & Chignon-Sicard, 2013).

There are several accepted treatment methods for Dupuytren contracture, with collagenase injections (Gilpin et al., 2010) and needle fasciotomy (van Rijssen, Gerbrandy, Linden, Klip, & Werker, 2006) becoming popular. Many approaches of nonoperative treatment methods have been proposed and abandoned, and their application is still experimental (Brauns, Nuffel, Smet, & Degreef, 2017). Often partial fasciectomy is the choice of treatment, in which the affected tissue is surgically removed to loosen the contracture (Townley, Baker, Sheppard, & Grobbelaar, 2006). Postoperative physical exercises and wearing a hand splint at night may be prescribed to stimulate recovery and prevent new contractures (Herweijer, Dijkstra, Nicolai, & van der Sluis, 2007; Larson & Jerosch-Herold, 2008).

During the recovery process of the tissues in the hand, cross-links are formed to connect collagen fibres (de Morree, 2014). If these fibres are formed with the fingers flexed — the natural position of a relaxed hand — new contractures can form. The function of the hand splint is to help maintain the extension ROM during the recovery process and to help correct any residual contractures (Collis, Collocott, Hing, & Kelly, 2013). In this process, the longer the joint is held in its end range, the greater the gain in ROM (Flowers & LaStayo, 1994).

Because the splint prevents flexion of the fingers, it applies an initial force on the fingers, stretching the tissues. At first the tissues will resist and the initial force (F_0) can be relatively high. However, over time the tissues will adapt

(a.o. due to viscoelastic stress-relaxation) and the process of wound recovery will change, both causing the force to decrease. Since the tissues adapt to their imposed new position, it can be assumed that the end range of the finger joint also shifts. Due to this shift, after a certain amount of time the finger joint will no longer be in its end range and the force on the fingers reaches a plateau.

According to the Total End Range Time principle (TERT), the longer a joint is held in its end range, the greater the gains in ROM (Flowers & LaStayo, 1994). Several researchers have found that in the treatment of contractures greater gains in ROM are achieved with longer splint use, both in hours and in weeks (Flowers & LaStayo, 1994; Glasgow, Fleming, Tooth, & Hockey, 2012; Glasgow, Wilton, & Tooth, 2003). However, neither Jerosch-Herold et al. (2011), Kemler, Houpt, and van der Horst (2012) nor Collis et al. (2013) could find evidence for the efficacy of postoperative night splinting in addition to hand therapy when compared to hand therapy alone. Some patients find the splint uncomfortable and compliance is an issue (Chojnowski, Wach, & Degreef, 2017; Kemler et al., 2012; Rives, Gelberman, Smith, & Carney, 1992). In light of these recent findings the function of postoperative splinting has become unclear and postoperative splinting for the treatment of Dupuytren contractures has become a controversial topic (Chojnowski et al., 2017; Huisstede, Gladdines, Randsdorp, & Koes, 2017).

We investigated the process of tissue adaptation in the treatment of Dupuytren contractures on macro-level by measuring the forces in a hand splint. Force is the main outcome measure in this study. One can imagine that a certain threshold force exists below which the body does not respond. For example, the gravitational force that acts on one's arms does not significantly elongate them. The force at the end of the relaxation period (the final force, F_{∞}) will reflect this threshold value. Mechanically speaking, the tissue behaves as a "spring-damper system". For such a system, the resisting force $F(t)$ depends on the adaptation rate (τ , in hours) and both the initial and threshold force (F_0 and F_{∞} , in Newtons) as follows:

$$F(t) = (F_0 - F_\infty) \cdot e^{-\frac{t}{\tau}} + F_\infty \quad (5.1)$$

The aim of current study is to characterize the adaptation rate of the tissue to give a new perspective in the discussion on the efficacy of postoperative splinting.

5.2 Methods

This is a descriptive study conducted at a single health centre specialised in hand care (Handencentrum Enschede). The Medical Ethical Evaluation Committee of the Medisch Spectrum Twente (MST) reviewed the study according to the declaration of Helsinki, and declared that it does not meet the criteria as stated by the Medical Research Involving Human Subjects Act (WMO), and therefore did not require their approval.

5.2.1 Participants

Patients of the Handencentrum Enschede that were treated for Dupuytren contracture with an open partial fasciectomy were asked to volunteer in this study. Our postoperative treatment protocol included application of a night splint. Only fingers with extension deficit contractures that were treated with a dorsal hand splint were included. Patients with infected or severely swollen wounds or with skin grafts were excluded. If patients were simultaneously treated for multiple fingers, the finger with the largest deficit was monitored. All patients provided written informed consent.

5.2.2 Study variables

Baseline information was collected including age, gender, affected digit (middle, ring or little), affected joint (metacarpophalangeal (MCP), proximal interpha-

langeal (PIP), distal interphalangeal (DIP)), pre-operation (pre-op) active range of motion (AROM) and treatment history. Post-operation (post-op) AROM and force of the hand on the splint were measured in the weeks after surgery. Total active extension (TAE, defined as the sum of active MCP, PIP and DIP joint extension) and total active flexion (TAF, defined as the sum of active MCP, PIP and DIP joint flexion) were calculated.

5.2.3 Materials

A pocket finger goniometer (Exacta™, North Coast Medical, Inc., Gilroy, CA, USA) was used to measure AROM of the MCP and PIP joint in both extension and flexion. For the DIP joint a JAMAR® goniometer was used (Preston, Jackson, MI, USA). An accuracy of $\pm 5^\circ$ is mentioned for the measurement of finger joint angles (Ellis & Bruton, 2002).

Force was measured using a custom-made force sensor (FIGURE 5.1A), based on inductive sensing (Kasemsadeh & LaPointe, 2016). This sensor was specifically developed to overcome the difficulties with commercially available thin force sensors, like drift, size or costs. A digital DS1825 thermometer (Maxim Integrated, San Jose, CA, USA) was included to allow correction for temperature drift. Extensive information about the force sensor and its performance is available in a separate paper (Giesberts, Sluiter, & Verkerke, 2018, CHAPTER 4). The specifics are summarized in TABLE 5.1.

A data logger equipped with a small battery and microSD card was used to locally store the measurements of both the force sensor and the thermometer (see FIGURE 5.1A). The system was programmed to store data continually for periods of ten seconds, six times per hour, and can operate for at least seven days without requiring any interventions.

Table 5.1 | Force sensor characteristics. More detailed information about the custom made inductive force sensor is available in another publication (Giesberts, Sluiter, & Verkerke, 2018, CHAPTER 4).

Characteristic	Value
Dimensions ($\varnothing \times$ thickness)	10×2.3 mm
Resolution	0.15×10^{-3} N
Accuracy	≈ 3.4 %
Sample rate	18 Hz
Drift	< 2.1 % / $\log_{10}(\text{hr})$
Hysteresis	≈ 6.0 %
Temperature drift	≈ -0.1 N / $^{\circ}\text{C}$

5.2.4 Measurement protocol

Three to five days after surgery, a custom-made hand splint was fitted to the patient. Before placement, the force sensors were carefully calibrated using ten weights of 100g each. Then the force sensor and thermometer were placed on the inside of the splint, underneath the customary layer of foam, to prevent any direct contact with the skin. Depending on the most affected joint (MCP or PIP), the force sensor was located at the distal part of the metacarpal or the proximal phalanx (FIGURE 5.1). The data logger was attached on the outside of the splint with Velcro.

Following standard treatment, the patient was instructed to use the splint at night only and return weekly to have the therapist assess the finger and adjust the splint if necessary. In certain cases, additional daytime use of 3×1 hour per day was advised.

At every weekly visit the battery and memory card were replaced and the AROM of the affected finger was measured. After three weeks the force sensor was removed from the splint. Night-time splint use continued for several months

Figure 5.1 | Use of the force sensor. A custom made inductive force sensor was placed on the handsplint to measure the force of the hand on the splint. c Wires were routed distally and the data logger was attached to the dorsal side of the splint with Velcro.

and a short-term follow-up AROM measurement was planned three months after surgery.

5.2.5 Data analysis

Each retrieved data set consisted of time, force and temperature data of one week. Donning and doffing resulted in a sudden change in temperature and corresponding times were manually identified in post-processing. Each identified selection was defined as a *wear period*.

Wear periods

For every wear period the duration and the time since surgery were determined. Each wear period's force data was fitted to EQUATION (5.1) using Matlab 2016b's *fit* function. This function generates values and 95% confidence intervals (CI) for F_0 , τ and F_∞ . The used restrictions (bounds) for this fit function were $0.001 < \tau < 2000$ hrs and $F_\infty < F_0$. The *adaptation time* was defined as 3τ , which represents the required time to reach 95% of the equilibrium.

Wear periods were excluded for further analysis if the duration was shorter than half an hour, the goodness of fit (R^2) was less than 0.01, or the 95% CI for τ included negative numbers.

Statistical analysis

The distribution of extracted data values were checked for normality using Liliefors tests. Finger AROM differences between pre-op, post-op and follow-up were tested for significance using Mann-Whitney U-tests. Statistical significance was defined as $p < 0.05$.

The influence of the time since surgery on variations in duration, adaptation rate (τ), initial force (F_0) and final force (F_∞) were analyzed with a linear regression. For this regression the logarithm of the adaptation rate was used because it had a log-normal distribution. The result of that regression was transformed with

$$\begin{aligned} \log(\tau) &= A + B \cdot t \\ \tau &= e^{(A+B \cdot t)} \\ \tau &= e^A \cdot e^{B \cdot t} \\ \tau &= C \cdot (e^B)^t, \end{aligned}$$

in which A and B and the resulting coefficients form the linear regression, C is the initial adaptation rate and e^B the decay rate.

5.3 Results

5.3.1 Subject characteristics

After signing informed consent, initially 17 subjects were enrolled of which 6 were excluded due to severely swollen wounds or a suspected infection. A total of 11 subjects were included in the study. Subject characteristics are summarized in TABLE 5.2.

Table 5.2 | Subject characteristics. This table presents the specifics for the selected fingers only. In the case of multiple affected fingers and joints, the most severe one was selected. Treatment history is any previous surgery related to the Dupuytren disease.

Variable	Results
Age (mean (range))	67 (59–75) years old
Gender (M/F)	6 / 5
treatment history (yes/no)	5 / 6
Hand (L/R)	6 / 5
Finger (middle/ring/little)	0 / 5 / 6
Joint (MCP/PIP/DIP)	8 / 3 / 0

5.3.2 Active range of motion

The surgery promptly improved AROM of the MCP joint for extension ($p = 0.006$) and remained improved in follow-up as shown in TABLE 5.3 and FIGURE 5.2. Flexion AROM decreased after surgery but improved in the weeks after surgery.

Figure 5.2 | AROM change. (left and top) Average AROM measurements of all subjects combined. Before surgery (blue) a substantial extension deficit in the MCP joint could be observed. After surgery (gray) this deficit was almost completely removed although finger flexion was hindered. In follow-up after three months (orange) the range of motion of the finger was improved both in extension and in flexion. (bottom right) Post-op AROM measurements of all subjects combined, grouped per week after surgery. TAE: Total Active Extension, TAF: Total Active Flexion.

Table 5.3 | AROM. Active range of motions (AROM) in the finger joints. Data is presented as median (range) in degrees. Statistically significant AROM differences between pre-op, post-op and follow-up are presented in bold.

Joint		Pre-op	Post-Op	Follow-Up	Pre vs Post	Pre vs Follow	Post vs Follow
n		9	10	6	p	p	p
MCP	Extension	32 (-10– 58)	1 (-7– 22)	3 (-3– 18)	0.006	0.023	0.624
	flexion	85 (68– 92)	41 (16– 66)	85 (75– 89)	0.000	0.670	0.000
PIP	Extension	10 (1– 61)	8 (2– 43)	5 (0– 28)	0.734	0.368	0.471
	flexion	92 (80–101)	66 (26– 94)	85 (80–104)	0.002	0.711	0.016
DIP	Extension	-7 (-21– 0)	-10 (-24– 12)	-11 (-21– -5)	0.796	0.212	0.474
	flexion	51 (39– 69)	31 (6– 50)	59 (36– 71)	0.001	0.508	0.005
TAE		30 (14– 95)	-2 (-16– 77)	3 (-14– 13)	0.002	0.000	0.654
TAF		224 (214–241)	135 (48–194)	228 (204–258)	0.000	0.796	0.000

5.3.3 Splint forces

With the exception of one subject in which the sensor failed after two days, the force of the affected finger on the splint was measured for several weeks (range 10–24 days). A typical measurement is shown in FIGURE 5.3. Of the 210 identified wear periods, 153 met the criteria stated in 5.2.5 and were included for further analysis. The extracted data from these wear periods is presented in TABLE 5.4. The force decreased over time with a mean adaptation rate of 0.85 hours, hence the adaptation time was 2.6 hours.

5.3.4 Influence of time since operation

With every day, the adaptation time decreased somewhat ($\approx 5\%$ /day, $p = 0.0005$, $R^2 = 0.08$) (see FIGURE 5.4). Similar trends were found for duration (-5 min/day, $p = 0.03$, $R^2 = 0.03$), initial force (-0.01 N/day, $p = 0.003$, $R^2 = 0.06$) and final force (-0.007 N/day, $p = 0.04$, $R^2 = 0.03$).

Table 5.4 | Wear periods. The values for adaptation rate are log-normally distributed (Liliefors, $p = 0.09$), so the transformed mean and 95% CI are given. The values for duration, F_0 and F_∞ are not normally distributed ($p = 0.002$, $p = 0.001$ and $p = 0.001$ respectively), so the median and interquartile ranges (IQR) are given.

Measure	Results	
N	153	
Duration	7.3	(IQR: 3.7–9.0) hrs
Adaptation rate (τ)	0.85	(95% CI: 0.1–10.6) hrs
Adaptation time (3τ)	2.55	(95% CI: 0.2–31.8) hrs
Initial force (F_0)	0.07	(IQR: 0.01–0.29) N
Final force (F_∞)	−0.08	(IQR: −0.31–0.01) N

5.4 Discussion

The extension deficit caused by the effects of Dupuytren were successfully treated in all subjects. Although MCP extension had significantly improved after surgery, flexion AROM temporarily worsened, most probably due to swelling and sensitivity of the wound. The follow-up AROM measurements showed that after three months the improvement of finger extension was effectively maintained and that finger flexion was regained. However, Dupuytren’s contracture is known to be recursive and follow-up was too short to make any statements about long-term results.

5.4.1 Total end range time

The TERT theory is based on a study looking at post-traumatic PIP joint stiffness. In their experiment, Flowers and LaStayo (1994) used plaster casts, which can be considered an even more static approach than using a hand splint. Still, they found that removing the cast after six days resulted in a significantly greater ROM than did removal after three days. Following their theory that ROM only

Figure 5.3 | One wear period. Typical example of the force measurement of one wear period. Six times per hour the force is logged for 10sec, represented by the vertically grouped blue dots. The prediction from the fit is presented as a solid blue line. The lightblue area represents the 95% confidence interval of the fit. After $t = \tau$, $t = 2\tau$ and $t = 3\tau$, the force has decreased by 63%, 86% and 95% respectively.

improves when a joint is held in its end range, and assuming that a certain threshold force is needed to keep a joint in its end range, then it can be concluded that a certain threshold force is needed to improve ROM. Our interpretation of Flowers and LaStayo (1994)'s measurements is that after three days the force must still have been above the threshold force. Our measurements consistently show that a plateau is reached within hours. Be aware that Flower's subjects had closed trauma whereas ours had open fasciectomy.

Figure 5.4 | Adaptation rate vs days. After surgery the adaptation time decreased somewhat. Every blue dot represents the calculated adaptation rate of one wear period, its opacity represents the goodness of fit (R^2) for that wear period. The solid orange line represents the fit to the data ($\approx 5\%/day$, $p = 0.0005$, $R^2 = 0.08$), with the dotted orange line the 95%CI interval to this fit.

5.4.2 Postoperative splinting

Without proper postoperative care, the surgically gained ROM might soon be lost again due to wound contraction and scar tissue formation in a relaxed (i.e. flexed) position. Hand therapy prevents stiffening of the joints and stimulates recovery. The hand splint may help to maintain the ROM that is gained with surgery.

A hand splint is intended to hold the affected joint in its end range, like the plaster cast did in the experiment of Flowers and LaStayo (1994). Our results consistently show that with this static approach the force decreases over time to

reach a mechanical equilibrium within hours. On average, the adaptation rate was calculated to be 0.85 hours, so in most cases 95% of the equilibrium (the adaptation time, 3τ) was reached after wearing the splint for about three hours. When the force has reached an equilibrium, the joint can no longer be in its end range, and — since the joint is immobilized by the splint — the end range must have shifted. Following the TERT principle the splint would only improve ROM during the few hours before reaching the mechanical equilibrium. However, other factors that were not measured such as compression might have an effect on the adaptation of the tissues. More research is needed to assess the minimal wear duration.

5.4.3 Wound contraction

During the proliferation phase of wound recovery, myofibroblasts surrounding the wound produce collagen and contract to close the wound (de Morree, 2014). Wound contraction is usually most prominent in the first weeks after surgery, and the greatest risk of losing ROM can be expected. This phenomenon is reflected in our observation that the adaptation time decreases in the weeks after surgery. It suggests that in the first days or weeks after surgery the tissue is more plastic, therefore more ROM can be regained and it takes longer to reach the equilibrium. In later weeks, the wound matures and less ROM is lost, hence the decreasing adaptation time. Following this principle the splint should be used at least as long as wound recovery continues. Indeed, hand therapists prescribe splint use for three to six months (Collis et al., 2013).

5.4.4 Study limitations

Because the force sensors were not placed directly on the skin but rather underneath a protective layer of foam, the measured forces did not equal the actual forces on the finger. This does not influence the magnitude of τ , but could give

unrealistic low values for F_∞ and F_0 since the force is distributed over a larger area than measured.

The exclusion of a wear period for further analysis based on its force data might create a risk of confirmation bias. However, the force data of these excluded wear periods would appear to have multiple spikes or an overall chaotic pattern, resulting in a low goodness of fit and a large 95% CI for τ when trying to fit the data to EQ. (5.1). This profile fits the description of subjects mentioning that occasionally they would readjust or remove the brace at night. In those cases the exact beginning and end of a wear period could not be determined in post-processing because these actions would not result in a sudden temperature change. Therefore, exclusion based on the fit analysis is justified.

For safety and comfort a relatively small battery was used which required the use of an intermittent measurement protocol to save battery power. Since the force and temperature were measured every ten minutes, this means that the exact beginning of each wear period has an maximum error of ten minutes. This error will have influenced the accuracy of the calculation of small adaptation rates ($< 0.5\text{hr}$). The error will also have influenced the magnitude of F_0 but not of F_∞ .

From a scientific point of view passive ROM might have been a more relevant measure than active ROM, but for ethical reasons we were unable to deviate from the standard treatment protocol.

The choice to only include the finger with the largest deficit in case of multiple affected fingers might have skewed the outcomes.

Although a clear decrease in force was observed, it does not necessarily say anything about the internal structures in the hand. No medical imaging techniques were used to be able to correlate the measured force to physical changes in the hand. The reached mechanical equilibrium within several hours suggests rapid tissue adaptation, but does not provide hard evidence since the correction

of contractures might be multi-factorial.

5.5 Conclusions

Forces in postoperative splinting in the treatment of Dupuytren contracture were measured. Each night the observed force decreased to reach an equilibrium after about three hours. In the weeks following surgery the adaptation time decreased slightly, so every week the equilibrium was reached earlier.

Chapter 6

Rapid decrease of cast-induced forces in the Ponseti method

6

BOB GIESBERTS, EDSKO HEKMAN, BART VERKERKE AND PATRICK
MAATHUIS

This chapter is accepted for publication (21-Sept 2018) as R.B. Giesberts, E.E.G. Hekman, G.J. Verkerke and P.G.M. Maathuis (n.d.) Rapid decrease of cast-induced forces in the Ponseti method, *The Bone & Joint Journal*

Abstract

Background: The Ponseti method is an effective evidence-based treatment for clubfoot. This treatment method uses gentle manipulation to adjust the position of the clubfoot in serial steps towards a more physiologic position. Casting is used to hold the newly achieved position. At first the clubfoot resists the new position imposed by the plaster cast, pressing against the cast, but over time the tissues in the clubfoot are expected to adapt to the new position and the force decreases. The aim of current study is to test this hypothesis by measuring the forces between clubfoot and cast in the Ponseti method.

Methods: Force measurements were taken during the treatment of 10 idiopathic clubfeet. Force data was collected for multiple weeks at the location of the talar neck (TN) and the first metatarsal (FM) to determine the adaptation rate of the clubfoot.

Results: In all measurements the force decreased over time. The median (inter-quartile range) of the half-life time was determined at 26 (20 – 53) minutes for FM and 22 (9 – 56) minutes for TN. This suggests that the clubfoot tissues adapt to the new position within several hours.

Conclusions: This is the first study that provides objective force data that supports the hypothesis of adaptation of the idiopathic clubfoot to the new position imposed by the cast. We showed that the expected decrease in corrective force over time does indeed exist. An important new insight on the physiology of the Ponseti method is that the required correction time may well be much shorter than the originally proposed and generally accepted period of one week.

6.1 Introduction

Clubfoot (*talipes equinovarus*) is a common congenital deformity which is often treated with the Ponseti method (Ponseti & Smoley, 1963). In this method, the treating physician manipulates the clubfoot by applying manual pressure on the medial side of the *first metatarsal* (FM) with counter pressure on the lateral side of the *talar neck* (TN). This pressure causes the forefoot to align properly with the talus while moving the calcaneus in the subtalar joint to its proper position. The newly imposed position stretches and remodels the clubfoot's tissues and is maintained for a week with a plaster cast. Most cases of clubfoot are corrected after five to six cast changes and, in many cases, a percutaneous Achilles tenotomy (Morcuende, Dolan, Dietz, & Ponseti, 2004). After the casting period, a foot abduction brace needs to be worn for several years to prevent relapse (Dobbs et al., 2004).

It is suggested that in serial casting most of the correction is obtained by the manipulation. As suggested by Carroll (1990) “*The plaster does not produce the correction*” (as cited in R. Brand et al., 2006), but it merely maintains the improved orientation of the foot, gained by the manipulation. The mechanical behavior of the clubfoot tissues can be characterized as “visco-elastic” (Lis, de Castro, & Nordin, 2012; Özkaya et al., 2012). That is, its behavior is partly elastic (i.e. deformation is recovered after removal of the load) and partly plastic (deformation is permanent). Initially the clubfoot's tissues resist the correction, acting as a spring pressing on the plaster cast. When the plaster cast is removed after a week, the clubfoot does not immediately return to its original position. Apparently over time these tissues have adapted to the applied force. The decreasing stress in response to an initial stretch is known as *stress-relaxation* and the rate at which the force decreases is the adaptation rate. It can be characterized by a half-life time $t_{1/2}$, which is the time required for the force to decrease to half of its original value.

Objectively quantifying the Ponseti method will create a better understanding of its physiology. Force sensors are used to obtain a measure for the correction forces. We hypothesize that *the force which the clubfoot applies to the plaster cast will decrease over time, as a result of the adaptation of the tissues in the clubfoot to the position imposed by the plaster cast*. The aim of this study is to test this hypothesis, to determine correction forces in time and to calculate the half-life time of a clubfoot in the Ponseti method.

6.2 Methods

This study is an exploratory observational study performed at the University Medical Center of Groningen (UMCG) situated in the Netherlands.

The medical ethical evaluation committee of the UMCG reviewed the study in accordance with the declaration of Helsinki, and declared that it does not meet the criteria as stated by the Medical Research Involving Human Subjects Act (WMO), and therefore did not require medical ethical approval (document number M16.196266).

6.2.1 Children

Children with idiopathic clubfoot, who were younger than 3 months, and who had had no form of prior treatment were selected for this study. After obtaining informed written consent from both parents, the children were included in the study. From the children who presented with bilateral clubfoot, the foot with the highest Pirani and Diméglio score was included for measurements.

Table 6.1 | Force sensor characteristics. For more details: (Giesberts, Sluiter, & Verkerke, 2018, CHAPTER 4).

Characteristic	Value
Dimensions ($\varnothing \times$ thickness)	10×2.3 mm
Resolution	0.15×10^{-3} N
Accuracy	3.4 %
Sample rate	18 Hz
Drift	< 2.1 % / $\log_{10}(\text{hr})$
Hysteresis	6.0 %
Temperature sensitivity	-0.1 N / $^{\circ}\text{C}$

6.2.2 Materials

FIGURE 6.1A shows a custom made force sensor, based on inductive sensing (Kasemsadeh & LaPointe, 2016), that was developed for the specific application of this study. The output signal of this force sensor is LHR.DATA, a 28-bit digital value proportional to the resonance frequency of the sensor, which is a measure for the applied force. More information about the force sensor and its performance is presented in CHAPTER 4 (Giesberts, Sluiter, & Verkerke, 2018). The specifics are summarized in TABLE 6.1. A digital DS1825 thermometer (*Maxim Integrated*, San Jose, CA, USA) was included to allow for the correction of temperature induced signal disturbance.

An acquisition unit equipped with a small battery and microSD card was used to locally store the measurements of both the force sensors and the thermometer (see FIGURE 6.1B).

Figure 6.1 | Use of the force sensor. Two custom made inductive force sensors were placed on the clubfoot to measure the force of the foot on the cast. Wires were routed distally and the acquisition unit was attached to the cast laterally.

6.2.3 Measurement protocol

Before use, the force sensors were calibrated using ten weights of 100g each. To assess the severity of the clubfoot deformity and the progress of the treatment, both the Pirani and Diméglio scores were taken.

Before manipulation of the foot, the treating physician carefully placed a force sensor on both the FM and TN using Cutinova Hydro (*Smith&Nephew*, Hull, England) to protect the skin underneath the sensor, and Medipore tape (*3M*, Neuss, Germany) to attach it to the skin (see FIGURE 6.1C). The thermometer was placed on the plantar aspect of the foot (see FIGURE 6.1B). A strip of Coban™ tape (*3M*, Neuss, Germany) was applied between the skin and the thermometer and wires.

The foot was treated following the Ponseti method. Wires were routed distally and, as soon as the plaster had set, the acquisition unit was attached to the cast laterally with Peha-haft bandage (*Hartmann*, Heidenheim, Germany) (see FIGURE 6.1B).

At the next scheduled meeting the cast and sensors were removed, after which the data was copied to a computer and the battery replaced while the parents had the opportunity to bathe their child. The foot was inspected for any signs of skin damage and the protocol was repeated until the final cast before tenotomy.

6.2.4 Data collection

The system was programmed to store time, force and temperature data continuously for the first four hours, after which the system entered a low power state to extend battery life. Four times per hour it woke up from this state to collect data for ten seconds (see FIGURE 6.2). The timing of specific steps in the protocol (placing the sensors, application of plaster cast, releasing the foot) was noted to identify important events in the data.

Figure 6.2 | Measurement protocol. The system was programmed to measure the force continuously for the first 4 hours after which it switched to a battery saving protocol with four 10 second long periods per hour.

6.2.5 Mathematical model of the mechanical behavior

Assuming that tissue mechanically behaves as a visco-elastic material, the resistive force when it is stretched to a certain length can be modeled using EQUATION (6.1) (Murata, 2012).

$$F(t) = (F_0 - F_\infty) \cdot 2^{-\frac{t}{\tau_{1/2}}} + F_\infty \quad (6.1)$$

In this equation, $F(t)$ is the (resistive) force at time t , and F_0 and F_∞ are the starting force ($t = 0$) and end force after a long time (at infinity), respectively. Parameter $t_{1/2}$ is known as the *half-life time*. What the equation means is that the force decays from starting value F_0 to final value F_∞ . The lower the half-life time, the faster the force decays. The half-life time $t_{1/2}$ represents the time after which the force has dropped from F_0 to halfway the final value F_∞ . FIGURE 6.3 illustrates how the force decays to 95% after 4.3 half-life times.

Figure 6.3 | Expected force decrease. The effect of EQUATION (6.1) is shown in this figure. After $t_{1/2}$ hours the force will have decreased from F_0 to F_∞ with 50%, after $2t_{1/2}$ with 75%. After $4.3t_{1/2}$ 95% of the decrease will have happened.

6.2.6 Data processing

Each retrieved data set consisted of time, force and temperature data of one week. All data was processed using Matlab version R2017b.

Erroneous data points (i.e. $LHR_DATA = 2^{28} - 1$ as explained in Texas Instruments (2014) or $T = 85^\circ C$ as explained in Maxim Integrated (2005)) were removed from the datasets.

The data was corrected for temperature changes and filtered to reduce the influence of any noise caused by movements of the child. As such, the first four hours of continuous force sensor data was processed using a moving standard deviation filter *stdfilt* with a window of 10sec to identify and remove spikes. Next, the data was low-pass filtered (cut-off frequency 1/60 Hz) and re-sampled to have one data point per 10sec. Each 10sec period in the force sensor data was filtered using the same standard deviation filter and the median was taken over the filtered data (see FIGURE 6.2).

The force sensor data (LHR_DATA) were then converted to a force value ($F(t)$) using the calibration data. The start of each measurement (t_0) was defined as the exact moment the physician released the foot (that is, when the casting was finished).

Each data set was then fit to EQUATION (6.1) using Matlab 2017b's *fit* function. This function generates values and 95% confidence intervals (CI) for the initial force F_0 , the final force F_∞ and the half-life time $t_{1/2}$. The used restrictions (bounds) for the fit function were $0.001 \cdot \ln(2) < t_{1/2} < 1000 \cdot \ln(2)$ hrs and $F_\infty < F_0$. Data sets for which the half-life time could not be calculated ($R^2 < 0.50$) were excluded from further analysis.

6.2.7 Statistics

The determined values for $t_{1/2}$, F_0 and F_∞ were tested for normality using Lilliefors' test and presented as median (interquartile range, IQR). Correlation of the calculated half-life time with the cast number and the Diméglio score and age at the moment of casting were determined using Spearman's rank correlation coefficient. Statistical significance was defined as $p < 0.05$.

Characteristic	Value
Children	10
Age (range)	7 (2-30) days
Boys / girls	9 / 1
Bilateral	6
Pre-treatment Pirani	4.3 $\pm$ 0.7
Pre-treatment Diméglio	14.0 $\pm$ 1.6

Table 6.2 | Child characteristics. Unless stated otherwise numbers are presented as mean $\pm$ standard deviation.

6.3 Results

Force measurements were performed in the treatment of ten clubfeet of which the child characteristics are shown in TABLE 6.2. Complications in the form of pressure marks on the skin were encountered in four cases. In two of those cases further measurements were canceled but the obtained data were included for analysis. Technical malfunctioning occurred in ten measurements (eight times insufficient battery power, one disconnected sensor, 1 disconnected thermometer). In total 29 successful measurements were performed on 10 clubfeet.

6.3.1 Half-life time

The results of one typical measurement are presented in FIGURE 6.4. For five data sets (three for FM, one for both FM & TN) the goodness of fit was too small ($R^2 < 0.50$) to determine a relation between the measured force and the elapsed time, and were excluded for further analysis. In total 25 data sets for the force on FM and 28 for the force on TN were analysed. The calculated values for $t_{1/2}$, F_0 and F_∞ are presented in TABLE 6.3.

Figure 6.4 | Decreasing force. Typical example of the measured force on FM (top) and TN (bottom). The raw data in grey is first processed to get the filtered data in orange and blue. The filtered data is fit to EQUATION (6.1) to find the half-life time $t_{1/2}$.

Sensor location	Parameter	Value	
FM	$t_{1/2}$ (h:mm)	0:26	(0:20 – 0:53)
	F_0 (N)	0.29	(0.20 – 0.49)
	F_∞ (N)	0.09	(0.00 – 0.21)
TN	$t_{1/2}$ (h:mm)	0:22	(0:09 – 0:56)
	F_0 (N)	0.41	(0.21 – 0.57)
	F_∞ (N)	0.03	(-0.42 – 0.16)

Table 6.3 | Data is presented as median (interquartile range) as it was not normally distributed.

Figure 6.5 | Correlations. In all three diagrams FM is represented as orange points and TN as blue points. The open circles refer to the measurements of one child that had much higher values for the half-life time than the others. Spearman's test was used to determine the correlation coefficient ρ .

6.3.2 Correlations

FIGURE 6.5 shows the correlation of the calculated half-life time with the cast number and the Diméglio score and age at the moment of casting. Trends could be observed for a lower half-life time with increasing cast number ($\rho = 0.46$, $p = 0.021$) and age ($\rho = 0.44$, $p = 0.030$), but only for FM and not for TN.

For unknown reasons the calculated half-life times for one child were much lower than for the others. This child was not particularly older or younger than the other children, nor were his clubfeet much different from the other included clubfeet. Excluding these measurements from the analysis made the initially found statistically significant correlations disappear, but introduced a negative correlation of the half-life time with the Diméglio score at the moment of casting for FM ($\rho = -0.51$, $p = 0.037$).

6.4 Discussion

Studies have estimated the forces (Giesberts et al., 2017, CHAPTER 3) and torque (Cohen et al., 2012) on the clubfoot, but never before was force measurement performed with measurements in multiple patients spanning several weeks each. In the present study the change of the applied force in the Ponseti method is made visual. A clear decrease in force was observed in all measurements, with almost all measurements showing a typical stress-relaxation curve. This supports our hypothesis (decrease of force over time). The adaptation may have taken place in the foot, in the cast or both. However, the creep behavior of cast materials is far too limited for this observed effect (Cohen et al., 2013). This indicates that the tissues in the clubfoot adapt to the position imposed by the cast.

6.4.1 Limitations

Since the sensors are attached to the skin with tape, each measurement has an unknown baseline force acting upon it. Additionally, the exact location of the sensor is unknown since both the skin and the padding are able to move in relation to the cartilaginous structure below. This makes the absolute values that are registered by the force sensors somewhat unreliable as can be seen in the unlikely negative values for F_{∞} . However, this does not affect the calculation of the half-life time $t_{1/2}$ since this is only determined by the curve's shape and not its absolute value.

The small population size did not allow for any statistical analysis of differences within the group such as laterality or sex.

Although a clear decrease in force was observed, it does not necessarily say anything about the internal structures in the clubfoot. No medical imaging techniques were used to be able to correlate the measured force to physical changes in the clubfoot. In every measurement a mechanical equilibrium was reached within

several hours, but this does not necessarily indicate that a biological equilibrium is reached within the same time span.

6.4.2 Stress-relaxation

As shown in FIGURE 6.3 after 4.3 times the half-life time ($4.3t_{1/2}$, which in our case corresponds to less than 2 hours), the corrective force has been reduced by 95% with respect to its final value F_{∞} . After this first period the forces do not change much: the tissues are in a stable position.

Magnetic Resonance Imaging (MRI) studies comparing tarsal bones before and about ten minutes after casting show immediate shape changes to the cartilage anlagen of the tarsal bones (R. Brand et al., 2006; Mahmoodian, 2010; Philip, 2014). Stress-relaxation is said to occur within 15–20 minutes after application (Fung, 1981) (as cited in Flowers, 2002), but therapists are warned that *“this phenomenon cannot be repeated many times in rapid succession, because once the relaxation has occurred, the tissue length is maintained by the elastic restraints within the tissue components such as covalent cross-links”* (Peacock, 1984) (as cited in Flowers, 2002). Our observation that the clubfoot reaches a mechanical equilibrium within 2 hours suggests that most of the correction is done long before the end of the week and treatment could be accelerated drastically. Decreasing the casting duration has been proposed before (R. Brand et al., 2006; Mahmoodian, 2010; Philip, 2014) and performed in studies in which the cast change interval was changed from weekly to twice a week (Elgohary & Abulsaad, 2015; R. J. Xu, 2011) and even three times per week (Harnett et al., 2011). The studies all report a reduced treatment time without any signs of adverse effects (Giesberts, van der Steen, et al., 2018, CHAPTER 2). More research is needed to find an optimal cast change interval. Our findings suggest that if there is a limitation to shortening the casting interval, it is biological rather than stress related.

6.5 Conclusion

In the treatment of clubfoot with the Ponseti method the foot is manipulated and held in place with a plaster cast. The applied forces rapidly decrease and reach an equilibrium within hours after casting. This suggests that:

- the clubfoot tissues adapt and remove stresses within hours
- for stress reasons there is no need to wait one week with changing the cast
- the treatment duration can potentially be drastically decreased

More research is needed to determine the optimal cast change interval.

Chapter 7

Ten Cold Clubfeet

BOB GIESBERTS, EDSKO HEKMAN, BART VERKERKE AND PATRICK
MAATHUIS

This chapter is published as R.B. Giesberts, E.E.G. Hekman, G.J. Verkerke and P.G.M. Maathuis (2018) Ten Cold Clubfeet, *Acta Orthopaedica*, doi: 10.1080/17453674.2018.1493046.

Abstract

Background and purpose: Idiopathic clubfeet are commonly treated with serial manipulation and casting, known as the Ponseti method. The use of Plaster of Paris as casting material causes both exothermic and endothermic reactions. The resulting temperature changes can create discomfort to the patients.

Patients and methods: In 10 patients, we used a digital thermometer with a data logger to measure below-cast temperatures to create a thermal profile of the treatment process.

Results: After the anticipated temperature peak, a surprisingly large dip was observed ($T_{min} = 26^{\circ}\text{C}$) that lasted 12 hours.

Interpretation: Evaporation of excess water from a cast might cause discomfort for clubfoot patients and subsequently, their caregivers.

7.1 Introduction

The common treatment of idiopathic clubfoot (*talipes equinovarus*) consists of serial manipulation and casting, known as the Ponseti method (Ponseti, 2008). The treatment is started in the first weeks after birth and includes on average five cast changes, often followed by a percutaneous Achilles tenotomy. The deformity is successfully corrected in over 90% of all cases (Morcuende et al., 2004). An abduction orthosis is worn for several years to prevent relapse (Dobbs et al., 2004).

The Ponseti method dictates toe-to-groin casts with the knee flexed to prevent slipping of the cast (Maripuri, Gallacher, Bridgens, Kuiper, & Kiely, 2013; Ponseti, 2008). Plaster of Paris (PoP) is the accepted casting material, because it is inexpensive, easily obtained and it can be easily molded (Aydin et al., 2015; Pitner et al., 2008). The use of PoP as casting material involves both an exothermic (hot) setting reaction and an endothermic (cold) drying process.

7.1.1 Setting

Bandage rolls impregnated with PoP are soaked in lukewarm water before being wrapped around the foot. The PoP has the following exothermic setting reaction with water:

Many studies have focused on the extent of the generated heat and the effects of the associated temperature peak, that appears minutes after application (Burghardt, Anderson, Reed, & Herzenberg, 2014; Shuler & Bates, 2013). Thermal injuries can occur if the temperature exceeds 40°C for longer periods of time, but can be prevented by optimizing the number of layers used, cast type and brand, dipping water temperature and cast padding thickness (Halanski et

al., 2007; Hutchinson & Hutchinson, 2008; Shuler & Bates, 2013). Only a few layers of cast tape are needed to keep the tiny clubfeet in position, therefore thermal injuries are uncommon in the Ponseti method.

7.1.2 Drying

Evaporation of the excess water in the cast is an endothermic process, meaning that it extracts energy from its surroundings, thereby cooling it. The required energy for the evaporation of the excess water – the latent heat – is drawn in part from the patient. The effect of this endothermic reaction in the treatment of clubfoot has not been studied before. However, anecdotal information from parents suggests that some children struggle to keep warm with the wet and cold casts (Bridgens & Kiely, 2010; van Doorn, 2016).

We assessed the below-cast temperature profile in the treatment of clubfoot with the Ponseti method, in terms of both temperature drop and the extend of this drop.

7.2 Patients and methods

7.2.1 Children

10 Children with idiopathic clubfoot, who were younger than 3 months old, and who had had no form of prior treatment were selected for this study (TABLE 7.1). From the patients who presented with bilateral clubfoot only one foot was included for measurements.

Characteristic	Value
Subjects	10
Age (range)	7 (2-30) days
Boys / girls	9 / 1
Bilateral	6
Pre-treatment Pirani score	4.3 ± 0.7
Pre-treatment Diméglio score	14.0 ± 1.6

Table 7.1 | Patient characteristics. Unless stated otherwise numbers are presented as mean ± standard deviation.

7.2.2 Sensors

A 1-wire DS1825 digital thermometer (*Maxim Integrated*, San Jose, CA, USA) was used to measure the temperature under the cast. The accuracy of the sensor is ±0.5°C (*Maxim Integrated*, 2005).

7.2.3 Protocol

After manipulation of the foot, the orthopedist placed the thermometer on the plantar aspect of the foot (see FIGURE 7.1A). A strip of Coban™ tape (*3M*, Neuss, Germany) was used between the skin and the thermometer and wires (see FIGURE 7.1B). The thermometer was connected to an electronic data logging system with a small battery and SD card, enclosed in a 3D-printed casing. After placing the sensor, the foot was treated as usual following the Ponseti method.

The cast technician cut one roll of Gypsona (*BSN Medical GmbH*, Hamburg, Germany) in half, dipped it in a sink filled with lukewarm (20–25°C) water and applied it to the clubfoot while the orthopedist maintained the position of the foot. Care was taken not to use more than three layers of cast tape. Wires were routed distally and, as soon as the plaster had set, the casing was attached to the cast laterally with a single layer of Peha-haft bandage (*Hartmann*, Heidenheim,

Figure 7.1 | Thermometer location. (a) Location of the thermometer on the foot, below the cast. The plaster is represented in blue. The casing for the data logging system was attached to the cast with a layer of Peha-haft bandage. (b) Schematics representation of the layers between the skin and the thermometer.

Germany) (see FIGURE 7.1A). Parents were instructed to elevate the legs of the child with a rolled up towel, as protocol directs.

At the next scheduled meeting the cast and thermometer were removed. While the parents bathed their child, the sensors were left in a stable position for at least 15 minutes for calibration after which the data was copied to a computer and the battery replaced. The foot was carefully inspected for any signs of skin damage and the protocol was repeated until the final cast before tenotomy.

7.2.4 Water content

The same measurement protocol was used to cast one rubber clubfoot model (*MD Orthopaedics, Inc., Wayland, IA, USA*). The weight of all components was documented and the total weight of the rubber foot plus cast was monitored for several days. This provided an estimation of the excess water content in the plaster cast and the time it takes to evaporate.

The latent heat of evaporation of water is described as $Q = m \cdot \Delta H_{ev}$, with

m the measured mass of the excess water and $\Delta H_{ev}(25^\circ\text{C}) = 2442 \text{ J/g}$ (Rajput, 2010).

7.2.5 Data collection

The system was programmed to store time and temperature data continuously (0.16 Hz) for the first four hours, after which the system entered a low power state (85 μA) to save battery power. Four times per hour it woke up from this state to collect data for ten seconds. Each 10s period resulted in two temperature measurements over which the average was taken. The system featured a button to wake up after the cast was removed.

7.2.6 Data processing

Data was processed using Matlab version R2016b. Potential sensor errors were filtered using the median filter *medfilt1* (Pratt, 2007). The start of the measurement (t_0) was defined as the moment the plaster cast touched the foot. The height of the temperature peak (T_{max}) was calculated as the maximum value over the first 3 hours, the final equilibrium temperature (T_{end}) as the mean value over the final three days of measurement, and the lowest temperature (T_{min}) as the minimum value after T_{max} .

The beginning of the temperature dip was defined as the moment the temperature reached a value smaller than one standard deviation of T_{end} . The moment the temperature reached a value within one standard deviation of T_{end} was defined as the end of the temperature dip.

The three sets of temperature data (T_{max} , T_{min} and T_{end}) were tested for normality using the Shapiro-Wilk test. Student T-Tests were used to test the difference between temperatures for statistical significance.

7.2.7 Ethics, funding, and potential conflicts of interest

The medical ethical evaluation committee of the UMCG reviewed the study in accordance with the declaration of Helsinki, and declared that it did not meet the criteria as stated by the Medical Research Involving Human Subjects Act (WMO), and therefore did not require their approval (document number M16.196266). After obtaining informed written consent from both parents, the patients were included in the study. Financial support was provided by Stichting voor de Technische Wetenschappen [grant P12-03]. No competing interests declared.

7.3 Results

Complications in the form of pressure marks were encountered in 4 cases. In two of those cases further measurements were canceled but the obtained data was included for analysis. Technical malfunctioning occurred in 5 measurements (disconnected sensors, disconnected thermometer, faulty measurement protocol, empty battery). 29 Successful measurements were performed on 10 clubfeet, giving more than 3000 hours of data.

7.3.1 Clubfoot model - Water content

The measurement with a clubfoot model showed that approximately 35g of water was used for one roll of PoP bandage of 75g. Over the course of two days the water content in the cast decreased to reach an equilibrium of 14g. This means that 14g of the original water content attached to the PoP to form hard gypsum and the remaining 21g of excess water evaporated. This ratio fits well within the molecular ratio of the chemical reaction 7.1, which also predicts the generation of 740 kJ of heat (see TABLE 7.2). The required energy for evaporation of 21g of excess water is 51 kJ.

Table 7.2 | Molecular ratios The 21g of excess water measured (column highlighted in grey) fits the prediction based on the molecular ratios of the exothermic setting reaction.

	$CaSO_4(H_2O)\frac{1}{2}$	+	$1\frac{1}{2}H_2O$	→	$CaSO_4(H_2O)_2$	+	H_2O	+	heat
Molecular mass (mol ⁻¹)	141.15g		$1\frac{1}{2} \cdot 18.015g$		172.17g		18.015g		1433kJ
Mass (mol)	0.52		0.78		0.52		excess		0.52
Mass (g)	75		14		89		21		740kJ

Figure 7.2 | Results from the temperature measurements. The orange line indicates the average temperature over all measurements, the blue area the standard deviation (SD).

7.3.2 Clubfoot measurements

All temperature measurements showed a clear temperature peak after the application of the plaster cast (T_{max}). None of the below-cast temperatures ever exceeded 40°C and no thermal injuries were observed (FIGURE 7.2 and TABLE 7.3).

After the peak, a decrease to a significantly lower temperature was observed ($p < 0.001$) (FIGURE 7.3). On average, this temperature dip lasted 12 (range

Table 7.3 | Temperature data. Time data presented as median (interquartile range, IQR). Temperature data presented as mean $\pm$ standard deviation (95% confidence interval). Temperature data was tested for normality using Shapiro-Wilk tests.

	Time (hrs)	Temperature ($^{\circ}\text{C}$)	Normality
T_{max}	0.15 (0.1– 0.8)	35.3 $\pm$ 1.6 (34.7–35.8)	(p = 0.484)
T_{min}	4.9 (4.0– 7.3)	26.1 $\pm$ 3.0 (25.0–27.2)	(p = 0.214)
T_{end}	11.0 (9.6–15.8)	34.9 $\pm$ 1.4 (34.4–35.5)	(p = 0.087)

3.2-30) hours followed by a gradual increase to a stable final temperature (T_{end}).

All three sets of temperature data were normally distributed (see TABLE 7.3). There was no statistical difference between the temperature peak and the equilibrium temperature (p = 0.305).

7.4 Discussion

This is the first study to present long-term temperature measurements underneath the cast in the treatment of clubfeet. A surprisingly large temperature dip was observed caused by the evaporation of excess water. In several cases temperatures as low as 22 $^{\circ}\text{C}$ were observed and they stayed low for more than 12 hours.

7.4.1 Evaluation of the results

Our measurements showed a temperature peak after 9 minutes (TABLE 7.3). BSN Medical’s own measurements with Gypsona show a peak after 12 minutes (see FIGURE 7.4). Burghardt et al. (2014) found a peak temperature after 9 minutes when using 16 layers of a different brand of plaster.

Immediately after reaching the peak, the cast started to cool. This is also

Figure 7.3 | Difference in the measured temperature. T_{max} represents the temperature peak shortly after casting, T_{min} the lowest temperature reached after the initial peak, and T_{end} the equilibrium temperature.

Figure 7.4 | Gypsona temperature profile. Measurements from BSN Medical show the steep increase and subsequent gradual decline of the temperature after application. Copied with permission from BSN Medical.

visible in BSN Medical's measurement in which the temperature drops below the ambient temperature after approximately one hour (FIGURE 7.4). Depending on environmental influences it can take two to three days for the cast to dry completely (Hayter, 2015). Indeed, in our water content measurement the cast reached a stable weight after two days. However, on the clubfeet it took an average of 11 hours before the temperature reached normal levels again. This suggests that the plaster dried much quicker when placed on the child's foot and that some of the child's energy was used to evaporate the water of the plaster cast.

7.4.2 Casting-induced hypothermia

In adults, one case study exists which describes observed hypothermia after the application of a plaster jacket for immobilization of the cervical spine (Vale, 1969). Caregivers of children with a clubfoot have mentioned that their children can be

a bit prickly the first 24h after a new cast is applied (Bridgens & Kiely, 2010; van Doorn, 2016). The observed cold period might be one of the causes for this behavior.

It would take 51 kJ to evaporate the excess water of one cast. Part of this energy will be extracted from the environment, but a substantial part of it must be produced by the child. However, the thermoregulatory system of newborns is not fully developed yet, and they experience great difficulties when facing a cold environmental challenge (Pierro, De Coppi, & Eaton, 2012). Moreover, newborns have a relatively large surface area, so much heat is lost in relation to their heat producing volume.

A newborn's total energy expenditure in the first month after birth is 1.0–1.3 MJ/day (FAO, 2001; Olgaher & Forsum, 2003). Up to 8% of this energy — approximately 100 kJ/day — is used for thermogenesis and thermoregulation (Pierro et al., 2012). Relatively, the required 51 kJ for the evaporation of excess water is quite substantial and a mild form of casting-induced hypothermia in the treatment of clubfoot seems plausible, especially with bilateral clubfeet in vulnerable patients, such as prematurely born children. In those cases, postponing treatment at least until after the first month might be beneficial (Alves, Escalda, Fernandes, Tavares, & Neves, 2009; Awang, Sulaiman, Munajat, & Fazliq, 2014; ltar et al., 2010).

Given that newborns are naturally well monitored by their parents, that hypothermia in newborns is easily detected and that no such case studies exist for clubfoot, it is unlikely that casting induces severe or moderate hypothermia. Still, having cold feet for half a day undoubtedly is uncomfortable and plausibly affects the newborn's mood.

7.4.3 Limitations

We acknowledge limitations to our study. First, we did not measure whether the wet and cold casts affect the child's core temperature. If they do, a more prominent effect would be expected in bilateral cases. However, the number of children in our study does not allow for any meaningful test for subgroup differences, and no trend could be identified.

Second, the sensors were placed on the skin on the plantar aspect of the foot. So the measurements only reflect the temperature of that part of the foot and not the core temperature of the patient. Hypothermia was neither measured nor diagnosed based on clinical symptoms. Future research should include measuring the core temperature during the first 24 hours after casting to study the possibility of casting induced hypothermia.

7.4.4 Clinical implications

So far only anecdotal information exists about cold clubfeet (Bridgens & Kiely, 2010; van Doorn, 2016). In some cases, especially with bilateral clubfeet in vulnerable patients, it might be better to postpone treatment to prevent any harmful effects of the wet and cold casts. This is an extra argument to defer treatment in premature babies for several weeks, as stated by the First European consensus meeting on Ponseti clubfoot treatment's agreement (Böhm & Sinclair, 2013).

Preventing a clubfoot from getting cold would be possible if less excess water needed to be evaporated. However, squeezing the water out of the plaster to use less water results in hotter plaster and increased risk of burns: the excess water is needed to release the heat from the exothermic reaction (Kaplan, 1981).

Using synthetic cast tape as an alternative casting material could be a viable option to decrease the discomfort of cold feet. The chemical reaction and the structure of the material both cause the drying process to be shorter than when

using PoP.

In summary, after an anticipated temperature peak, a surprisingly large temperature dip was observed. More research is needed to assess whether the Ponseti method affects the child's core temperature.

Chapter 8

Design of a dynamic clubfoot brace

8.1 Introduction

The Ponseti method is the golden standard for the treatment of clubfoot (Hennessey, 2012). Many studies have provided convincing evidence for the efficacy of the treatment in various settings. Comparative studies have shown the superiority of the Ponseti method over surgery (Herzenberg et al., 2002) and Kite's method (Derzsi et al., 2015; Matos & de Oliveira, 2010) and gives comparable results as the French method (Richards et al., 2008). Many of the details in the Ponseti method such as casting material and starting age have been studied and all seem to lead to the conclusion that one should not deviate from the Ponseti method: the details matter (Nogueira, 2016; Zhao et al., 2014). Moreover, Funk (2018) says that there is no alternative to the Ponseti method. In this setting there is limited room for critical evaluation on the method, leading to a risk of confirmation bias, overconfidence of the treatment and an overly conservative

attitude towards possible improvements.

8.1.1 The problem

Parents' perspective

We did an online survey with 49 caregivers of children treated with the Ponseti method and found several downsides to the practical aspects of the Ponseti method (van Doorn, 2016). The fact that bathing of the child was not possible was often mentioned, but also that diapers would sometimes leak onto the cast, leaving dirty stains on the cast. Another point of concern was the cast being wet and cold the first few hours after casting, causing discomfort to the children who, due to their age, are unable to warm themselves. This notion was later supported by temperature measurements under the cast (Giesberts, Hekman, Verkerke, & Maathuis, 2018b, CHAPTER 7).

Only few other studies have noted downsides of the Ponseti method (Dibello, Di Carlo, & Carbone, 2016; Malagelada, Mayet, Firth, & Ramachandran, 2016; Nogueira et al., 2013). The procedure of serial casting impacts family function because it requires frequent hospital visits, it makes washing the child impossible and cuddling uncomfortable. Pletch et al. (2015) found that caregivers of children treated with the Ponseti method are concerned about the judgment of others when seeing the child in casts. Caregivers are afraid that others might think that there is something wrong with the child which might have been caused by the caregivers. In our survey several parents explained how they had received unfriendly comments from outsiders such as “*did you drop your child?*”, interpreting the plaster cast as a treatment for broken bones.

In our survey, caregivers all mentioned to gladly accept the downsides because their child's clubfoot was successfully treated. It is therefore not unthinkable that most of these concerns will never reach the physicians, since the caregivers would

be reluctant to criticize the one who treats their child.

Inefficiency

Another point of criticism on the Ponseti method is its apparent inefficiency. Many studies have shown that the cast change rate can be increased from weekly to at least twice weekly without any harmful effects (Giesberts, van der Steen, et al., 2018, CHAPTER 2). Biomechanical analysis of the forces under the cast suggests that the clubfoot tissues have adapted within hours after casting (CHAPTER 6). As soon as the biological tissues have adapted, the function of the cast seems questionable. Its function appears to be to maintain the correction, rather than actively applying any correction itself (Carroll, 1990).

8.1.2 The strategy

Orthosis

In the treatment of other contractures serial casting has been replaced by more user-friendly alternatives like static splints (constant position), static-progressive splints (incremental positions, like a tie-wrap) or dynamic splints (constant force) (Glasgow et al., 2010). The application of a constant position (stretch) induces an initial stress which reduces over time: *stress-relaxation*. On the other hand, the application of a constant force (stress) induces a stretch which increases over time: *creep*. The treatment of clubfoot, or any contracture in general, aims at stretching the affected tissue to reach a larger range of motion of a joint. Since the goal is to stretch the affected tissues, by definition, creep is a more efficient method than stress-relaxation (Özkaya et al., 2012).

Chanu, Bimol, Rai, Wangjam, and Singh (2016) used serial static splints rather than plaster casts to successfully treat 21 clubfeet (FIGURE 8.1A). Parents were instructed how to manipulate the foot three times a day and apply the

(a) Serial static splints (b) Dynamic brace

Figure 8.1 | Work by others. a Chanu et al. (2016) used serial static splints and instructed caregivers how to do manipulation of the clubfoot. b Cunningham (2017) made a dynamic clubfoot brace.

orthosis correctly. After several weeks all feet were corrected which was confirmed with X-ray images. The static splint is soft to the touch and allows caregivers to bathe their child. Although this approach is revolutionary and gives clear benefits to the caregivers, the paper describing the method has not been cited by anyone before.

Cunningham (2017) developed a dynamic brace with a polypropylene strip spiraling around the lower leg (FIGURE 8.1B). The shape of this strip creates a complex combination of moments on the clubfoot and is intended to apply the required corrections in the order as described by the Ponseti method. Prototypes have been made but pilot tests have been limited and unreported.

8.1.3 Goal

The goal of this project has been to investigate if it would be possible to apply the strategy of using an orthosis to the treatment of clubfoot, and as such develop concepts for this strategy, while still following the proven Ponseti method. More specific, the goal for this chapter has been to present concepts for a dynamic clubfoot brace and to discuss their feasibility.

8.2 Requirements

We have established a set of requirements for the proposed dynamic clubfoot brace, presented in TABLE 8.1.

8.2.1 Functions

Functions of dynamic clubfoot brace are as follows:

Move the clubfoot in the desired end position

1. Generate a force to correct the cavus deformity of the foot
2. Generate a constant force to correct the adductus and varus deformity of the foot
3. Hold the child's foot and leg
 - (a) Hold the knee in flexion
 - (b) Stabilize the talus in relation to the leg
4. Fit to all children between 0 and 12 months
5. Enable donning and doffing of the brace
6. Generate information about the position of the foot
 - (a) Generate information about its current position
 - (b) Generate information about when the desired correction is reached

8.3 Concepts

In the course of this project many students have worked on the design of a dynamic clubfoot brace which resulted in a variety of concepts with all kinds of

Table 8.1 | Brace requirements.

#	Requirement	Specification
1	The brace treats clubfoot	Resulting in “a functional pain-free, normal-looking, plantigrade foot, with good mobility, without calluses, and requiring no modified shoes” (Ponseti, 2008)
	... just as effective as with casting	minimal 95% success rate
	... faster than with casting	maximal 2 weeks
	... requiring fewer hospital visits than the casting phase in the Ponseti method	maximal 4 visits
	... with similar costs as the casting phase in the Ponseti method	maximal 500 euro
	... involve caregivers	Caregivers should be stimulated to leave the brace on (van Belle, 2017)
2	The brace follows the Ponseti method	Following the correct order (Ponseti, 2008)
	... correct cavus	Forefoot pronation
	... correct adductus	Foot abduction to 60°
	... correct varus	Foot eversion
	... correct equinus	Foot dorsiflexion to 20°. (Achieved with an Achilles tenotomy)
3	The brace uses the correct pressure points	Pressure on First Metatarsal (FM) and Talar Neck (TN) (Ponseti, 2008)
	Location FM	One pressure point at the medial side of the first metatarsal bone
	Location TN	One pressure point at the lateral side of the talar neck acting as fulcrum
	upper leg	One pressure point at the medial side of the lower and upper leg for counter pressure
	The heel must be free	No pressure on calcaneus (“Kite’s error”)

functional mechanisms ranging from a more conventional spring-damper system to magnetic ferrofluids. In this process Baat Medical and Orthin helped to guide

Table 8.2 | Brace requirements (continued)

4 The brace keeps the clubfoot in its end range	TERT theory (Flowers & LaStayo, 1994)
Constant moment	$M = (0.62 \pm 0.12) \text{ Nm}$ (Cohen et al., 2012)
Sufficient force, higher than equilibrium	$F > 1 \text{ N}$ (P. W. Brand & Hollister, 1999; Giesberts, Hekman, et al., 2018a)
5 The brace is safe	
Prevent skin damage	Force must be distributed over a large area to prevent pressure sores ($P < 0.50 \text{ N/mm}^2$ (P. W. Brand & Hollister, 1999))
... no cutting	No sharp edges
... no choking	No loose parts that can be swallowed
... no constrictions	ISO standards
6 The brace can be completely removed for short periods of time	Allowing bathing, diaper changes, etc. (Nogueira et al., 2013)
Donning and doffing is fast	Brace can be taken off and be put back on within 60sec
Donning and doffing is easy	Intuitive design
Not necessary to remove the brace for diaper changes	
7 Using the brace is a happy experience	
Soft feel	
Friendly appearance	No metal parts visible
Friendly colors	
8 The brace can be used by most clubfoot patients	Idiopathic clubfoot, younger than 6 months old
The brace can be adjusted for size	7 – 14 cm
The brace can be adjusted for severity	Pirani: 1.0 – 6.0
The brace can be adjusted for stiffness	Diméglio: 5 – 20

the design process and help with the construction of prototypes. Below the three most realistic concepts are presented.

8.3.1 Concept: Soft robotics

While in traditional robotics movement is initiated through to mechanical interaction of components, in soft robotics movement is initiated through thermal, electrical, physical or chemical properties of materials and interaction between materials. This principle was used in a dynamic clubfoot brace in the shape of a actuated rubber finger, that bends and presses on the foot when filled with a fluid. The fluid is under constant pressure in a container attached to the flexible finger. Because of the constant pressure, as the position of the clubfoot changes, the finger bends further, co-adapting to the position of the clubfoot.

Figure 8.2 | Soft robotics. Prototyping by Esther Keulers and Baat Medical.

PROS	CONS
Soft touch	Slow development
Safe	Limited range of motion

8.3.2 Concept: One part

Set with the challenge to create a dynamic clubfoot brace that could be made with additive manufacturing (3D printing), a one-part, monomaterial concept was created. Flexure hinges and negative stiffness were used to construct a mechanism that would apply a torque on the forefoot with the head of the talus as fulcrum.

Figure 8.3 | Concept one-part. Prototyping by Annelies de Jong

PROS	CONS
3D printing	Fixed axis of rotation
Simple design	No soft touch
Adjustable	

8.3.3 Concept: Spiral

Inspired by the design of Cunningham (Cunningham, 2017) a spiral-shaped brace was made that applies a complex moment on the foot without any fixed axes. The brace has a shell that is attached to the upper leg and the foot with strips of Velcro.

PROS	CONS
Elegant design	Unsure if the corrections are applied in the proper order
Attractive appearance	
Free axis of rotation	

8.4 Reflection

All concepts have elements that show the potential for a brace as a new treatment method for clubfoot. Several approaches are presented to apply a constant force on the clubfoot to treat the deformity. The selection for a best concept is outside the scope of this thesis.

A common feature in all ideas and concepts is that the brace must have a soft look and feel, which is reflected in the choice of materials for all concepts. Many parents have mentioned that they had missed cuddling with their child during the casting phase due to the rough feeling of the plaster casts, a soft brace allows for better cuddling.

8.4.1 Brace compliance

In the Ponseti method after the casting phase an abduction brace is worn for several years to prevent relapse. Brace non-compliance is known to be the biggest predictor for relapse of the clubfoot deformity (Dobbs et al., 2004). That is, if parents fail to put the brace on long enough the foot tends to get back into the typical clubfoot shape. In this phase of the treatment the parents are made responsible for the success of the treatment.

As part of her master track in Human-Technology Relations, Jonne van Belle studied the new relation in the treatment of clubfoot with a clubfoot brace rather than with plaster casts (van Belle, 2017). In the current treatment with plaster casts the parents have a relative passive relation to the treatment: together with their child they are subjected to the treatment and have little influence on the success of the treatment. Even if it seems like the child does not like the cast, parents simply cannot remove the plaster cast without braking it and thereby stopping the treatment. This relation changes when a dynamic clubfoot brace is used that can be taken off. Now the parents get a more active relation to the

treatment: they are responsible for putting the brace on correctly and for the right duration.

It might be hard for parents to resist the urge to take the brace off if their child seems to be uncomfortable with it. To help the parents in this responsibility to successfully treat the clubfoot of their child one can think of strict surveillance or alarms, but Jonne proposed three more elegant concepts to focus on the active participation of the parents (FIGURE 8.4). The concept of an item with matching appearance that caregivers should wear as a constant reminder of their responsibility seems most elegant and can already be tested in an existing brace where compliance is an issue.

8.4.2 Roadmap

We expect that for many practitioners the change from plaster to a dynamic clubfoot brace might be too large, i.e. the direct introduction of a dynamic brace would be unfeasible. Smaller steps are required and therefore a roadmap has been proposed for the gradual acceptance of an alternative treatment of clubfoot. Not all steps might be absolutely necessary and could be done in parallel, but the general order of steps should be followed to ensure good acceptance of the product as a new treatment method.

- Weekly cast changes (current)
- Twice weekly cast changes (current)
- Daily cast changes
- Serial static splinting (such as Chanu et al. (2016))
- Static progressive splinting
- Custom-made dynamic clubfoot brace
- Adjustable dynamic clubfoot brace

(a) Parent participation

(b) Baby toy

(c) Keep it close

Figure 8.4 | Concepts. Ideas to stimulate parents to take responsibility in the treatment of their child by making sure that the brace is always on when possible. Concepts made by Jonne van Belle.

8.5 Conclusions and future work

Technically a dynamic brace for the treatment of clubfoot is feasible. Since it is challenging to find the exact requirements for the magnitude and direction of the forces and torques, a clinical trial with a prototype in which the applied forces are monitored might be the most efficient and effective approach. In addition to this clinical trial, more time is needed for the public view on the Ponseti method to become more critical and to acknowledge the downsides of the method, creating room for the acceptance of a new dynamic clubfoot brace.

Chapter 9

General discussion

9.1 Introduction

Although the Ponseti method is highly effective in the treatment of clubfoot, there are many clues indicating that the Ponseti method is in need of a correction. The plaster casts of the Ponseti method prevents bathing, gets in the way with diaper changes and interferes with cuddling. Additionally the casts appear to cool the child's feet, to provoke judgment from others and its position-controlled approach is likely to be inefficient. In our experience these issues are unknown to most treating physicians and easily dismissed. Several studies have indicated the need for medical professionals to acknowledge that the Ponseti method causes increased stress for the families (Malagelada et al., 2016; Nogueira et al., 2013). For some caregivers these issues are real and require a solution.

In the correction process of deformations and contractures often a position-controlled approach is used (e.g. static bracing for scoliosis, plaster cast for clubfoot) while our philosophy is that the use of a force-controlled approach is more efficient and experienced as more comfortable (e.g. the *TriaC* brace for

scoliosis, dynamic braces for stiff fingers (Glasgow, Tooth, & Fleming, 2008)). The hypothesis that the correction of a clubfoot can be done with the application of a small but constant force has been studied in this thesis and conceptualized in a dynamic clubfoot brace.

9.2 Interpretation of findings

Only few studies have looked at the parents' perspective in the treatment of clubfoot (Dibello et al., 2016; Malagelada et al., 2016; Nogueira et al., 2013; Pletch et al., 2015). It was found that the casting phase is the most difficult for families (Malagelada et al., 2016; Nogueira et al., 2013). Our own questionnaire (CHAPTER 8) had similar findings to that of Pletch et al. (2015), showing the impact of the plaster casts on hygiene, cuddling and judgment. Complaints from parents about wet and cold casts were supported with temperature measurements under the cast (CHAPTER 7). It is clear that the lack of comfort of the Ponseti method can be a problem for many parents and caregivers.

When looking at the description of the Ponseti method, one of the first questions any engineer would ask is *“how hard do you press on the foot?”*. A treating physician would only be able to give subjective answers like *“as hard as necessary”* without giving any numbers. It might therefore be no surprise that we found large variations in applied forces on a clubfoot model (CHAPTER 3). With the numbers available we still do not have any clearer answer to this question than probably less than 1N, but we do have a new insight in the functioning of the Ponseti method: considering that all participants in this study had years of experience and are very well able to treat a clubfoot, and that the measured forces varied largely, we can assume that the exact magnitude of the applied force might not be relevant.

Our study with force sensors under the plaster cast in the treatment of clubfoot (CHAPTER 6) showed that the applied force decreased rapidly and reached a

plateau within several hours. Could it be that the exact magnitude of the applied force seems irrelevant because it decreases so rapidly?

It is clear that it is unnecessary to wait a full week to change the plaster cast on a clubfoot. Our review has provided strong evidence that an accelerated protocol can be safe and effective (CHAPTER 2). Our force measurements described in CHAPTER 6 give a reason for this finding and suggest that even shorter cast change intervals should be possible: the applied forces reach a stable low value within several hours. A randomized controlled trial with weekly, twice weekly, 2-days and daily intervals is needed to find an optimal cast change interval. We hypothesize that from a mechanobiological perspective the optimum is less than 24h, but feel that for serial casting a daily interval would be a more practical limit.

9.3 Clubfoot mechanics

To our knowledge the mechanobiological principles behind the Ponseti method are hardly studied, but a general description can be found in *Lovell and Winter's Pediatric Orthopaedics*: “*The efficacy of manipulation and casting is attributable to the viscoelastic, or rate-dependent, behavior of the ligaments and tendons. **The creep phenomenon occurs during manipulation.** A constant stress (force per unit area), or load, is applied. The tissues undergo deformation (strain) relatively quickly at first, then more slowly, until there is little additional elongation possible at that time. The cast is applied with the same load used during manipulation. **Stress relaxation takes place in the cast.** The tissues undergo a small additional amount of elongation in the cast. As soon as that occurs, the stress on the tissues effectively decreases, and will not increase again until the next manipulation.*” (Mosca, 2001).

From a mechanical perspective, the Ponseti method consists of three parts.

9.3.1 Manipulation

Cyclic loading of the foot is performed to precondition the tissues for correction (Cohen, 2015). It consists of the application of a series of constant forces which induces a creep response. The Ponseti method prescribes a manipulation time of 1 minute (Staheli, 2009) while R. J. Xu (2011) experimented with a longer manipulation time of 2 minutes and in his book, Ponseti (2008) recommends 2–3 minutes. For older patients Loureno and Morcuende (2007) used 5–10 minutes to allow “greater stretching of the soft tissues which are less elastic”. Cohen (2015) found that in her specimens of medial clubfoot tissue preconditioning convergence already began after 8 cycles. In the treatment of other contractures a manipulation time of 10–15 minutes is often prescribed (Bell-Krotoski & Figarola, 1995; Flowers & LaStayo, 1994; McClure et al., 1994).

It is interesting to see that for other contractures the manipulation time is up to ten times longer than for clubfeet. Could this mean that the creep strain rate (i.e. how fast deformation results from the applied stress) for clubfoot tissue is much higher than tissue in contractures in adults? Or does this mean that clubfoot treatment could benefit from a longer preconditioning time?

9.3.2 Plaster casting

The foot is held at a constant position for a week, translating to a constant elongation (stress relaxation) of the soft tissue of the clubfoot (Cohen, 2015). Cohen (2015) found that the medial fibrotic tissue of the clubfoot had relaxed by approximately 30% after 100 seconds of constant strain.

Although the process of tissue adaptation is well described with microscale models, its effect on macroscale has seldomly been measured. Strik (2000) measured the force of a plaster cast on the skin in contractures on several joints in spastic patients and observed clear stress-relaxation behavior, reaching 50% re-

laxation in 4 hours. Our own measurements in clubfeet (CHAPTER 6) show the same pattern of stress-relaxation with 95% relaxation in several hours. Several models are used to describe the stress-relaxation function $G(t)$ tissue, often consisting of two or three exponential elements, one describing the “fast” response and one describing the “slow” response (Anssari-Benam, Bucchi, & Bader, 2015; Screen, Toorani, & Shelton, 2013; B. Xu, Li, & Zhang, 2013). Because a variation of models is used, a comparison of stress-relaxation rate is not intuitive.

We conclude that the microscale (tissues) behavior of stress-relaxation is visible on macroscale (joints) as a similar effect. *In vivo* measurements studying the adaptation of the involved tissues are necessary to bridge the gap between both scales. Cohen (2015) suggests the use of fluoroscopy, ultrasound and elastography to get a better understanding of the exact mechanics of the Ponseti method and the treatment of contractures in general.

9.3.3 Bracing

The function of the bracing period is to maintain the correction that is achieved in the manipulation and casting phase. Not adequately wearing the brace severely increases the risk of a relapse (Dobbs et al., 2004). This indicates that although the position of the foot is corrected, its internal tissues might still have the tendency to create a new deformity. The bracing phase has not been studied in this PhD project.

9.4 Limitations

9.4.1 User input

Clubfoot patients are too young to clearly express their feelings. It is therefore not possible to ask patients about their opinion on the current treatment method

or possible improvements. Caregivers act as a proxy and indicate that their children can be a bit agitated after a new cast is applied, but the exact cause for this agitated feeling is a guess. Still, the treatment method also affects the caregivers and they are seen as the primary 'users' in this scenario. Asking their opinion is therefore justified.

As a pilot study, an online questionnaire was used to get the opinion of caregivers on several topics of the treatment of their child, discussed in CHAPTER 8. However, no scientifically sound survey was performed to get a better understanding of the user opinion. Such a survey was fully prepared and planned but interfered with the agenda of the WKO (*Werkgroep Kinder Orthopedie*) and was halted by the WKO.

The plaster phase is a relative short one during a stressful and exhausting period for the new parents. Combined with the few clubfeet our local hospital treats, it has been challenging to find available people that could act as potential users. As such, attempts to gather more user input — such as home visits to observe the existing problems with the plaster casts in real life — failed. In the design of any product it is paramount to understand what the user wants (Eger, Bonnema, Lutters, & van der Voort, 2012) and this we still do not. It is clear that there are problems but not how they should be solved.

The following example illustrates our experience with improper understanding of the users' needs. Soiling of the plaster cast is mentioned as one of the difficulties of the Ponseti method (Patel, Jeer, & Cornell, 2007). In an attempt to improve comfort, a design was made with part plaster cast, part brace (FIGURE 9.1). The lower leg would be wrapped in plaster cast as normal procedure and the upper leg would be held with a plastic shell that could be temporarily removed to allow easy diaper changes and prevent soiling of the cast. When the responders of the online questionnaire were asked for their opinion on this concept, there were no signs that anyone would prefer this solution over the standard Ponseti casts. We do not question our assumption that soiling of the cast is a problem, but without

Figure 9.1 | Hybrid concept. The lower part of the clubfoot is wrapped in plaster cast, as usual, with a plastic insert included in the cast. Attached to this insert is a shell that can be temporarily removed to make diaper changes easier.

the input of users any solution is just a shot in the dark. Apparently our proposed design was not the solution users were waiting for.

One approach to this situation with missing user input is to invite parents during the treatment of their child to join a focus group. In sessions with parents different problems can be discussed and together a solution could be developed. Future work should focus on co-design to create realistic solutions (Sanders & Stappers, 2008).

9.4.2 Clinical trial

As with the development of any biomedical product, in this project the development of a dynamic clubfoot brace has been slow. Several concepts were made and functional prototypes were made that would be suitable for a future clinical trial. This clinical trial is important to get valuable feedback from both caregivers and treating physicians.

We feel that in this project the prototype-users relation has been a circular problem: without a good prototype we could not do a trial, without feedback from a trial we could not improve the prototype. To break from this circle

attempts were made to construct a functional model of a clubfoot. Such a model would allow us to see how the brace manipulates the foot and make a decision if it would be acceptable to give it to children. Many researchers have tried to come up with a good model for a clubfoot, but so far no functional model exists (DiMeo, Lalush, Grant, & Morcuende, 2012; Jain, Dhande, & Vyas, 2009; Khas, Pandey, & Ray, 2013; Mahmoodian, 2010; Windisch et al., 2007). Despite our efforts, we too have not been able to create model to predict the response of a clubfoot when subjected to external forces.

It requires much courage to start testing a prototype on babies without being able to validate the prototype on a functional model first. With the Ponseti method being well established as the treatment method of choice it is challenging to deviate from this with a new method. For many treating physicians and most caregivers the Ponseti method is new and a welcome substitute for the surgical methods. There is no demand-pull for a better method, rendering our current attempts with the dynamic clubfoot brace a technology-push, a strategy that can be much slower (Nemet, 2009). We feel that what is needed most in the process of the development of a dynamic clubfoot orthosis, is time. Over time attractive attributes (in this case the Ponseti method being non-invasive) become one-dimensional attributes and then further on become basic needs (Shen, Tan, & Xie, 2000). In time the newness of the Ponseti method will fade and physicians and caregivers will be more critical to the downsides of it, creating room for the acceptance of a new treatment method such as a dynamic clubfoot brace.

9.5 Conclusion

The Ponseti method is the golden standard in the treatment of clubfoot. However, for the caregivers of children with a clubfoot the method has several important downsides. The plaster casts make it impossible to bath the child, interfere with cuddling and provoke judgment from outsiders. Temperature measurements show

that the drying of the casts has a cooling effect on the child's legs.

There is no need to wait one full week to change the cast. There is strong evidence that the cast change interval can safely be shortened and force measurements suggest that even daily cast changes would be effective.

The concept of a dynamic clubfoot brace could provide a solution for the problems caregivers face with the plaster casts, as well as for the inefficiency of the position-controlled approach of the static plaster casts. This brace could be temporarily removed to allow bathing, would be soft to the touch and have a appealing appearance, while applying a constant low force on the foot to efficiently correct the clubfoot deformity.

References

- Alves, C., Escalda, C., Fernandes, P., Tavares, D., & Neves, C. (2009). Ponseti method does age at the beginning of treatment make a difference? [Journal Article]. *Clinical orthopaedics and related research*, 467(5), 1271-1277. doi: 10.1007/s11999-008-0698-1
- Anssari-Benam, A., Bucchi, A., & Bader, D. L. (2015). Unified viscoelasticity: Applying discrete element models to soft tissues with two characteristic times [Journal Article]. *Journal of biomechanics*, 48(12), 3128-3134.
- Awang, M., Sulaiman, A. R., Munajat, I., & Fazliq, M. E. (2014). Influence of age, weight, and pirani score on the number of castings in the early phase of clubfoot treatment using ponseti method [Journal Article]. *The Malaysian journal of medical sciences: MJMS*, 21(2), 40-43.
- Aydin, B. K., Sofu, H., Senaran, H., Erkocak, O. F., Acar, M. A., & Kirac, Y. (2015). Treatment of clubfoot with ponseti method using semirigid synthetic softcast [Journal Article]. *Medicine*, 94(47), e2072. doi: 10.1097/MD.0000000000002072
- Baindurashvili, A., Kenis, V., & Stepanova, Y. (2012). Soft-tissue complications during treatment of children with congenital clubfoot [Journal Article]. *EWMA Journal*, 12(3).
- Bell-Krotoski, J. A., & Figarola, J. H. (1995). Biomechanics of soft-tissue growth and remodeling with plaster casting [Journal Article]. *Journal of Hand Therapy*, 8(2), 131-137.

- Besselaar, A. T., Kamp, M. C., Reijman, M., & van der Steen, M. C. (n.d.). Incidence of congenital idiopathic clubfoot in the netherlands. *JPO B*. (Accepted for publication)
- Besselaar, A. T., Sakkers, R. J. B., Schuppers, H. A., Witbreuk, M. M. E. H., Zeegers, E. V. C. M., Visser, J. D., ... Burger, K. N. J. (2017). Guideline on the diagnosis and treatment of primary idiopathic clubfoot. *Acta Orthopaedica*, 88(3), 305-309. Retrieved from <https://doi.org/10.1080/17453674.2017.1294416> (PMID: 28266239) doi: 10.1080/17453674.2017.1294416
- Böhm, S., & Sinclair, M. (2013). Report of the 1st european consensus meeting on ponseti clubfoot treatment [Journal Article]. *Journal of Children's Orthopaedics*, 7(3), 251-254. doi: 10.1007/s11832-013-0497-4
- Bonifas, A., Kim, J., & Suh, J. (2015, 09 3). *Force responsive inductors for force sensors* (Patent No. WO 2015130610 A1). Google Patents. Retrieved from <https://google.com/patents/WO2015130610A1> (WO Patent App. PCT/US2015/017,096)
- Brand, P. W., & Hollister, A. M. (1999). *Clinical mechanics of the hand* (3rd ed.) [Book]. St. Louis, Missouri, USA: Mosby, Inc.
- Brand, R., Siegler, S., Pirani, S., Morrison, W., & Udupa, J. (2006). Cartilage anlagen adapt in response to static deformation [Journal Article]. *Medical hypotheses*, 66(3), 653-659. doi: 10.1016/j.mehy.2005.10.006
- Brauns, A., Nuffel, M. V., Smet, L. D., & Degreef, I. (2017). A clinical trial of tension and compression orthoses for Dupuytren contractures. *Journal of Hand Therapy*, 30(3), 253 - 261. Retrieved from <http://www.sciencedirect.com/science/article/pii/S0894113016302368> doi: 10.1016/j.jht.2016.11.011
- Bridgens, J., & Kiely, N. (2010). Current management of clubfoot (congenital talipes equinovarus) [Journal Article]. *BmJ*, 340(6), 308-312. doi: 10.1136/bmj.c355
- Burghardt, R. D., Anderson, J. G., Reed, R. A., & Herzenberg, J. E. (2014).

- Exothermic properties of plastersynthetic composite casts [Journal Article]. *Journal of children's orthopaedics*, 8(2), 193-201. doi: 10.1007/s11832-014-0563-6
- Carr, J. J. (1993). *Sensors and circuits: sensors, transducers, and supporting circuits for electronic instrumentation, measurement, and control* [Book]. Englewood Cliffs, New Jersey, USA: PTR Prentice-Hall, Inc.
- Carroll, N. (1990). Clubfoot [Book Section]. In W. W. Lovell, R. T. Morrissy, & R. B. Winter (Eds.), *Lovell and winter's pediatric orthopaedics* (Vol. 2, p. 927956). Philadelphia: Lippincott Williams & Wilkins.
- Chanu, A. R., Bimol, N., Rai, B., Wangjam, K., & Singh, N. R. (2016). Manipulation and orthosis in the management of congenital talipes equinovarus [Journal Article]. *Journal of Medical Society*, 30(1), 27.
- Chojnowski, A., Wach, W., & Degreef, I. (2017). Controversy: Splinting for dupuytren contracture. In P. Werker, J. Dias, C. Eaton, B. Reichert, & W. Wach (Eds.), *Dupuytren disease and related diseases - the cutting edge* (pp. 323-329). Switzerland: Springer International Publishing. doi: 10.1007/978-3-319-32199-8_44
- Chu, A., Labar, A. S., Sala, D. A., van Bosse, H. J., & Lehman, W. B. (2010). Clubfoot classification: correlation with ponseti cast treatment [Journal Article]. *Journal of Pediatric Orthopaedics*, 30(7), 695-699.
- Cohen, T. L. (2015). *Biomechanical characterization and evaluation of conservative clubfoot correction* (PhD Thesis, Marquette University). Retrieved from https://epublications.marquette.edu/dissertations_mu/603/
- Cohen, T. L., Altiok, H., Tarima, S., Smith, P. A., & Harris, G. F. (2012, aug). Creep evaluation of (orthotic) cast materials during simulated clubfoot correction. In *34th annual international conference of the ieee engineering in medicine and biology society* (p. 3352-3355). IEEE. doi: 10.1109/EMBC.2012.6346683
- Cohen, T. L., Altiok, H., Wang, M., McGrady, L. M., Tarima, S., Krzak, J., ... Harris, G. F. (2013). Evaluation of cast creep occurring during simulated

- clubfoot correction [Journal Article]. *Proceedings of the Institution of Mechanical Engineers, Part H: Journal of Engineering in Medicine*, 227(8), 919-927. doi: 10.1177/0954411913485044
- Coleman, S., Nixon, J., Keen, J., Wilson, L., McGinnis, E., Dealey, C., ... Nelson, E. A. (2014). A new pressure ulcer conceptual framework [Journal Article]. *Journal of Advanced Nursing*, 70(10), 2222-2234. Retrieved from <http://dx.doi.org/10.1111/jan.12405> doi: 10.1111/jan.12405
- Collis, J., Collocott, S., Hing, W., & Kelly, E. (2013). The effect of night extension orthoses following surgical release of dupuytren contracture: A single-center, randomized, controlled trial. *The Journal of Hand Surgery*, 38(7), 1285-1294.e2. doi: 10.1016/j.jhssa.2013.04.012
- Cooper, D. M., & Dietz, F. R. (1995). Treatment of idiopathic clubfoot. a thirty-year follow-up note [Journal Article]. *Journal of Bone & Joint Surgery - American Volume*, 77(10), 1477-1489. Retrieved from <https://www.ncbi.nlm.nih.gov/pubmed/7593056>
- Cosma, D. I., & Vasilescu, D. E. (2014). Ponseti treatment for clubfoot in romania: a 9-year single-centre experience [Journal Article]. *Journal of Pediatric Orthopaedics B*, 23(6), 512-516. Retrieved from http://journals.lww.com/jpo-b/Fulltext/2014/11000/Ponseti_treatment_for_clubfoot_in_Romania___a.4.aspx doi: 10.1097/bpb.0000000000000081
- Cummings, R. J., Davidson, R. S., Armstrong, P. F., & Lehman, W. B. (2002). Congenital clubfoot [Journal Article]. *The Journal of Bone & Joint Surgery*, 84(2), 290-308.
- Cunningham, J. (2017). *Clubfoot bracing - cunningham prosthetic care*. <http://www.cunninghamprostheticcare.com/clubfoot-bracing/>. (Accessed: 2018-07-17)
- Cyr, L. M., & Ross, R. G. (1998). How controlled stress affects healing tissues [Journal Article]. *Journal of Hand Therapy*, 11(2), 125-130. Retrieved from <http://www.sciencedirect.com/science/article/pii/>

- S0894113098800094 doi: 10.1016/S0894-1130(98)80009-4
- Davis, F., & De Vita, R. (2012). A nonlinear constitutive model for stress relaxation in ligaments and tendons [Journal Article]. *Annals of Biomedical Engineering*, 40(12), 2541-2550. Retrieved from <http://dx.doi.org/10.1007/s10439-012-0596-2> doi: 10.1007/s10439-012-0596-2
- de Hemptinne, C., Swann, N. C., Ostrem, J. L., Ryapolova-Webb, E. S., San Luciano, M., Galifianakis, N. B., & Starr, P. A. (2015). Therapeutic deep brain stimulation reduces cortical phase-amplitude coupling in parkinson's disease [Journal Article]. *Nature Neuroscience*, 18, 779-786. doi: 10.1038/nn.3997
- de Morree, J. J. (2014). *Dynamiek van het menselijk bindweefsel. functie, beschadiging en herstel* (6th ed.) [Book]. Houten: Bohn Stafleu van Loghum. Retrieved from <http://books.google.nl/books?id=xgBX5hqU5RoC>
- Derzsi, Z., Nagy, r., Gozar, H., Gurzu, S., & Pop, T. S. (2015). Kite versus ponseti method in the treatment of 235 feet with idiopathic clubfoot: Results of a single romanian medical center [Journal Article]. *Medicine*, 94(33).
- Dibello, D., Di Carlo, V., & Carbone, M. (2016, September). The ponseti method for the treatment of congenital clubfoot: the parental opinion. In *Proc. 37th sicot orthopaedic world congress (sicot 2016)* (p. 245). SICOT. Retrieved from <http://www.sicot.org/sites/default/files/images/Rome/Abstract-Book-Free-Papers.pdf> (43791)
- Diméglio, A., Bensahel, H., Souchet, P., Mazeau, P., & Bonnet, F. (1995). Classification of clubfoot [Journal Article]. *Journal of Pediatric Orthopaedics B*, 4(2), 129-136. Retrieved from http://journals.lww.com/jpo-b/Fulltext/1995/04020/Classification_of_Clubfoot_2.aspx
- DiMeo, A. J., Lalush, D. S., Grant, E., & Morcuende, J. A. (2012). Development of a surrogate biomodel for the investigation of clubfoot bracing [Journal Article]. *Journal of Pediatric Orthopaedics*, 32(7), e47-e52.
- Dobbs, M. B., & Gurnett, C. A. (2012). Genetics of clubfoot [Journal Article]. *Journal of Pediatric Orthopaedics B*, 21(1), 7-9. Re-

- trieved from https://journals.lww.com/jpo-b/Fulltext/2012/01000/Genetics_of_clubfoot.3.aspx doi: 10.1097/BPB.0b013e328349927c
- Dobbs, M. B., Rudzki, J., Purcell, D. B., Walton, T., Porter, K. R., & Gurnett, C. A. (2004). Factors predictive of outcome after use of the ponseti method for the treatment of idiopathic clubfeet [Journal Article]. *The Journal of Bone & Joint Surgery*, 86(1), 22-27.
- Dyer, P., & Davis, N. (2006). The role of the pirani scoring system in the management of club foot by the ponseti method [Journal Article]. *Journal of Bone & Joint Surgery-British Volume*(88), 1082-4.
- Eastwood, D. M. (2009). The clubfoot: congenital talipes equinovarus [Book Section]. In *Children's orthopaedics and fractures* (p. 541-558). Springer.
- Eger, A., Bonnema, M., Lutters, E., & van der Voort, M. (2012). *Product design* (1st ed.). Amsterdam: Boom Uitgevers.
- Elgohary, H. S., & Abulsaad, M. (2015). Traditional and accelerated ponseti technique: a comparative study [Journal Article]. *European Journal of Orthopaedic Surgery & Traumatology*, 1-5. doi: 10.1007/s00590-015-1594-5
- Ellis, B., & Bruton, A. (2002). A study to compare the reliability of composite finger flexion with goniometry for measurement of range of motion in the hand. *Clinical Rehabilitation*, 16(5), 562-570. Retrieved from <http://journals.sagepub.com/doi/abs/10.1191/0269215502cr513oa> doi: 10.1191/0269215502cr513oa
- FAO. (2001, October). *Human energy requirements. report of a joint fao/who/unu expert consultation* (Report). Rome: Food and Agriculture Organization of the United Nations (FAO).
- Farmer, S. E., & James, M. (2009). Contractures in orthopaedic and neurological conditions: a review of causes and treatment. *Disability and Rehabilitation*(13), 549-558. doi: 10.1080/09638280010029930
- Ferguson-Pell, M., Hagisawa, S., & Bain, D. (2000). Evaluation of a sensor for low interface pressure applications. *Medical Engineering & Physics*, 22(9), 657 - 663. Retrieved from <http://www.sciencedirect.com/science/article/>

- pii/S1350453300000801 doi: 10.1016/S1350-4533(00)00080-1
- Flowers, K. R. (2002). A proposed decision hierarchy for splinting the stiff joint, with an emphasis on force application parameters. *Journal of Hand Therapy*, 15(2), 158–162. Retrieved from <http://www.sciencedirect.com/science/article/pii/S0894113002500302> doi: 10.1053/hanthe.2002.v15.015015
- Flowers, K. R., & LaStayo, P. (1994). Effect of total end range time on improving passive range of motion [Journal Article]. *Journal of Hand Therapy*, 7(3), 150-157.
- Foissac, R., Camuzard, O., Dumas, P., Dumontier, C., & Chignon-Sicard, B. (2013). Traitement des brides de la maladie de dupuytren par la collagnase injectable. *Chirurgie de la Main*, 32(4), 199 - 205. Retrieved from <http://www.sciencedirect.com/science/article/pii/S1297320313000784> doi: <https://doi.org/10.1016/j.main.2013.05.004>
- Frick, S. L. (2009). Fundamentals of clubfoot treatment: understanding some critical principles needed for success. In *29th annual oscar miller day symposium*.
- Frost, H. M. (1987). Bone mass and the mechanostat: a proposal [Journal Article]. *The anatomical record*, 219(1), 1-9.
- Frost, H. M. (2003). Bone's mechanostat: A 2003 update. *The Anatomical Record Part A: Discoveries in Molecular, Cellular, and Evolutionary Biology*, 275A(2), 1081-1101. Retrieved from <https://onlinelibrary.wiley.com/doi/abs/10.1002/ar.a.10119> doi: 10.1002/ar.a.10119
- Fung, Y.-C. (1981). *Biomechanics: Mechanical properties of living tissues* (1st ed.) [Book]. Springer-Verlag.
- Funk, J. F. (2018). Klumpfußbehandlung in der täglichen Praxis ist Ponseti alternativlos? *Orthopädische und Unfallchirurgische Praxis*, 7(8), 426-432. doi: 10.3238/oup.2016.04260432
- Giesberts, R. B. (2016). *Mbed: Inductive_sensor_3* [Web page]. https://os.mbed.com/users/bobgiesberts/code/Inductive_Sensor_3.

- Retrieved 2018-02-01, from https://os.mbed.com/users/bobgiesberts/code/Inductive_Sensor_3
- Giesberts, R. B., Hekman, E. E. G., Maathuis, P. G. M., & Verkerke, G. J. (2017). Quantifying the ponseti method [Journal Article]. *Journal of the Mechanical Behavior of Biomedical Materials*, 66, 45-49. doi: 10.1016/j.jmbbm.2016.10.021
- Giesberts, R. B., Hekman, E. E. G., Verkerke, G. J., & Maathuis, P. G. M. (2018a). *Rapid decrease of cast-induced forces in the Ponseti method* [Unpublished Work]. (Accepted for publication in Bone and Joint Journal (21-Sept 2018))
- Giesberts, R. B., Hekman, E. E. G., Verkerke, G. J., & Maathuis, P. G. M. (2018b). Ten cold clubfeet [Journal Article]. *Acta Orthopaedica*, 0(0), 1-5. (PMID: 29985745) doi: 10.1080/17453674.2018.1493046
- Giesberts, R. B., Sluiter, V. I., & Verkerke, G. J. (2018). Design and test of a new inductive force sensor [Technical Note]. *Sensors*, 18(7). doi: 10.3390/s18072079
- Giesberts, R. B., van der Steen, M. C., Maathuis, P. G. M., Besselaar, A. T., Hekman, E. E. G., & Verkerke, G. J. (2018, 06). Influence of cast change interval in the Ponseti method: a systematic review. *PLOS One*, 13(6), 1-12. doi: 10.1371/journal.pone.0199540
- Gilani, S. F., Ahmed, S., Obaid-Ur-Rehman, & Bashir, M. A. (2014). Comparison of accelerated vs standard ponseti method in management of idiopathic clubfoot [Journal Article]. *Journal of Pakistan Orthopaedic Association*, 26(1), 7-10.
- Gilpin, D., Coleman, S., Hall, S., Houston, A., Karrasch, J., & Jones, N. (2010). Injectable collagenase clostridium histolyticum: A new nonsurgical treatment for dupuytren's disease. *The Journal of Hand Surgery*, 35(12), 2027 - 2038.e1. doi: 10.1016/j.jhssa.2010.08.007
- Glasgow, C., Fleming, J., Tooth, L. R., & Hockey, R. L. (2012). The long-term relationship between duration of treatment and contracture resolution

- using dynamic orthotic devices for the stiff proximal interphalangeal joint: a prospective cohort study [Journal Article]. *Journal of Hand Therapy*, 25(1), 38-47.
- Glasgow, C., Tooth, L., & Fleming, J. (2008). Which splint? dynamic versus static progressive splinting to mobilise stiff joints in the hand [Journal Article]. *The British Journal of Hand Therapy*, 13(4), 104-110. Retrieved from <http://hth.sagepub.com/content/13/4/104.abstract>
- Glasgow, C., Tooth, L. R., & Fleming, J. (2010). Mobilizing the stiff hand: combining theory and evidence to improve clinical outcomes [Journal Article]. *Journal of Hand Therapy*, 23(4), 392-401.
- Glasgow, C., Wilton, J., & Tooth, L. (2003). Optimal daily total end range time for contracture: Resolution in hand splinting [Journal Article]. *Journal of Hand Therapy*, 16(3), 207-218.
- Goldstein, R. Y., Seehausen, D. A., Chu, A., Sala, D. A., & Lehman, W. B. (2014). Predicting the need for surgical intervention in patients with idiopathic clubfoot [Journal Article]. *Journal of pediatric orthopedics*.
- Gurnett, C. A., Alaei, F., Kruse, L. M., Desruisseau, D. M., Hecht, J. T., Wise, C. A., ... Dobbs, M. B. (2008). Asymmetric lower-limb malformations in individuals with homeobox *hox11* gene mutation [Journal Article]. *The American Journal of Human Genetics*, 83(5), 616-622. doi: 10.1016/j.ajhg.2008.10.004
- Gurnett, C. A., Boehm, S., Connolly, A., Reimschisel, T., & Dobbs, M. B. (2008). Impact of congenital talipes equinovarus etiology on treatment outcomes [Journal Article]. *Developmental Medicine & Child Neurology*, 50(7), 498-502.
- Halanski, M. A., Halanski, A. D., Oza, A., Vanderby, R., Munoz, A., & Noonan, K. J. (2007). Thermal injury with contemporary cast-application techniques and methods to circumvent morbidity [Journal Article]. *JBJS*, 89(11), 2369-2377. doi: 10.2106/JBJS.F.01208
- Harnett, P., Freeman, R., Harrison, W. J., Brown, L. C., & Beckles, V. (2011).

- An accelerated ponseti versus the standard ponseti method: a prospective randomised controlled trial [Journal Article]. *Journal of Bone & Joint Surgery, British Volume*, 93-B(3), 404-408. Retrieved from <http://www.bjj.boneandjoint.org.uk/content/93-B/3/404.abstract> doi: 10.1302/0301-620x.93b3.24450
- Hattori, K., Sano, H., Saijo, Y., Kita, A., Hatori, M., Kokubun, S., & Itoi, E. (2007). Measurement of soft tissue elasticity in the congenital clubfoot using scanning acoustic microscope [Journal Article]. *Journal of Pediatric Orthopaedics B*, 16(5), 357-362.
- Hayter, F. (2015). *Casting directives* [Presentation]. Retrieved 2018-03-07, from <http://slideplayer.com/slide/3475355/>
- Hennessey, T. A. (2012). Congenital clubfoot and the ponseti method: a review of recent literature [Journal Article]. *Current Orthopaedic Practice*, 23(5), 442-447.
- Herweijer, H., Dijkstra, P. U., Nicolai, J.-P. A., & van der Sluis, C. K. (2007). Postoperative hand therapy in dupuytren's disease [Journal Article]. *Disability and Rehabilitation*, 29(22), 1736-1741. doi: 10.1080/09638280601125106
- Herzenberg, J. E., Radler, C., & Bor, N. (2002). Ponseti versus traditional methods of casting for idiopathic clubfoot [Journal Article]. *Journal of Pediatric Orthopaedics*. Retrieved from https://journals.lww.com/pedorthopaedics/Fulltext/2002/07000/Ponseti_Versus_Traditional_Methods_of_Casting_for.19.aspx
- Higgins, J. P., Altman, D. G., & Sterne, J. (2011). Chapter 8: Assessing risk of bias in included studies [Book Section]. In J. P. Higgins & S. Green (Eds.), *Cochrane handbook for systematic reviews of interventions* (Version 5.1.0 (updated March 2011) ed., chap. 8). The Cochrane Collaboration.
- Hollinger, A., & Wanderley, M. M. (2006). Evaluation of commercial force-sensing resistors. In *International conference on new interfaces for musical expression (NIME06)*. Paris, France: IRCAM. Retrieved from <http://www.idmil>

- .org/_media/publications/hollinger_wanderley_fsrevaluation.pdf
- Holubeva, N. (2015, June). *Configuring inductive-to-digital-converters for parallel resistance (R_p) variation in L-C tank sensors* (Application Report No. SNAA221A). Dallas, Texas: Texas Instruments. Retrieved from <http://www.ti.com/lit/an/snaa221a/snaa221a.pdf>
- Hueston, J. T. (n.d.). Dupuytren's contracture: The trend to conservatism: Hunterian lecture delivered at the royal college of surgeons of england on 8th october 1964. *Annals of The Royal College of Surgeons of England*, 36(3), 134-51. Retrieved from <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC2311814/pdf/annrcse00212-0008.pdf>
- Huisstede, B. M., Gladdines, S., Randsdorp, M. S., & Koes, B. W. (2017). Effectiveness of conservative, surgical, and postsurgical interventions for trigger finger, Dupuytren disease, and De Quervain disease: A systematic review. *Archives of Physical Medicine and Rehabilitation*. doi: 10.1016/j.apmr.2017.07.014
- Hutchinson, M. J., & Hutchinson, M. R. (2008). Factors contributing to the temperature beneath plaster or fiberglass cast material [Journal Article]. *Journal of orthopaedic surgery and research*, 3(10), 1. doi: 10.1186/1749-799X-3-10
- Hutton, B., Salanti, G., Chaimani, A., Caldwell, D. M., Schmid, C., Thorlund, K., ... Moher, D. (2014, 03). The quality of reporting methods and results in network meta-analyses: An overview of reviews and suggestions for improvement. *PLOS ONE*, 9(3), 1-9. Retrieved from <https://doi.org/10.1371/journal.pone.0092508> doi: 10.1371/journal.pone.0092508
- Ibraheem, G., Adegbehingbe, O., Babalola, O., Agaja, S., Ahmed, B., Olawepo, A., & Medubi, M. (2017). Evaluation of an accelerated ponseti protocol for the treatment of talipesequinovarus in nigeria [Journal Article]. *East and Central African Journal of Surgery*, 22(1), 28-38.
- Interlink Electronics. (n.d.). Fsr™force sensing resistors™, fsr™integration guide (Computer software manual No. 94-00004 Rev B). Retrieved from <https://>

www.rapidonline.com/pdf/182546_in_en_01.pdf

- Jain, M. L., Dhande, S. G., & Vyas, N. S. (2009). A biomechanical control mechanism for correction of clubfoot deformity in babies [Journal Article]. *International Journal of Biomechatronics and Biomedical Robotics*, 1(1), 51-56.
- Jerosch-Herold, C., Shepstone, L., Chojnowski, A. J., Larson, D., Barrett, E., & Vaughan, S. P. (2011). Night-time splinting after fasciectomy or dermo-fasciectomy for dupuytren's contracture: a pragmatic, multi-centre, randomised controlled trial. *BMC musculoskeletal disorders*, 12, 136-136. doi: 10.1186/1471-2474-12-136
- Jowett, C. R., Morcuende, J. A., & Ramachandran, M. (2011). Management of congenital talipes equinovarus using the ponseti method: a systematic review [Journal Article]. *Journal of Bone & Joint Surgery, British Volume*, 93-B(9), 1160-1164. Retrieved from <http://dx.doi.org/10.1302/0301-620x.93b9.26947> doi: 10.1302/0301-620x.93b9.26947
- Kannan, M., & Umashankar, R. (2018). Analysis of functional outcome of congenital talipes equinovarus by Ponseti method in our institution. *International Journal of Scientific Research*, 7(4), 61-63.
- Kaplan, S. (1981). Burns following application of plaster splint dressings. *The Journal of Bone and Joint Surgery*, 63(4), 670-672.
- Kasemsadeh, B., & LaPointe, L. (2016, June). *Inductive sensing touch-on-metal buttons design guide* (Application Report No. SNOA951). Dallas, Texas: Texas Instruments. Retrieved from <http://www.ti.com/lit/an/snoa951/snoa951.pdf>
- Kemler, M., Houpt, P., & van der Horst, C. (2012). A pilot study assessing the effectiveness of postoperative splinting after limited fasciectomy for dupuytren's disease. *Journal of Hand Surgery (European Volume)*, 37(8), 733-737. doi: 10.1177/1753193412437631
- Kent, A., Kecskes, Z., Shadbolt, B., & Falk, M. (2007). Normative blood pressure data in the early neonatal period [Journal Article]. *Pediatric Nephrology*,

- 22(9), 1335-1341. Retrieved from <http://dx.doi.org/10.1007/s00467-007-0480-8> doi: 10.1007/s00467-007-0480-8
- Khas, K. S., Pandey, P. M., & Ray, A. R. (2013). Rapid manufacturing of a clubfoot model imitating soft tissue and bones [Journal Article]. *Virtual and Physical Prototyping*, 8(3), 187-192.
- Khodasevych, I., Parmar, S., & Troynikov, O. (2017). Flexible sensors for pressure therapy: Effect of substrate curvature and stiffness on sensor performance. *Sensors*, 17(10). Retrieved from <http://www.mdpi.com/1424-8220/17/10/2399> doi: 10.3390/s17102399
- Laaveg, S. J., & Ponseti, I. V. (1980). Long-term results of treatment of congenital club foot [Journal Article]. *The Journal of Bone & Joint Surgery*, 62(1), 23-31. Retrieved from <http://sfx.utwente.nl:3210/prod?sid=google&auinit=SJ&aulast=Laaveg&atitle=Long-term%20results%20of%20treatment%20of%20congenital%20club%20foot.&id=pmid%3A7351412>
- Lampasi, M., Abati, C. N., Bettuzzi, C., Stilli, S., & Trisolino, G. (2018, Mar 29). Comparison of dimeglio and pirani score in predicting number of casts and need for tenotomy in clubfoot correction using the ponseti method. *International Orthopaedics*. Retrieved from <https://doi.org/10.1007/s00264-018-3873-3> doi: 10.1007/s00264-018-3873-3
- Larson, D., & Jerosch-Herold, C. (2008, Jul 21). Clinical effectiveness of post-operative splinting after surgical release of dupuytren's contracture: a systematic review. *BMC Musculoskeletal Disorders*, 9(1), 104. Retrieved from <https://doi.org/10.1186/1471-2474-9-104> doi: 10.1186/1471-2474-9-104
- Lis, A., de Castro, C., & Nordin, M. (2012). Basic biomechanics of the musculoskeletal system [Book]. In (4th ed., chap. Biomechanics of Tendons and Ligaments). Baltimore, MD: Lippincott Williams & Wilkins.
- Loureno, A. F., & Morcuende, J. A. (2007). Correction of neglected idiopathic club foot by the ponseti method [Journal Article]. *Journal of Bone &*

- Joint Surgery, British Volume*, 89-B(3), 378-381. Retrieved from <http://www.bjj.boneandjoint.org.uk/content/89-B/3/378.abstract> doi: 10.1302/0301-620x.89b3.18313
- ltar, S., Uysal, M., Alemdarolu, K. B., Aydoan, N. H., Kara, T., & Atlhan, D. (2010). Treatment of clubfoot with the ponseti method: Should we begin casting in the newborn period or later? [Journal Article]. *The Journal of Foot and Ankle Surgery*, 49(5), 426-431. Retrieved from <http://www.sciencedirect.com/science/article/pii/S1067251610002656> doi: 10.1053/j.jfas.2010.06.010
- Maathuis, P. G., de Jong, E., Hekman, E., Bulstra, S., & Verkerke, G. (2009). Krachtmetingen tijdens ponseti klompvoetredressie [Journal Article]. *Nederlands Tijdschrift voor Orthopaedie*, 16(1), 48.
- Mageshwaran, S., Murali, V. B., Devendran, R., Yoosuf, A., & Anandan, H. (2016). Evaluation of outcome of correction of clubfoot by conventional ponseti and accelerated ponseti [Journal Article]. *International Journal of Scientific Study*, 4(8), 199-202.
- Magnusson, S. (1998). Passive properties of human skeletal muscle during stretch maneuvers [Journal Article]. *Scandinavian journal of medicine & science in sports*, 8(2), 65-77.
- Mahmoodian, R. (2010). *Mechanics and development of tarsal cartilage anlagen — insights from a human study* (PhD Thesis). Drexel University.
- Malagelada, F., Mayet, S., Firth, G., & Ramachandran, M. (2016). The impact of the ponseti treatment method on parents and caregivers of children with clubfoot: a comparison of two urban populations in europe and africa [Journal Article]. *Journal of Children's Orthopaedics*, 10(2), 101-107. Retrieved from <http://dx.doi.org/10.1007/s11832-016-0719-7> doi: 10.1007/s11832-016-0719-7
- Marieb, E., & Hoehn, K. (2007). *Human anatomy & physiology* [Book]. San Francisco, California, USA: Pearson Benjamin Cummings. Retrieved from <https://books.google.nl/books?id=x1uEB68iitwC>

- Maripuri, S. N., Gallacher, P. D., Bridgens, J., Kuiper, J. H., & Kiely, N. T. (2013). Ponseti casting for club foot above-or below-knee? [Journal Article]. *Bone Joint J*, 95(11), 1570-1574. doi: 10.1302/0301-620X.95B11.31899
- Matos, M. A., & de Oliveira, L. A. A. (2010). Comparison between ponseti's and kite's clubfoot treatment methods: a meta-analysis. *The Journal of Foot and Ankle Surgery*, 395-397. doi: 10.1053/j.jfas.2010.04.020
- Maxim Integrated. (2005, Jan). DS1825 programmable resolution 1-wire digital thermometer with 4-bit ID [Computer software manual]. Retrieved from <https://datasheets.maximintegrated.com/en/ds/DS1825.pdf> (Rev 020105)
- McClure, P. W., Blackburn, L. G., & Dusold, C. (1994). The use of splints in the treatment of joint stiffness: biologic rationale and an algorithm for making clinical decisions [Journal Article]. *Physical Therapy*, 74(12), 1101-1107.
- Miller, N. H., Carry, P. M., Mark, B. J., Engelman, G. H., Georgopoulos, G., Graham, S., & Dobbs, M. B. (2015). Does strict adherence to the ponseti method improve isolated clubfoot treatment outcomes? a two-institution review [Journal Article]. *Clinical Orthopaedics and Related Research*, 1-7. Retrieved from <http://dx.doi.org/10.1007/s11999-015-4559-4> doi: 10.1007/s11999-015-4559-4
- Mohan, S., Kumar, S., Prashanth, L., & Krishnagopal, R. (2018). Management of idiopathic clubfoot by Ponseti's technique. *International Journal of Orthopaedics Sciences*, 4(1), 92-97. doi: 10.22271/ortho.2018.v4.i1b.16
- Mohan, S. M., Pawan, T., Mohan, S. S., & Madhu, T. (2014). Modified ponseti technique for the management of congenital talipes equino varus (ctev) in neonates: Our experience in a rural teaching institution [Journal Article]. *Journal of Evolution of Medical and Dental Sciences*, 3(14), 3618-3626. doi: 10.14260/jemds/2014/2331
- Morcuende, J. A., Abbasi, D., Dolan, L. A., & Ponseti, I. V. (2005). Results of an accelerated ponseti protocol for clubfoot [Journal Article]. *Journal of Pediatric Orthopaedics*, 25(5), 623-626.

- Morcuende, J. A., Dolan, L. A., Dietz, F. R., & Ponseti, I. V. (2004). Radical reduction in the rate of extensive corrective surgery for clubfoot using the ponseti method [Journal Article]. *Pediatrics*, 113(2), 376-380.
- Mosca, V. S. (2001). The foot [Book Section]. In R. T. Morrissy & S. L. Weinstein (Eds.), *Lovell and winter's pediatric orthopaedics* (5th ed., Vol. 2, p. 1151-1215). Philadelphia, PA, USA: Lippincott Williams & Wilkins.
- Murata, H. (2012). Rheology - theory and application of biomaterials [Book Section]. In A. De Souza Gomes (Ed.), *Polymerization* (chap. 17). InTech. doi: 10.5772/48393
- Nemet, G. F. (2009). Demand-pull, technology-push, and government-led incentives for non-incremental technical change. *Research Policy*, 38(5), 700-709. Retrieved from <http://www.sciencedirect.com/science/article/pii/S0048733309000080> doi: 10.1016/j.respol.2009.01.004
- Nogueira, M. P. (2016). Corr insights®: Does strict adherence to the ponseti method improve isolated clubfoot treatment outcomes? a two-institution review. *Clinical Orthopaedics and Related Research®*, 474(1), 244-245. doi: 10.1007/s11999-015-4600-7
- Nogueira, M. P., Farcetta, M., Fox, M. H., Miller, K. K., Pereira, T. S., & Morcuende, J. A. (2013). Treatment of congenital clubfoot with the ponseti method: the parents' perspective [Journal Article]. *Journal of Pediatric Orthopaedics B*, 22(6), 583-588.
- NXP. (2014, Aug). Kinetis KL25 sub-family (Computer software manual No. KL25P80M48SF0). Retrieved from <https://www.nxp.com/docs/en/data-sheet/KL25P80M48SF0.pdf> (Rev 5 08/2014)
- Oberhauser, C. (2015, March). *Ldc sensor design* (Application Report No. SNOA930). Dallas, Texas: Texas Instruments. Retrieved from <http://www.ti.com/lit/an/snoa930/snoa930.pdf>
- Oberhauser, C. (2016, February). *Optimizing l measurement resolution for the ldc161x and ldc1101* (Application Report No. SNOA944). Dallas, Texas: Texas Instruments. Retrieved from <http://www.ti.com/lit/an/snoa944/>

snoa944.pdf

- Olgaher, A., & Forsum, E. (2003). Total energy expenditure, body composition and weight gain in moderately preterm and full-term infants at term postconceptional age. *Acta Paediatrica*, 92, 1327-1334. doi: 10.1080/08035250310005396
- Özkaya, N., Nordin, M., Goldsheyder, D., & Leger, D. (2012). Mechanical properties of biological tissues [Book Section]. In *Fundamentals of biomechanics: Equilibrium, motion and deformation* (3rd ed., p. 221-236). New York: Springer. doi: 10.1007/978-1-4614-1150-5_15
- Parmar, S., Khodasevych, I., & Troynikov, O. (2017). Evaluation of flexible force sensors for pressure monitoring in treatment of chronic venous disorders [Journal Article]. *Sensors*, 17(8). doi: 10.3390/s17081923
- Patel, N. K., Jeer, P., & Cornell, M. S. (2007). Using 'cling film' to protect lower limb plaster casts in babies with club foot [Journal Article]. *British Journal of Nursing*, 16(18), 1140-1142. doi: 10.12968/bjon.2007.16.18.27509
- Peacock, E. E. (1984). *Wound repair* [Book]. Philadelphia, PA, USA: W.B. Saunders.
- Philip, P. (2014). *Volumetric and growth changes induced by serial casting treatment of congenital clubfoot on tarsal bones* (MSc thesis, Drexel University). Retrieved 2018-1-4, from <http://hdl.handle.net/1860/4490>
- Pierro, A., De Coppi, P., & Eaton, S. (2012). Neonatal physiology and metabolic considerations [Book Section]. In A. G. Coran & N. S. Adzick (Eds.), *Pediatric surgery (seventh edition)* (7th ed., p. 89-107). Philadelphia: Mosby. Retrieved from <http://www.sciencedirect.com/science/article/pii/B9780323072557000064> doi: 10.1016/B978-0-323-07255-7.00006-4
- Pirani, S. (2009). Clubfoot scoring [Book Section]. In S. Lynn (Ed.), *Clubfoot: Ponseti management* (3rd ed., p. 27). Global Help. Retrieved from <http://global-help.org/products/clubfoot-ponseti-management/>
- Pittner, D. E., Klingele, K. E., & Beebe, A. C. (2008). Treatment of clubfoot with the ponseti method: a comparison of casting materials [Jour-

- nal Article]. *Journal of Pediatric Orthopaedics*, 28(2), 250-253. doi: 10.1097/BPO.0b013e318164f8e7
- Pletch, A., Morcuende, J., Barriga, H., Segura, J., & Salas, A. (2015). Photovoice and clubfoot: Using a participatory research method to study caregiver adherence to the ponseti method in per [Journal Article]. *The Iowa orthopaedic journal*, 35, 160.
- Pohl, M., Rckriem, S., Mehrholz, J., Ritschel, C., Strik, H., & Pause, M. R. (2002). Effectiveness of serial casting in patients with severe cerebral spasticity: A comparison study [Journal Article]. *Archives of Physical Medicine and Rehabilitation*, 83(6), 784-790. Retrieved from <http://www.sciencedirect.com/science/article/pii/S0003999302603016> doi: 10.1053/apmr.2002.32821
- Ponseti, I. V. (2008). *Congenital clubfoot: Fundamentals of treatment* (2nd ed.) [Book]. New York, USA: Oxford University Press Inc.
- Ponseti, I. V., & Smoley, E. N. (1963). Congenital club foot: the results of treatment [Journal Article]. *The Journal of Bone and Joint Surgery*, 45(2), 261-344.
- Pratt, W. K. (2007). *Digital image processing* (4th ed.). Hoboken, NJ, USA: John Wiley & Sons. Retrieved from <https://nl.mathworks.com/help/signal/ref/medfilt1.html>
- Radler, C. (2013). The ponseti method for the treatment of congenital club foot: review of the current literature and treatment recommendations [Journal Article]. *International Orthopaedics*, 37(9), 1747-1753. Retrieved from <http://dx.doi.org/10.1007/s00264-013-2031-1> doi: 10.1007/s00264-013-2031-1
- Rahman, T., Akins, R. E., Wee, J., & Mackenzie, W. G. (2011). Continuous force measurement in limb lengthening [Journal Article]. *Journal of Bioengineering and Biomedical Sciences*, 1(2), 104. doi: 10.4172/2155-9538.1000104
- Rajput, R. (2010). *Thermal engineering* [Book]. Laxmi Publications Pvt Limited. Retrieved from <https://books.google.nl/books?id=65gxCX2dC84C>

- Ramahenina, H., OConnor, R. J., & Chamberlain, A. M. (2016). Problems encountered by parents of infants with clubfoot treated by the ponseti method in madagascar: a study to inform better practice [Journal Article]. *Journal of Rehabilitation Medicine*, 48(5), 481-483. doi: 10.2340/16501977-2082
- Reswick, J. B., & Rogers, J. E. (1976). Experience at rancho los amigos hospital with devices and techniques to prevent pressure sores [Book Section]. In R. M. Kenedi & J. M. Cowden (Eds.), *Bed sore biomechanics: Proceedings of a seminar on tissue viability and clinical applications organised in association with the department of biomedical engineering, the institute of orthopaedics (university of london), royal national orthopaedic hospital, stanmore, london, and held at the university of strathclyde, glasgow, in august, 1975* (p. 301-310). London: Macmillan Education UK. Retrieved from http://dx.doi.org/10.1007/978-1-349-02492-6_38 doi: 10.1007/978-1-349-02492-6_38
- Richards, B. S., Faulks, S., Rathjen, K. E., Karol, L. A., Johnston, C. E., & Jones, S. A. (2008). A comparison of two nonoperative methods of idiopathic clubfoot correction: The ponseti method and the french functional (physiotherapy) method. *The Journal of Bone and Joint Surgery*, 2313-2321. doi: 10.2106/JBJS.G.01621
- Riley, W. (2008). *Handbook of frequency: Stability analysis*. National Institute of Standards and Technology. Retrieved 2018-03-09, from <https://nvlpubs.nist.gov/nistpubs/Legacy/SP/nistspecialpublication1065.pdf>
- Rives, K., Gelberman, R., Smith, B., & Carney, K. (1992). Severe contractures of the proximal interphalangeal joint in dupuytren's disease: Results of a prospective trial of operative correction and dynamic extension splinting. *The Journal of Hand Surgery*, 17(6), 1153-1159. Retrieved from <http://www.sciencedirect.com/science/article/pii/S036350230991084X> doi: 10.1016/S0363-5023(09)91084-X
- Sahu, B., Rajavelu, R., & Tudu, B. (2015). Management of idiopathic congenital talipes equinovarus by standard versus accelerated ponseti plaster

- technique: A prospective study [Journal Article]. *Journal of Orthopedics, Traumatology and Rehabilitation*, 8(1), 30.
- Salim, M., Ahmed, M., & Rehman, M. U. (2018). Management of club foot by Ponseti method: our experience. *Pakistan Armed Forces Medical Journal*, 68(3), 589-593. Retrieved from http://www.pafmj.org/art_pdf/31-%202017-01-012-%20Dr%20Muneeb%20Ur%20Rehman.pdf
- Sanders, E. B. N., & Stappers, P. J. (2008). Co-creation and the new landscapes of design. *CoDesign*, 4(1), 5-18. Retrieved from <https://doi.org/10.1080/15710880701875068> doi: 10.1080/15710880701875068
- Sangiorgio, S. N., Ebrahimzadeh, E., Morgan, R. D., & Zions, L. E. (2017). The timing and relevance of relapsed deformity in patients with idiopathic clubfoot [Journal Article]. *The Journal of the American Academy of Orthopaedic Surgeons*, 25(7), 536-545. doi: 10.5435/JAAOS-D-16-00522
- Sanjeevi, R. (1982). A viscoelastic model for the mechanical properties of biological materials [Journal Article]. *Journal of biomechanics*, 15(2), 107-109.
- Sanjeevi, R., Somanathan, N., & Ramaswamy, D. (1982). A viscoelastic model for collagen fibres [Journal Article]. *Journal of biomechanics*, 15(3), 181-183.
- Sano, H., Uhthoff, H. K., Jarvis, J. G., Mansingh, A., & Wenkebach, G. F. (1998). Pathogenesis of soft-tissue contracture in club foot [Journal Article]. *J Bone Joint Surg Br*, 80(4), 641-4.
- Scher, D. M., Feldman, D. S., van Bosse, H. J., Sala, D. A., & Lehman, W. B. (2004). Predicting the need for tenotomy in the ponseti method for correction of clubfeet [Journal Article]. *Journal of Pediatric Orthopaedics*, 24(4), 349-352.
- Screen, H. R. C., Toorani, S., & Shelton, J. C. (2013). Microstructural stress relaxation mechanics in functionally different tendons [Journal Article]. *Medical Engineering & Physics*, 35(1), 96-102. Retrieved from <http://www.sciencedirect.com/science/article/pii/S1350453312000902> doi: 10.1016/j.medengphy.2012.04.004
- Shabtai, L., Specht, S. C., & Herzenberg, J. E. (2014). Worldwide spread of

- the ponseti method for clubfoot [Journal Article]. *World J Orthop*, 5(5), 585-590. doi: 10.5312/wjo.v5.i5.585
- Sharma, P., Yadav, V., Verma, R., Gohiya, A., & Gaur, S. (2016). Comparative analysis of results between conventional and accelerated ponseti technique for idiopathic congenital clubfoot [Journal Article]. *Orthopaedic Journal of MP Chapter*, 22(1).
- Shaw, R. B. J., Chong, A. K. S., Zhang, A., Hentz, V. R., & Chang, J. (2007). Dupuytren's disease: History, diagnosis, and treatment. *Plastic and Reconstructive Surgery*, 120(3), 44e-54e. doi: 10.1097/01.prs.0000278455.63546.03
- Shen, X., Tan, K., & Xie, M. (2000). An integrated approach to innovative product development using kanos model and qfd. *European Journal of Innovation Management*, 3(2), 91-99. doi: 10.1108/14601060010298435
- Shuler, F. D., & Bates, C. M. (2013). Skin temperatures generated following plaster splint application [Journal Article]. *Orthopedics*, 36(5), 364-367. doi: 10.3928/01477447-20130426-06
- Staheli, L. (2009). *Clubfoot: ponseti management* (Third edition ed.) [Book]. Global HELP Organization.
- Strik, H. (2000). Langzeittonusmessung unter redressierender gelenkbehandlung bei schwerer spastik [Journal Article]. *Akt Neurol*, 27. Retrieved from http://forschung.deutsche-rentenversicherung.de/ForschPortalWeb/ressource?key=tagungsband_10_reha-kolloqu.pdfpage307
- Sugandhavesa, N., Cheewawattanachai, C., Luevitoonvechkij, S., & Khunsree, S. (2015). Results of shortened program of ponseti technique for congenital clubfoot [Journal Article]. *Journal of the Medical Association of Thailand*, 98(1), 88-92.
- Sutcliffe, A., Vaea, K., Pouliwaati, J., & Evans, A. M. (2013). fast casts: Evidence based and clinical considerations for rapid ponseti method [Journal Article]. *The Foot and Ankle Online Journal*, 6(9). Retrieved from <http://faoj>

- .files.wordpress.com/2013/10/fastcastsfaoj.pdf doi: 10.3827/faoj.2012.0609.0002
- Tabary, J., Tabary, C., Tardieu, C., Tardieu, G., & Goldspink, G. (1972). Physiological and structural changes in the cat's soleus muscle due to immobilization at different lengths by plaster casts [Journal Article]. *The Journal of physiology*, 224(1), 231-244.
- Tekscan. (2012, Oct). Flexiforce®sensors users manual (Computer software manual No. Rev I). Retrieved from <https://www.tekscan.com/resources/product-guide/flexiforce-user-manual>
- Terrazas-Lafargue, G., & Morcuende, J. A. (2007). Effect of cast removal timing in the correction of idiopathic clubfoot by the ponseti method [Journal Article]. *The Iowa Orthopaedic Journal*, 27, 24-27.
- Texas Instruments. (2014, Dec). LDC1612, LDC1614 multi-channel 28-bit inductance to digital converter (LDC) for inductive sensing (Computer software manual No. SNOSCY9). Retrieved from <http://www.ti.com/lit/ds/symlink/ldc1614.pdf> (Initial release)
- Townley, W. A., Baker, R., Sheppard, N., & Grobbelaar, A. O. (2006). Dupuytren's contracture unfolded. *BMJ*, 332(7538), 397-400. Retrieved from <https://www.bmj.com/content/332/7538/397> doi: 10.1136/bmj.332.7538.397
- Treadwell, M. C., Stanitski, C. L., & King, M. (1999). Prenatal sonographic diagnosis of clubfoot: implications for patient counseling. *Journal of Pediatric Orthopaedics*, 19(1), 8-10.
- Uhl, T. L., & Jacobs, C. A. (2012). Torque measures of common therapies for the treatment of loss of knee flexion [Journal Article]. *Sports Health: A Multidisciplinary Approach*, 4(2), 101-106.
- Vale, R. (1969). Post-operative accidental hypothermia [Journal Article]. *Anaesthesia*, 24(3), 449-452.
- van Belle, J. S. (2017). *Improving brace compliance in clubfoot treatment: Design proposals based on philosophical analysis* (Capita Selecta Design As-

- signment). University of Twente.
- van Doorn, E. (2016). *A design for a new clubfoot treatment, the parents' view* (BSc Thesis, University of Twente). Retrieved 2018-1-4, from <http://essay.utwente.nl/70558/>
- van Meer, B. L., Meuffels, D. E., van Eijnden, W. A., Verhaar, J. A., Bierma-Zeinstra, S. M., & Reijman, M. (2015). Which determinants predict tibiofemoral and patellofemoral osteoarthritis after anterior cruciate ligament injury? a systematic review [Journal Article]. *British journal of sports medicine*, bjsports-2013-093258.
- van Rijssen, A. L., Gerbrandy, F. S., Linden, H. T., Klip, H., & Werker, P. M. (2006). A comparison of the direct outcomes of percutaneous needle fasciotomy and limited fasciectomy for dupuytren's disease: A 6-week follow-up study. *The Journal of Hand Surgery*, 31(5), 717 - 725. doi: 10.1016/j.jhssa.2006.02.021
- Van Tulder, M., Furlan, A., Bombardier, C., Bouter, L., & Group, E. B. o. t. C. C. B. R. (2003). Updated method guidelines for systematic reviews in the cochrane collaboration back review group [Journal Article]. *Spine*, 28(12), 1290-1299.
- Wang, H., Kow, J., de Boer, G., Jones, D., Alazmani, A., & Culmer, P. (2017, Oct). A low-cost, high-performance, soft tri-axis tactile sensor based on eddy-current effect [Journal Article]. In *2017 IEEE Sensors* (p. 1-3). doi: 10.1109/ICSENS.2017.8234098
- Werler, M. M., Yazdy, M. M., Mitchell, A. A., Meyer, R. E., Druschel, C. M., Anderka, M., ... Mahan, S. T. (2013). Descriptive epidemiology of idiopathic clubfoot [Journal Article]. *American Journal of Medical Genetics Part A*, 161(7), 1569-1578. Retrieved from <http://dx.doi.org/10.1002/ajmg.a.35955> doi: 10.1002/ajmg.a.35955
- Wess, K. M. (2009). *Charakterisierung des kollagens im ligamentum deltoideum beim kongenitalen klumpfu und beim gesunden fu* (Thesis). Heinrich-Heine-University.

- Williams, P. E. (1990). Use of intermittent stretch in the prevention of serial sarcomere loss in immobilised muscle [Journal Article]. *Annals of the Rheumatic Diseases*, 49(5), 316-317.
- Windisch, G., Salaberger, D., Rosmarin, W., Kastner, J., Exner, G. U., HaldiBrindle, V., & Anderhuber, F. (2007). A model for clubfoot based on microct data [Journal Article]. *Journal of anatomy*, 210(6), 761-766.
- Wolff, J. (1892). *Das gesetz der transformation der knochen*. [Book]. Berlin, Germany: Hirschwald.
- Xu, B., Li, H., & Zhang, Y. (2013). Understanding the viscoelastic behavior of collagen matrices through relaxation time distribution spectrum [Journal Article]. *Biomatter*, 3(3).
- Xu, R. J. (2011). A modified ponseti method for the treatment of idiopathic clubfoot: a preliminary report [Journal Article]. *Journal of Pediatric Orthopaedics*, 31(3), 317-319. doi: 10.1097/BPO.0b013e31820f7358
- Zhang, W., Richards, B. S., Faulks, S. T., Karol, L. A., Rathjen, K. A., & Browne, R. H. (2012). Initial severity rating of idiopathic clubfeet is an outcome predictor at age two years [Journal Article]. *Journal of Pediatric Orthopaedics B*, 21(1), 16-19. Retrieved from http://journals.lww.com/jpo-b/Fulltext/2012/01000/Initial_severity_rating_of_idiopathic_clubfeet_is.5.aspx doi: 10.1097/BPB.0b013e32834c31a2
- Zhao, D., Li, H., Zhao, L., Liu, J., Wu, Z., & Jin, F. (2014). Results of clubfoot management using the ponseti method: do the details matter? a systematic review. *Clinical Orthopaedics and Related Research®*, 472(4), 1329-1336.
- Zinober, S., & Has, R. (2014, 01 9). *Wirbelstromsensor und verfahren zum messen einer kraft* (Patent No. WO 2014005769 A2). Google Patents. Retrieved from <https://google.com/patents/WO2014005769A2> (WO Patent App. PCT/EP2013/061,024)
- Zionts, L. E., Jew, M. H., Ebrahimzadeh, E., & Sangiorgio, S. N. (2017). The influence of sex and laterality on clubfoot severity [Journal Article].

- cle]. *Journal of Pediatric Orthopedics*, 37(2), e129-e133. doi: 10.1097/BPO.0000000000000603
- Zosia, M. (2003). Congenital talipes equinovarus (clubfoot): a disorder of the foot but not the hand. *Journal of Anatomy*, 202(1), 37-42. Retrieved from <https://onlinelibrary.wiley.com/doi/abs/10.1046/j.1469-7580.2003.00147.x> doi: 10.1046/j.1469-7580.2003.00147.x

Acknowledgments

This project would not have been finished without the support, supervision, input and distraction provided by my family, friends and colleagues.

To all parents of the children that participated in this research: thank you. I can't imagine how stressful and scary it must be having a newborn child in the first place, let alone trusting us do our measurements on the foot of your child. Without your help none of our work would have been possible and I'm really grateful for your confidence. A big thanks to all participants of the Dupuytren study. You were all so kind and interested, the whole study felt like a big team effort. Finally, I would like to thank all doctors that participated in the rubber foot study.

I would like to thank Bart and Edsko for being my supervisors again. For my bachelor and master assignment you were great supervisors so I was happy when you invited me to opt for this PhD project. With Patrick on the team you gave me confidence that we would be able to fulfill my ambitions to improve (a small part of) the world with this project. I have always seen you two as one single entity "Bartenedsko" and only started to appreciate your individual values at a rather late stage of this project. Thank you for your guidance, trust and patience.

Patrick, you have told me time and time again that my *"enthousiasme is aanstekelijk"* but also that my approach towards science with medical doctors was

“na ief”. You have taught me valuable lessons about how medical doctors think and work and how to manage expectations. Without your frequent moderation I might have offended many more doctors than necessary. At the start of this project you warned me that to many physicians the Ponseti method is sacred. It would be hard, if not impossible, to have other physicians see any downsides of the Ponseti method. When the WKO relentlessly blocked our questionnaire to the parents’ experience with the Ponseti method I was really moved by how you stood up for me towards the WKO. Thank you very much!

I would like to thank the rest of my graduation committee for their time and effort in reading and commenting on my work. Marieke, although you weren’t my official supervisor or colleague, you were the closest thing to having a partner in this research. I like to think of our combined work as the clubfoot research department of the Netherlands with frequent exchanges of knowledge, people and material. I am proud of our review paper in PLOS One, thank you!

I would like to thank all members of the user committee for their involvement in this project. At every meeting we had very interesting discussions and I’m happy that you put so much time and effort in this project.

Robert (Nederlandse Klompvoet Vereniging) and Remco (Spierziekten Nederland), I think it was the very first day of my project that we sat with the three of us to discuss how we would revolutionize the clubfoot-world. Back then it all seemed so simple and I can recall that we even discussed what to do with all that spare time after we would have had developed this dynamic clubfoot brace. Thank you for continually stressing the relevance and importance of this project.

Marion (Orthin) you told me that in Groningen alone you saw about 100 clubfeet per year and it was quite a shock when Eindhoven’s research showed that it was rather 180 clubfeet per year, but for the whole of Netherlands! For a moment I thought that you would loose it, but I’m happy that you didn’t and that you stayed involved until the very end. Thank you for your continuous

support and advice. Karel, I think we had good fun trying to come up with new ideas for a clubfoot brace and I admire your patience with all students that you helped during the course of this project.

Laurens (Baat medical) and later Dymphy and Emma thank you for efforts in guiding this project towards the end. Sometimes it seemed like you could hardly hide your doubts in the approach and planning of this project and I wish you would have stood up a bit more to tell us how to do it better. I respect your professional attitude and I have really enjoyed our brainstorming at your location where Harry was absolutely key.

This project was part of the Symbionics project. Arjen, thank you for your starting and coordinating this project and bringing us all together. Although I think our projects do not have strong links, we as a people fit rather well! Niek, Stergios, Kostas, Teun, Claudia, Laura, Roy, Ronald, Nauzef, Hans and even Mariska and Joan, with a varying mix of people we really had some great times in some forest, in Egmond aan Zee and in Cape Town! I hope we'll keep running into each other like we did in Addo National Park and in the Karwei in Groningen. You guys are great!

Many people from outside our department have helped me understand my project. Joop taught me how to cast rubber, Nadja showed how to analyze it, Cindy helped to keep the project ethically sound, Remco explained and checked how my own sensor works, and De Vlam helped with understanding the thermodynamics of a baby foot wrapped in plaster cast. Every time I entered a new department I was amazed about the many fascinating projects we do at the University of Twente and felt proud to be part of that.

Hans-André, André and all others from ZGT Hengelo, and André and all others from UMCG, thanks for having me and my students!

Anne, Martijn, Alex, Hanneke and all others from Handencentrum Enschede, it has been such a fun experience to set up a study together in a field that I knew

so little about. I'm very proud of our paper in the Journal of Hand Therapy. Martijn, climbing was fun!

For years the sign on the door of our office said Kostas, Stergios, Alex and Bob. So, where were you? Luckily Alex sometimes dropped by with some great facts about time travel or tea or a forgotten religion. I was so happy when we moved downstairs and soon got a great new roomie in Tycho. Thanks for letting me harass you with constant questions about common filters, Matlab and tea. The final few months Eline joined and a couple of times, when both Zhara and Noortje were there too, we had all desks occupied! Even though our projects did not have any overlap, I'm so happy to have had you as my direct colleagues.

Some of my colleagues (certainly Victor and Niek) will feel that they have done most of the work in this PhD and perhaps they are right. However, to me it still felt like a ton of work so I wish you would have done a bit more! I am sorry that I could not pay back properly when you asked me for advice on presentations, fonts or colors. Victor, Niek, Arvid and sometimes even Gijs, thanks for teaching me everything that's too abstract for my petty designer brains to process. With you as direct neighbors developing a force sensor felt easy! I bet that any of you could have done within a week what have cost me most of my PhD. Martijn and Kyrian, you make everything seem easy by saying "*why don't you just...*" and then say something I completely didn't think of before. To all of my colleagues, some of you have left years ago, thank you!

Lianne and Jeanine, without your management the actual management of the department could not have managed. You were always there to help out if we made a mess of it again, or when things got too frustrating. I am happy to have had you as our secretary for all these years.

To all students that have helped in this project: thank you so much for your enthusiasm and ambition. I really mean it when I say that you made this project fun! To all students from Integrative Design of Biomedical Products I'd like

to say sorry for telling you to make a function analysis and a morphological scheme. I still think these tools are utterly useless and we should not bore you with useless tools. Despite this, you were still capable of creating the craziest creative ideas (liquid magnets, what?!) for a clubfoot brace and you always found ways to surprise us. Mariska, Elleke, Esther, Annelies, Ainhua, Eva thank you for the good ideas for the design of a clubfoot brace. Without your efforts the last chapter of this thesis would have been much shorter. Together with Elleke and Esther we have come to know this clubfoot-world as a frustrating one, with difficult doctors to work with. Jonne, your perspective on this project was so insanely broad that for a brief moment it even changed Edsko's utopian view on technology (not Bart's of course). I'd like to thank Hamid and Jasper for helping in the development of the force sensor, Marnick and Robin for their patience with testing our first prototypes on habitual toe walkers, and Maaïke, Els and Rohan for the measurements with clubfoot. Karlijn, Lenardo, Rohan and Marc, thank you so much for your work on the development of a clubfoot model. We were warned by other researchers that developing such a model would be a PhD project on its own and I guess they were right. Still, I think together we have learned so much about the movement of a clubfoot.

My futsal team! I still remember when Hielke asked me, five minutes before the match, if I wanted to join. Hell yeah, I'll take every excuse to leave my computer screen for a while! Few minutes later I was sweating like a pig while playing in my jeans and normal shoes. After the match I had no idea what to do behind my desk in my soaking wet T-shirt and just went home. Back then it was mostly Vincenzo, René, Jin, Jan and Hielke and we were rather serious and good. When you guys left I took over the honorable task of being captain and everything changed. In my team I only accepted girls (Astrid, Marjolein, Simone, Maaïke, Christina) or guys that had absolutely no ball skills (Teun, Marc, Stefan, Jan, Ronald, Kostas). I thought we played for fun and not to win! Strange enough you guys always won when I had to miss a match. I'm sure you guys have started to win matches again since Ronald took over my powers as captain. Thanks for

all the great matches and fun!

I hope I didn't bother any of you with my research, but to all people from Cube, TCT, Gropo, Radium and the rest of our climbing community: thanks for showing me that bouldering is really the most fun thing in life. I'm looking forward to meeting all of you again at the next event, competition or trip to Fontainebleau.

Then my family. I'm sure I've bothered you with it, and I'm happy I did! You were patient enough just to hear my complaints without immediately suggesting a solution that I was not ready for. I'm looking forward not to discuss work anymore but rather talk about what to explore in the south of Limburg.

We're getting towards the most important people in finishing this thesis.

Jort. You've done the exact same process, so why didn't you warn me? Next time, please just tell me not to do it! Will it be the same for the bouldergym that you first let me start it and then after a few years tell me that it sucks? We desperately need some nice weekends in Bleau to heal from this crazy PhD-battle.

Niek, I've mentioned you before, it was really great to have you as a roommate. I bet that our neighbors must have thought us of as quite a cute couple, which we were! I'm glad it's over though, and I'm sure soon you'll be done too. Hold on! Endure! Persevere! The same goes for you Eline, you were quite a nice roomie in the office but I prefer to finally have you as my roomie at home. I was so sad when you left to England and missed you really hard. It's great to have you close by and start a new adventure in Maastricht. You're the best!

Bob Giesberts Maastricht, May 2018

About the author

Bob Giesberts was born March 7th, 1987, in Groningen (the Netherlands). He got his high school degree in 2005 at the Reitdiep College in Groningen. He studied Industrial Design for his Bachelor's degree at the University of Twente, with a minor in Psychology at the Norwegian University of Science and Technology in Trondheim.

During his Master's degree in Industrial Design Engineering at the University of Twente he followed the track Design of Biomedical Products within the specialization of Emerging Technology Design. He received his Master's degree, cum laude, at the University of Twente after spending a month at Harapan Jaya in Pe-matang Siantar, North-Sumatra, Indonesia to work on the development of lower leg prostheses. He won the best MSc thesis award for the faculty of Engineering Technology for the academic year 2012/2013.

During and after his Master's degree he was the secretary of the MoveAid Foundation which aimed at implementation of biomedical technologies in developing countries, in which he supervised several international projects. Additionally he helped developing the project of Product Co-Creation Centers in slums in Brazil.

At the beginning of 2015 he started his PhD at the department of Biomechanical Engineering at the University of Twente, under supervision of prof. dr.

ir. Bart Verkerke and ir. Hekman. He supervised eleven BSc students and five MSc students from various studies.

He is currently manager and co-owner of bouldergym Radium Boulders in Maastricht.

Publications

Journal publications

1. **Giesberts, R.B.**, Hekman, E.E.G., Maathuis, P.G.M., & Verkerke, G.J. (2017). Quantifying the Ponseti method. *Journal of the Mechanical Behavior of Biomedical Materials*, 66, 45-49. doi: 10.1016/j.jmbbm.2016.10.021.
2. **Giesberts, R.B.**, van der Steen, M.C., Maathuis, P.G.M., Besselaar, A.T., Hekman, E.E.G. & Verkerke, G.J. (2018) Influence of cast change interval in the Ponseti method: a systematic review. *PLoS One*, 13(6): e0199540. doi: 10.1371/journal.pone.0199540.
3. **Giesberts, R.B.**, Hekman, E.E.G., Verkerke, G.J. & Maathuis, P.G.M. (n.d.) Rapid decrease of cast-induced forces in the Ponseti method. Accepted for publication in *The Bone & Joint Journal* (21-Sept-2018).
4. **Giesberts, R.B.**, Hekman, E.E.G., Verkerke, G.J. & Maathuis, P.G.M. (2018) Ten Cold Clubfeet. *Acta Orthopaedica*, 0(0) 1-5. doi: 10.1080/17453674.2018.1493046.
5. **Giesberts, R.B.**, ter Haar, A.M., Sanderman, G.M., Hekman, E.E.G. & Verkerke, G.J. (n.d.) Tissue adaptation rate in the treatment of Dupuytren contracture. Accepted for publication in *Journal of Hand Therapy* (22-Sept-2018).

6. **Giesberts, R.B.**, Sluiter, V.I., & Verkerke, G.J. (2018) Design and Test of a New Inductive Force Sensor. *MDPI Sensors*, 18(7), 2079. doi: 10.3390/s18072079.
7. **Giesberts, R.B.**, Ennion, L., Hjelmstrom, O., Karma, A., Lechler, K., Hekman, E., & Bergsma A. (2018) The modular socket system in a rural setting in Indonesia. *Prosthetics and Orthotics International*, 42(3), 336-343. doi: 10.1177/0309364617741917

International conference publications

1. van Belle, J., **Giesberts, R.B.**, & Eggink, W. (2018) The Use of Philosophy of Technology in Design: A Research-Through-Design Case of Treatment Compliance. Proceedings of the biannual Design Research Society conference (DRS): Catalyst. Storni, C., Leahy, K., McMahon, M., Lloyd, P. & Bohemia, E. (eds.). Limerick: Design Research Society, p. 218-228.
2. **Giesberts, R.B.**, Hekman, E.E.G., Maathuis, P.G.M., & Verkerke, G.J. (2017) Temperature changes in serial casting in the treatment of clubfoot. Poster session presented at XXVI Congress of the International Society of Biomechanics, ISB2017, Brisbane, Australia.
3. **Giesberts, R.B.**, Steen, van der, M.C., Hekman, E.E.G., Maathuis, P.G.M., Besselaar, A.T., & Verkerke, G.J. (2017). Length of the casting phase in the treatment of clubfoot. Poster session presented at XXVI Congress of the International Society of Biomechanics, ISB2017, Brisbane, Australia.
4. **Giesberts, R.B.**, Fink, A., Hekman, E.E.G., & Verkerke, G.J. (2017). Adaptation Rate In The Treatment Of Stiff Fingers. Poster session presented at 16th ISPO Wold Congress 2017, Cape Town, South Africa.
5. **Giesberts, R.B.**, Maten, A.C.A., Kolk, E., Hekman, E.E.G., Verkerke, G.J., & Bergsma, A. (2015). Development of a Lower Leg Prosthesis for

a Mobile Workshop in rural Indonesia. Poster session presented at 15th ISPO World Congress 2015, Lyon, France.

Ponseti's clubfoot treatment

a method in need of correction?

A clubfoot is a common congenital deformity of the foot. Worldwide the Ponseti method is the accepted treatment method for clubfoot. In this method, the treating physician manipulates the foot into a slightly better position and fixates this position with a plaster cast that stays on for a week. After five or six weekly cast changes the position of the foot is corrected. An abduction brace is worn for several years to prevent relapse.

The research in this thesis describes how aspects of the Ponseti method are quantified and gives pointers for improvements. Both a literature study and a study with force sensors on the foot suggest that the weekly cast change interval is unnecessary long. A surprising observation was a temperature drop due to water evaporating from the cast, creating an uncomfortably long cold period for the children.

Concepts for a dynamic clubfoot brace were developed as an alternative treatment method. Such a brace would apply a constant force on the foot rather than a constant position, making the correction process more efficient and possibly more comfortable. Parents would be able to temporarily remove the brace to allow bathing of their child and the soft materials of the brace makes cuddling a pleasant experience again.

*Bob Giesberts
PhD Candidate
University of Twente*

UNIVERSITY OF TWENTE.

ISBN 978-90-365-4587-7

9 789036 545877